

Label location



The label will be located at the rear side panel of the patch unit.
The FCC part 15 section 15.19(3) statement included in User Guide section 1.3.

1		2		3		4			
Rev.		Description				ECO No. <small>(USER #)</small>		DRWN <small>(REV APVCS)</small>	
01		First Issue						Vadim D.	

cat #: 0155-032 Rev. 01

SOLO PUMP


FCC ID: WNR08SOLORP4

REF




0114-11

SN:

XXXXXX

www.medingo.com

IPX4



Rx only

Original size

30.5

13.5

SOLO PUMP


FCC ID: WNR08SOLORP4

REF




0114-11

SN:



XXXXXX

www.medingo.com

IPX4



Rx only

CONFIDENTIAL. PROPERTY OF MEDINGO MEDICAL SOLUTIONS. CANNOT BE DIVULGED WITHOUT WRITTEN AUTHORIZATION.											
<div><div>MEDINGO</div><div>Medical Solutions</div><div>ISO STANDARDS</div><div>DO NOT SCALE DRAWING.</div><div>MODEL</div><div>DWG.</div><div>CONFIG.</div></div>		PART NAME		NAME		DATE		SIGN		ALL DIMENSIONS ARE IN mm. GENERAL TOL. ± 0.1 ANGLES $\pm 0.5^\circ$ SCALE 1:1	
		SOLO Pump Lable		Vadim							
		PART NO.		R & D <small>(PARENT ADM)</small>							
		SUBJECT		MANUF <small>(DRWN)</small>							
		DWG NO.		Q.A. <small>(APVCS)</small>							
0155-032		REV. 01		MATERIAL & HARDNESS Polyaster Transparent Label 14*31 mm							
		FINISH/TREATMENT									
		NA									

PAPER SIZE: A4 SHEET: 1 OF: 1