

Subject : Response to Inquiry to FCC (Tracking Number 572418) (TCB)  
Date : Wednesday, March 20, 2013 5:40 am  
Linked to: Generic Office of Engineering (Main FCC Record)  
From : [oetech@fccsun27w.fcc.gov](mailto:oetech@fccsun27w.fcc.gov)  
To : [<karens@ComplianceTesting.com>](mailto:karens@ComplianceTesting.com)



[FCC Home](#) | [Search](#) | [RSS](#) | [Updates](#) | [E-Filing](#) | [Initiatives](#) | [Consumers](#) | [Find People](#)



## Office of Engineering and Technology

### **Inquiry on 03/19/2013 :**

#### **Inquiry:**

To Whom It May Concern,

Compliance Testing is requesting FCC guidance for RF Exposure Exemption on the Biomet Vanguard Knee Balancer, FCC ID: XNL-ORTHOSNSR2.

Attached are the following documents - User's Manual, FCC Part 1.1310 RF Exposure Test Report and the FCC Part 95 Test Report

We look forward to hearing from you shortly.

Respectfully submitted,

Karen Springer

### **FCC response on 03/20/2013**

The implant device is eligible for SAR exclusion under KDB 447498. The SAR analysis report provided in this KDB must be included with the application. Please be reminded that this device is subject to PBA under KDB 388624 (B)(8).

#### **Attachment Details:**

Do not reply to this message. Please select the [Reply to an Inquiry Response](#) link from the OET Inquiry System to add any additional information pertaining to this inquiry.