




Care Innovations

Coordinator

FCC 15.247:2013

Report #: CARE0015.7



Report Prepared By Northwest EMC Inc.

NORTHWEST EMC – (888) 364-2378 – www.nwemc.com

California – Minnesota – Oregon – New York – Washington

CERTIFICATE OF TEST

Last Date of Test: May 1, 2013
Care Innovations
Model: Coordinator


Emissions

Test Description	Specification	Test Method	Pass/Fail
Occupied Bandwidth	FCC 15.247:2013	ANSI C63.10:2009	Pass
Output Power	FCC 15.247:2013	ANSI C63.10:2009	Pass
Band Edge Compliance	FCC 15.247:2013	ANSI C63.10:2009	Pass
Spurious Conducted Emissions	FCC 15.247:2013	ANSI C63.10:2009	Pass
Power Spectral Density	FCC 15.247:2013	ANSI C63.10:2009	Pass
Spurious Radiated Emissions	FCC 15.247:2013	ANSI C63.10:2009	Pass
AC Powerline Conducted Emissions	FCC 15.207:2013	ANSI C63.10:2009	Pass

Deviations From Test Standards

None

Approved By:



Tim O'Shea, Operations Manager



NVLAP Lab Code: 200630-0

Test Facility

The measurement facility used to collect the data is located at:

Northwest EMC, Inc.
22975 NW Evergreen Parkway, Suite 400
Hillsboro, OR 97124

Phone: (503) 844-4066 Fax: 844-3826

This site has been fully described in a report filed with and accepted by the FCC (Federal Communications Commission) and Industry Canada (Site filing #2834D-1).

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government of the United States of America.

Product compliance is the responsibility of the client, therefore the tests and equipment modes of operation represented in this report were agreed upon by the client, prior to testing. This Report may only be duplicated in its entirety. The results of this test pertain only to the sample(s) tested. The specific description is noted in each of the individual sections of the test report supporting this certificate of test.

REVISION HISTORY

Revision Number	Description	Date	Page Number
00	None		

Barometric Pressure

The recorded barometric pressure has been normalized to sea level.

United States

FCC - Designated by the FCC as a Telecommunications Certification Body (TCB). Certification chambers, Open Area Test Sites, and conducted measurement facilities are listed with the FCC.

A2LA - Accredited by A2LA to ISO / IEC Guide 65 as a product certifier. This allows Northwest EMC to certify transmitters to FCC and IC specifications.

NVLAP - Each laboratory is accredited by NVLAP to ISO 17025

Canada

IC - Recognized by Industry Canada as a Certification Body (CB). Certification chambers and Open Area Test Sites are filed with IC.

European Union

European Commission – Validated by the European Commission as a Conformity Assessment Body (CAB) under the EMC directive and as a Notified Body under the R&TTE Directive.

Australia/New Zealand

ACMA - Recognized by ACMA as a CAB for the acceptance of test data.

Korea

KCC / RRA - Recognized by KCC's RRA as a CAB for the acceptance of test data.

Japan

VCCI - Associate Member of the VCCI. Conducted and radiated measurement facilities are registered.

Taiwan

BSMI – Recognized by BSMI as a CAB for the acceptance of test data.

NCC - Recognized by NCC as a CAB for the acceptance of test data.

Singapore

IDA – Recognized by IDA as a CAB for the acceptance of test data.

Hong Kong

OFTA – Recognized by OFTA as a CAB for the acceptance of test data.

Vietnam

MIC – Recognized by MIC as a CAB for the acceptance of test data.

Russia

GOST – Accredited by Certinform VNIINMASH, CERTINFO, SAMTES, and Federal CHEC to perform EMC and Hygienic testing for Information Technology products to GOST standards.

SCOPE

For details on the Scopes of our Accreditations, please visit:

<http://www.nwemc.com/accreditations/>

Measurement Uncertainty

When a measurement is made, the result will be different from the true or theoretically correct value. The difference is the result of tolerances in the measurement system that cannot be completely eliminated. To the extent that technology allows us, it has been our aim to minimize this error. Measurement uncertainty is a statistical expression of measurement error qualified by a probability distribution.


A measurement uncertainty estimation has been performed for each test per our internal quality document WP 342. The estimation is used to compare the measured result with its "true" or theoretically correct value. The expanded measurement uncertainty (K=2) for each test is listed below. Our measurement data meets or exceeds the measurement uncertainty requirements of the applicable specification; therefore, the test data can be compared directly to the specification limit to determine compliance. The calculations for estimating measurement uncertainty are based upon ETSI TR 100 028 (or CISPR 16-4-1 as applicable), and are available upon request.

The following table represents the Measurement Uncertainty (MU) budgets for each of the tests that may be contained in this report.

Test	+ MU	- MU
Frequency Accuracy (Hz)	0.12	-0.01
Amplitude Accuracy (dB)	0.49	-0.49
Conducted Power (dB)	0.41	-0.41
Radiated Power via Substitution (dB)	0.69	-0.68
Temperature (degrees C)	0.81	-0.81
Humidity (% RH)	2.89	-2.89
Field Strength (dB)	3.80	-3.80
AC Powerline Conducted Emissions (dB)	2.94	-2.94



Oregon Labs EV01-12 22975 NW Evergreen Pkwy Hillsboro, OR 97124 (503) 844-4066	California Labs OC01-13 41 Tesla Irvine, CA 92618 (949) 861-8918	New York Labs NY01-04 4939 Jordan Rd. Elbridge, NY 13060 (315) 685-0796	Minnesota Labs MN01-08 9349 W Broadway Ave. Brooklyn Park, MN 55445 (763) 425-2281	Washington Labs NC01-05, SU02, SU07 19201 120 th Ave. NE Bothell, WA 98011 (425) 984-6600
VCCI				
A-0108	A-0029		A-0109	A-0110
Industry Canada				
2834D-1, 2834D-2	2834B-1, 2834B-2, 2834B-3		2834E-1	2834C-1
NVLAP				
NVLAP Lab Code: 200630-0	NVLAP Lab Code: 200676-0	NVLAP Lab Code: 200761-0	NVLAP Lab Code: 200881-0	NVLAP Lab Code: 200629-0





PRODUCT DESCRIPTION

Client and Equipment Under Test (EUT) Information

Company Name:	Care Innovations
Address:	20270 NW Amberglen Court
City, State, Zip:	Beaverton, OR 97006
Test Requested By:	Bill Morse
Model:	Coordinator
First Date of Test:	April 18, 2013
Last Date of Test:	May 1, 2013
Receipt Date of Samples:	April 17, 2013
Equipment Design Stage:	Production
Equipment Condition:	No Damage

Information Provided by the Party Requesting the Test

Functional Description of the EUT (Equipment Under Test):
Wireless monitoring technology that uses a 2.4 GHz ISM radio module, 802.15.4 compliant with 1 antenna.
Testing Objective:
To demonstrate compliance to FCC 15.247 requirements.

Configuration CARE0015- 1

EUT			
Description	Manufacturer	Model/Part Number	Serial Number
Coordinator	Care Innovations	QC101200-01	001D40000031002F

Peripherals in test setup boundary			
Description	Manufacturer	Model/Part Number	Serial Number
Laptop	Dell	Latitude 2100	00196-063-869-320

Cables					
Cable Type	Shield	Length (m)	Ferrite	Connection 1	Connection 2
USB to Radio Programmer	No	0.5	No	Laptop	Coor/Router/Door
PA = Cable is permanently attached to the device. Shielding and/or presence of ferrite may be unknown.					

Configuration CARE0016- 2

EUT			
Description	Manufacturer	Model/Part Number	Serial Number
Coordinator	Care Innovations	QC101200-01	001D40000031002F

Peripherals in test setup boundary			
Description	Manufacturer	Model/Part Number	Serial Number
Facility Server	Kontron	KB-KTUS15-04	QC101601491267
DC Power Supply	FSP Group Inc	FSP060-DBAB1	H2381001241
Radio programmer	Texas Instruments	CC Debugger	None
UPS	APC	Pro 1500	3B1217X21928

Remote Equipment Outside of Test Setup Boundary			
Description	Manufacturer	Model/Part Number	Serial Number
Laptop	Dell	Latitude 2100	00196-063-869-320

Cables					
Cable Type	Shield	Length (m)	Ferrite	Connection 1	Connection 2
DC Power	No	1.4	No	DC Power Supply	Facility Server
AC Power	No	1.8	No	AC Mains	UPS
AC Power	No	2.1	Yes	AC Mains	DC Power Supply
USB to Radio Programmer	No	0.5	Yes	Laptop	Coordinator
USB to Mod 10	No	0.5	Yes	UPS (Mod 10)	Facility Server (USB)
PA = Cable is permanently attached to the device. Shielding and/or presence of ferrite may be unknown.					

Equipment Modifications

Item	Date	Test	Modification	Note	Disposition of EUT
1	4/18/2013	Spurious Radiated Emissions	Tested as delivered to Test Station.	No EMI suppression devices were added or modified during this test.	EUT remained at Northwest EMC following the test.
2	4/22/2013	Power Spectral Density	Tested as delivered to Test Station.	No EMI suppression devices were added or modified during this test.	EUT remained at Northwest EMC following the test.
3	4/22/2013	Band Edge Compliance	Tested as delivered to Test Station.	No EMI suppression devices were added or modified during this test.	EUT remained at Northwest EMC following the test.
4	4/22/2013	Occupied Bandwidth	Tested as delivered to Test Station.	No EMI suppression devices were added or modified during this test.	EUT remained at Northwest EMC following the test.
5	4/22/2013	Output Power	Tested as delivered to Test Station.	No EMI suppression devices were added or modified during this test.	EUT remained at Northwest EMC following the test.
6	4/22/2013	Spurious Conducted Emissions	Tested as delivered to Test Station.	No EMI suppression devices were added or modified during this test.	EUT remained at Northwest EMC following the test.
7	5/1/2013	AC Powerline Conducted Emissions	Modified from delivered configuration.	Added ferrite to Facility Server power supply the USB cable connecting the laptop to the Coordinator. Modification authorized by Bill Morse.	Scheduled testing was completed.

Occupied Bandwidth

Testing was performed using the mode(s) of operation and configuration(s) noted within the report. The individuals and/or the organization requesting the test provided the modes, configurations and settings used to complete the evaluation. The actual test parameters are specified in the test data, this includes items such as investigated frequency range (scanned) and test levels. The testing methods and performance specifications, as well as the test site used for the evaluation are indicated in the test data.

TEST EQUIPMENT

Description	Manufacturer	Model	ID	Last Cal.	Interval
Attenuator, 6dB	S.M. Electronics	18N-06	AWN	3/25/2013	12
MXG Vector Signal Generator	Agilent	N5182A	TIF	NCR	0
Power Meter	Gigatronics	8651A	SPM	1/9/2012	24
Power Sensor	Gigatronics	80701A	SPL	7/8/2011	24
EV06 Direct Connect Cable	ESM Cable Corp.	TT	ECA	NCR	0
Attenuator 20 dB, SMA M/F 26GHz	S.M. Electronics	SA26B-20	AUY	8/2/2012	12
40GHz DC Block	Miteq	DCB4000	AMD	6/25/2012	12
Spectrum Analyzer	Agilent	E4440A	AFD	7/5/2012	24

TEST DESCRIPTION

The 6dB occupied bandwidth was measured using 100 kHz resolution bandwidth and 300 kHz video bandwidth. The 26 dB (99.9%) emission bandwidth (EBW) was also measured at the same time.

The EUT was set to low, medium and high transmit frequencies. The measurement was made using a direct connection between the RF output of the EUT and the spectrum analyzer. The EUT was transmitting at the data rate(s) listed in the datasheet.




Occupied Bandwidth

XMit 2013.02.28
PsaTx 2013.01.10

EUT: Coordinator		Work Order: CARE0015	
Serial Number: 001D40000031002F		Date: 04/22/13	
Customer: Care Innovations		Temperature: 23°C	
Attendees: Bill Morse		Humidity: 29%	
Project: None		Barometric Pres.: 1032	
Tested by: Brandon Hobbs, Rod Peloquin		Power: 5VDC	
		Job Site: EV06	
TEST SPECIFICATIONS		Test Method	
FCC 15.247:2013		ANSI C63.10:2009	
COMMENTS			
The EUT was operating at 100% duty cycle while under test.			
DEVIATIONS FROM TEST STANDARD			
None			
Configuration #	1	Signature <i>Brandon Hobbs</i>	
		Value	Limit
2400 MHz - 2483.5 MHz Band			Result
OQPSK			
Low Channel 2405 MHz		1.577 MHz	> 500 kHz
Mid Channel 2445 MHz		1.632 MHz	> 500 kHz
High Channel 2480 MHz		1.533 MHz	> 500 kHz


2400 MHz - 2483.5 MHz Band, OQPSK, Low Channel 2405 MHz

Value	Limit	Result
1.577 MHz	> 500 kHz	Pass




2400 MHz - 2483.5 MHz Band, OQPSK, Mid Channel 2445 MHz

Value	Limit	Result
1.632 MHz	> 500 kHz	Pass



2400 MHz - 2483.5 MHz Band, OQPSK, High Channel 2480 MHz

Value	Limit	Result
1.533 MHz	> 500 kHz	Pass



Output Power

Testing was performed using the mode(s) of operation and configuration(s) noted within the report. The individuals and/or the organization requesting the test provided the modes, configurations and settings used to complete the evaluation. The actual test parameters are specified in the test data, this includes items such as investigated frequency range (scanned) and test levels. The testing methods and performance specifications, as well as the test site used for the evaluation are indicated in the test data.

TEST EQUIPMENT

Description	Manufacturer	Model	ID	Last Cal.	Interval
Attenuator, 6dB	S.M. Electronics	18N-06	AWN	3/25/2013	12
MXG Vector Signal Generator	Agilent	N5182A	TIF	NCR	0
Power Meter	Gigatronics	8651A	SPM	1/9/2012	24
Power Sensor	Gigatronics	80701A	SPL	7/8/2011	24
EV06 Direct Connect Cable	ESM Cable Corp.	TT	ECA	NCR	0
Attenuator 20 dB, SMA M/F 26GHz	S.M. Electronics	SA26B-20	AUY	8/2/2012	12
40GHz DC Block	Miteq	DCB4000	AMD	6/25/2012	12
Spectrum Analyzer	Agilent	E4440A	AFD	7/5/2012	24

TEST DESCRIPTION

The transmit frequency was set to the required channels in each band. The transmit power was set to its default maximum. A direct connection was made between the RF output of the EUT and a spectrum analyzer. Attenuation and a DC block were used. The reference level offset on the spectrum analyzer was adjusted to compensate for cable loss and the external attenuation used between the RF output and the spectrum analyzer input.

Method Option 1 found in KDB 558074 DTS D01 Measurement Section 9.1.1 was used because the RBW on the analyzer was greater than the Emission Bandwidth of the radio.

De Facto EIRP Limit: Per 47 CFR 15.247 (b)(1-3), the EUT meets the de facto EIRP limit of +36 dBm.




Output Power

XMit 2013.02.28
PsaTx 2013.01.10

EUT: Coordinator		Work Order: CARE0015	
Serial Number: 001D40000031002F		Date: 04/22/13	
Customer: Care Innovations		Temperature: 23°C	
Attendees: Bill Morse		Humidity: 29%	
Project: None		Barometric Pres.: 1032	
Tested by: Brandon Hobbs, Rod Peloquin		Power: 5VDC	
		Job Site: EV06	
TEST SPECIFICATIONS		Test Method	
FCC 15.247:2013		ANSI C63.10:2009	
COMMENTS			
The EUT was operating at 100% duty cycle while under test.			
DEVIATIONS FROM TEST STANDARD			
None			
Configuration #	1	Signature <i>Rodry Le Pellego</i>	
		Value	Limit
2400 MHz - 2483.5 MHz Band			Result
OQPSK			
Low Channel 2405 MHz		10.72 mW	< 1 W
Mid Channel 2445 MHz		7.73 mW	< 1 W
High Channel 2480 MHz		5.257 mW	< 1 W


2400 MHz - 2483.5 MHz Band, OQPSK, Low Channel 2405 MHz

Value	Limit	Result
10.72 mW	< 1 W	Pass




2400 MHz - 2483.5 MHz Band, OQPSK, Mid Channel 2445 MHz

Value	Limit	Result
7.73 mW	< 1 W	Pass



2400 MHz - 2483.5 MHz Band, OQPSK, High Channel 2480 MHz

Value	Limit	Result
5.257 mW	< 1 W	Pass



Band Edge Compliance

Testing was performed using the mode(s) of operation and configuration(s) noted within the report. The individuals and/or the organization requesting the test provided the modes, configurations and settings used to complete the evaluation. The actual test parameters are specified in the test data, this includes items such as investigated frequency range (scanned) and test levels. The testing methods and performance specifications, as well as the test site used for the evaluation are indicated in the test data.

TEST EQUIPMENT

Description	Manufacturer	Model	ID	Last Cal.	Interval
Attenuator, 6dB	S.M. Electronics	18N-06	AWN	3/25/2013	12
MXG Vector Signal Generator	Agilent	N5182A	TIF	NCR	0
Power Meter	Gigatronics	8651A	SPM	1/9/2012	24
Power Sensor	Gigatronics	80701A	SPL	7/8/2011	24
Spectrum Analyzer	Agilent	E4440A	AFD	7/5/2012	24
EV06 Direct Connect Cable	ESM Cable Corp.	TT	ECA	NCR	0
Attenuator 20 dB, SMA M/F 26GHz	S.M. Electronics	SA26B-20	AUY	8/2/2012	12
40GHz DC Block	Miteq	DCB4000	AMD	6/25/2012	12

TEST DESCRIPTION

The spurious RF conducted emissions at the edges of the authorized bands were measured with the EUT set to low and high transmit frequencies in each available band. The channels closest to the band edges were selected. The measurement was made using a direct connection between the RF output of the EUT and the spectrum analyzer. The EUT was transmitting at the data rate(s) listed in the datasheet.

The spectrum was scanned below the lower band edge and above the higher band edge.




Band Edge Compliance

XMit 2013.02.28
PsaTx 2013.01.10

EUT: Coordinator		Work Order: CARE0015	
Serial Number: 001D40000031002F		Date: 04/22/13	
Customer: Care Innovations		Temperature: 23°C	
Attendees: Bill Morse		Humidity: 29%	
Project: None		Barometric Pres.: 1032	
Tested by: Brandon Hobbs, Rod Peloquin		Power: 5VDC	
		Job Site: EV06	
TEST SPECIFICATIONS		Test Method	
FCC 15.247:2013		ANSI C63.10:2009	
COMMENTS			
The EUT was operating at 100% duty cycle while under test.			
DEVIATIONS FROM TEST STANDARD			
None			
Configuration #	1	Signature <i>Rodry Le Pellego</i>	
		Value	Limit
2400 MHz - 2483.5 MHz Band			Result
OQPSK			
Low Channel 2405 MHz		-41.44 dBc	≤ -20 dBc
High Channel 2480 MHz		-32.35 dBc	≤ -20 dBc
			Pass
			Pass

2400 MHz - 2483.5 MHz Band, OQPSK, Low Channel 2405 MHz

Value	Limit	Result
-41.44 dBc	≤ -20 dBc	Pass



2400 MHz - 2483.5 MHz Band, OQPSK, High Channel 2480 MHz

Value	Limit	Result
-32.35 dBc	≤ -20 dBc	Pass



Spurious Conducted Emissions

Testing was performed using the mode(s) of operation and configuration(s) noted within the report. The individuals and/or the organization requesting the test provided the modes, configurations and settings used to complete the evaluation. The actual test parameters are specified in the test data, this includes items such as investigated frequency range (scanned) and test levels. The testing methods and performance specifications, as well as the test site used for the evaluation are indicated in the test data.

TEST EQUIPMENT

Description	Manufacturer	Model	ID	Last Cal.	Interval
Attenuator, 6dB	S.M. Electronics	18N-06	AWN	3/25/2013	12
MXG Vector Signal Generator	Agilent	N5182A	TIF	NCR	0
Power Meter	Gigatronics	8651A	SPM	1/9/2012	24
Power Sensor	Gigatronics	80701A	SPL	7/8/2011	24
EV06 Direct Connect Cable	ESM Cable Corp.	TT	ECA	NCR	0
Attenuator 20 dB, SMA M/F 26GHz	S.M. Electronics	SA26B-20	AUY	8/2/2012	12
40GHz DC Block	Miteq	DCB4000	AMD	6/25/2012	12
Spectrum Analyzer	Agilent	E4440A	AFD	7/5/2012	24

TEST DESCRIPTION

The spurious RF conducted emissions were measured with the EUT set to low, medium and high transmit frequencies. The measurements were made using a direct connection between the RF output of the EUT and the spectrum analyzer. The EUT was transmitting at the data rate(s) listed in the datasheet. For each transmit frequency, the spectrum was scanned throughout the specified frequency range.




Spurious Conducted Emissions


XMit 2013.02.28
PsaTx 2013.01.10

EUT: Coordinator		Work Order: CARE0015			
Serial Number: 001D4000031002F		Date: 04/22/13			
Customer: Care Innovations		Temperature: 23°C			
Attendees: Bill Morse		Humidity: 29%			
Project: None		Barometric Pres.: 1032			
Tested by: Brandon Hobbs, Rod Peloquin		Power: 5VDC			
		Job Site: EV06			
TEST SPECIFICATIONS		Test Method			
FCC 15.247:2013		ANSI C63.10:2009			
COMMENTS					
The EUT was operating at 100% duty cycle while under test.					
DEVIATIONS FROM TEST STANDARD					
None					
Configuration #	1	Signature <i>Rodry Le Peloquin</i>			
		Frequency Range	Value	Limit	Result
2400 MHz - 2483.5 MHz Band					
OQPSK					
Low Channel 2405 MHz		Fundamental	N/A	N/A	N/A
Low Channel 2405 MHz		30 MHz - 12.5 GHz	-56.07 dBc	≤ -20 dBc	Pass
Low Channel 2405 MHz		12.5 GHz - 25 GHz	-59.34 dBc	≤ -20 dBc	Pass
Mid Channel 2445 MHz		Fundamental	N/A	N/A	N/A
Mid Channel 2445 MHz		30 MHz - 12.5 GHz	-52.21 dBc	≤ -20 dBc	Pass
Mid Channel 2445 MHz		12.5 GHz - 25 GHz	-57.68 dBc	≤ -20 dBc	Pass
High Channel 2480 MHz		Fundamental	N/A	N/A	N/A
High Channel 2480 MHz		30 MHz - 12.5 GHz	-47.38 dBc	≤ -20 dBc	Pass
High Channel 2480 MHz		12.5 GHz - 25 GHz	-56.29 dBc	≤ -20 dBc	Pass


2400 MHz - 2483.5 MHz Band, OQPSK, Low Channel 2405 MHz				
Frequency Range		Value	Limit	Result
Fundamental		N/A	N/A	N/A




2400 MHz - 2483.5 MHz Band, OQPSK, Low Channel 2405 MHz				
Frequency Range		Value	Limit	Result
30 MHz - 12.5 GHz		-56.07 dBc	≤ -20 dBc	Pass




2400 MHz - 2483.5 MHz Band, OQPSK, Low Channel 2405 MHz				
Frequency Range	Value	Limit	Result	
12.5 GHz - 25 GHz	-59.34 dBc	≤ -20 dBc	Pass	




2400 MHz - 2483.5 MHz Band, OQPSK, Mid Channel 2445 MHz				
Frequency Range	Value	Limit	Result	
Fundamental	N/A	N/A	N/A	




2400 MHz - 2483.5 MHz Band, OQPSK, Mid Channel 2445 MHz				
Frequency Range	Value	Limit	Result	
30 MHz - 12.5 GHz	-52.21 dBc	≤ -20 dBc	Pass	




2400 MHz - 2483.5 MHz Band, OQPSK, Mid Channel 2445 MHz				
Frequency Range	Value	Limit	Result	
12.5 GHz - 25 GHz	-57.68 dBc	≤ -20 dBc	Pass	




2400 MHz - 2483.5 MHz Band, OQPSK, High Channel 2480 MHz				
Frequency Range	Value	Limit	Result	
Fundamental	N/A	N/A	N/A	



2400 MHz - 2483.5 MHz Band, OQPSK, High Channel 2480 MHz				
Frequency Range	Value	Limit	Result	
30 MHz - 12.5 GHz	-47.38 dBc	≤ -20 dBc	Pass	



2400 MHz - 2483.5 MHz Band, OQPSK, High Channel 2480 MHz			
Frequency Range	Value	Limit	Result
12.5 GHz - 25 GHz	-56.29 dBc	≤ -20 dBc	Pass



Power Spectral Density

Testing was performed using the mode(s) of operation and configuration(s) noted within the report. The individuals and/or the organization requesting the test provided the modes, configurations and settings used to complete the evaluation. The actual test parameters are specified in the test data, this includes items such as investigated frequency range (scanned) and test levels. The testing methods and performance specifications, as well as the test site used for the evaluation are indicated in the test data.

TEST EQUIPMENT

Description	Manufacturer	Model	ID	Last Cal.	Interval
Attenuator, 6dB	S.M. Electronics	18N-06	AWN	3/25/2013	12
MXG Vector Signal Generator	Agilent	N5182A	TIF	NCR	0
Power Meter	Gigatronics	8651A	SPM	1/9/2012	24
Power Sensor	Gigatronics	80701A	SPL	7/8/2011	24
EV06 Direct Connect Cable	ESM Cable Corp.	TT	ECA	NCR	0
Attenuator 20 dB, SMA M/F 26GHz	S.M. Electronics	SA26B-20	AUY	8/2/2012	12
40GHz DC Block	Miteq	DCB4000	AMD	6/25/2012	12
Spectrum Analyzer	Agilent	E4440A	AFD	7/5/2012	24

TEST DESCRIPTION

The maximum power spectral density measurements were measured with the EUT set to the required transmit frequencies in each band. The measurement was made using a direct connection between the RF output of the EUT and the spectrum analyzer. The EUT was transmitting at the lowest, middle, and maximum data rate for each modulation type available.

Per the procedure outlined in FCC KDB 558074 D01 DTS Measurement Section 10.2 Method PKPSD (peak PSD), the spectrum analyzer was used as follows:

- RBW = 100 kHz
- VBW = 300 kHz
- Detector = Peak (to match method used for power measurement)
- Trace = Max hold

The observed power level is then scaled to an equivalent value in 3 kHz by adding a Bandwidth Correction Factor (BWCF) where:

$$BWCF = 10 \cdot \log(3 \text{ kHz} / 100 \text{ kHz}) = -15.2 \text{ dB}$$



Power Spectral Density


XMit 2013.02.28
PsaTx 2013.01.10

EUT: Coordinator		Work Order: CARE0015				
Serial Number: 001D40000031002F		Date: 04/22/13				
Customer: Care Innovations		Temperature: 23°C				
Attendees: Bill Morse		Humidity: 29%				
Project: None		Barometric Pres.: 1032				
Tested by: Brandon Hobbs, Rod Peloquin		Power: 5VDC				
		Job Site: EV06				
TEST SPECIFICATIONS		Test Method				
FCC 15.247:2013		ANSI C63.10:2009				
COMMENTS						
The EUT was operating at 100% duty cycle while under test.						
DEVIATIONS FROM TEST STANDARD						
None						
Configuration #	1	Signature <i>Rodry Le Pellego</i>				
		Value dBm/100kHz	dBm/100kHz To dBm/3kHz	Value dBm/3kHz	Limit dBm/3kHz	Result
2400 MHz - 2483.5 MHz Band						
OQPSK						
Low Channel 2405 MHz		6.295	-15.2	-8.905	8	Pass
Mid Channel 2445 MHz		5.082	-15.2	-10.118	8	Pass
High Channel 2480 MHz		3.201	-15.2	-11.999	8	Pass

2400 MHz - 2483.5 MHz Band, OQPSK, Low Channel 2405 MHz						
	Value	dBm/100kHz	dBm/100kHz	Value	Limit	Result
			To dBm/3kHz	dBm/3kHz		
		6.295	-15.2	-8.905	8	Pass



2400 MHz - 2483.5 MHz Band, OQPSK, Mid Channel 2445 MHz						
	Value	dBm/100kHz	dBm/100kHz	Value	Limit	Result
			To dBm/3kHz	dBm/3kHz		
		5.082	-15.2	-10.118	8	Pass



2400 MHz - 2483.5 MHz Band, OQPSK, High Channel 2480 MHz

Value	dBm/100kHz	Value	Limit	Result
dBm/100kHz	To dBm/3kHz	dBm/3kHz	dBm/3kHz	
3.201	-15.2	-11.999	8	Pass



Spurious Radiated Emissions

Testing was performed using the mode(s) of operation and configuration(s) noted within the report. The individuals and/or the organization requesting the test provided the modes, configurations and settings used to complete the evaluation. The actual test parameters are specified in the test data, this includes items such as investigated frequency range (scanned) and test levels. The testing methods and performance specifications, as well as the test site used for the evaluation are indicated in the test data. The test data represents the configuration / operating mode/ model that produced the highest emission levels as compared to the specification limit.

MODES OF OPERATION

On transmitting 802.11 Zigbee

POWER SETTINGS INVESTIGATED

5VDC

CONFIGURATIONS INVESTIGATED

CARE0015 - 1

FREQUENCY RANGE INVESTIGATED

Start Frequency	30 GHz	Stop Frequency	26 GHz
-----------------	--------	----------------	--------

SAMPLE CALCULATIONS

Radiated Emissions: Field Strength = Measured Level + Antenna Factor + Cable Factor - Amplifier Gain + Distance Adjustment Factor + External Attenuation

TEST EQUIPMENT


Description	Manufacturer	Model	ID	Last Cal.	Interval
LP Filter	Micro-Tronics	LPM50004	LFD	7/6/2012	24 mo
Pre-Amplifier	Miteq	AMF-6F-18002650-25-10P	AVU	9/11/2012	12 mo
Antenna, Horn	ETS Lindgren	3160-09	AIV	NCR	0 mo
EV01 Cables	N/A	Standard Gain Horns Cables	EVF	2/27/2013	12 mo
EV01 Cables	N/A	Bilog Cables	EVA	6/26/2012	12 mo
Pre-Amplifier	Miteq	AM-1616-1000	AOL	6/26/2012	12 mo
Antenna, Biconilog	EMCO	3141	AXG	4/10/2012	36 mo
Pre-Amplifier	Miteq	AMF-3D00100800-32-13P	AVF	9/11/2012	12 mo
Pre-Amplifier	Miteq	AMF-6F-08001200-30-10P	AVC	2/27/2013	12 mo
Antenna, Horn	ETS	3160-07	AHU	NCR	0 mo
EV01 Cables	N/A	Double Ridge Horn Cables	EVB	6/27/2012	12 mo
High Pass Filter	Micro-Tronics	HPM50111	HFO	7/6/2012	24 mo
Pre-Amplifier	Miteq	AMF-4D-010100-24-10P	APW	6/27/2012	12 mo
Antenna, Horn	ETS	3115	AIZ	1/24/2011	36 mo
Spectrum Analyzer	Agilent	E4446A	AAQ	2/7/2012	24 mo

MEASUREMENT BANDWIDTHS

Frequency Range (MHz)	Peak Data (kHz)	Quasi-Peak Data (kHz)	Average Data (kHz)
0.01 - 0.15	1.0	0.2	0.2
0.15 - 30.0	10.0	9.0	9.0
30.0 - 1000	100.0	120.0	120.0
Above 1000	1000.0	N/A	1000.0


TEST DESCRIPTION

The highest gain of each type of antenna to be used with the EUT was tested. The EUT was configured for low, mid, and high band transmit frequencies. For each configuration, the spectrum was scanned throughout the specified range. In addition, measurements were made in the restricted bands to verify compliance. While scanning, emissions from the EUT were maximized by rotating the EUT on a turntable, adjusting the position of the EUT and the EUT antenna in three orthogonal axis, and adjusting measurement antenna height and polarization. A preamp and high pass filter were used for this test in order to provide sufficient measurement sensitivity.

Work Order:	CARE0015	Date:	04/18/13		
Project:	None	Temperature:	23.8 °C		
Job Site:	EV01	Humidity:	31.3% RH		
Serial Number:	001D40000031002F	Barometric Pres.:	1032 mbar		
EUT:	Coordinator			Tested by:	Brandon Hobbs
Configuration:	1				
Customer:	Care Innovations				
Attendees:	Bill Morse Stan Telson				
EUT Power:	5VDC				
Operating Mode:	On transmitting 802.11 Zigbee				
Deviations:	None				
Comments:	Please reference the data comments for frequency and EUT orientation. 10Hz video averaging was used for all average test points. The EUT was operating at 100% duty cycle while under test.				


Test Specifications	FCC 15.247:2013	Test Method	ANSI C63.10:2009
---------------------	-----------------	-------------	------------------

Run #	15	Test Distance (m)	3	Antenna Height(s)	1-4m	Results	Pass
-------	----	-------------------	---	-------------------	------	---------	------




Freq (MHz)	Amplitude (dBuV)	Factor (dB)	Antenna Height (meters)	Azimuth (degrees)	Duty Cycle Correction Factor (dB)	External Attenuation (dB)	Polarity/Transducer Type	Detector	Distance Adjustment (dB)	Adjusted (dBuV/m)	Spec. Limit (dBuV/m)	Compared to Spec. (dB)	Comments
7438.697	52.3	19.5	1.9	179.0	0.0	0.0	Horz	PK	0.0	71.8	74.0	-2.2	2480MHz, EUT Vert
7438.533	52.0	19.5	1.2	78.0	0.0	0.0	Horz	PK	0.0	71.5	74.0	-2.5	2480MHz, EUT On Side
7438.553	51.7	19.5	1.9	197.0	0.0	0.0	Vert	PK	0.0	71.2	74.0	-2.8	2480MHz, EUT Vert
7438.588	41.0	19.5	1.9	179.0	-10.8	0.0	Horz	AV	0.0	49.7	54.0	-4.3	(10Hz) 2480MHz, EUT Vert
7438.653	40.8	19.5	1.2	78.0	-10.8	0.0	Horz	AV	0.0	49.5	54.0	-4.5	(10Hz) 2480MHz, EUT On Side
7438.541	40.8	19.5	1.9	197.0	-10.8	0.0	Vert	AV	0.0	49.5	54.0	-4.5	(10Hz) 2480MHz, EUT Vert
7438.733	49.4	19.5	2.2	312.0	0.0	0.0	Horz	PK	0.0	68.9	74.0	-5.1	2480MHz, EUT On End
7333.652	49.7	19.1	1.2	171.0	0.0	0.0	Horz	PK	0.0	68.8	74.0	-5.2	2445MHz, EUT Vert
7438.300	48.9	19.5	3.1	136.0	0.0	0.0	Vert	PK	0.0	68.4	74.0	-5.6	2480MHz, EUT On Side
7438.833	48.2	19.5	1.8	236.0	0.0	0.0	Vert	PK	0.0	67.7	74.0	-6.3	2480MHz, EUT On End
7438.613	38.6	19.5	2.2	312.0	-10.8	0.0	Horz	AV	0.0	47.3	54.0	-6.7	(10Hz) 2480MHz, EUT On End
7333.637	38.9	19.1	1.2	171.0	-10.8	0.0	Horz	AV	0.0	47.2	54.0	-6.8	(10Hz) 2445MHz, EUT Vert
7438.580	37.2	19.5	1.8	236.0	-10.8	0.0	Vert	AV	0.0	45.9	54.0	-8.1	(10Hz) 2480MHz, EUT On End
7333.341	46.8	19.1	1.1	85.0	0.0	0.0	Vert	PK	0.0	65.9	74.0	-8.1	2445MHz, EUT Vert
7438.673	37.1	19.5	1.2	172.0	-10.8	0.0	Vert	AV	0.0	45.8	54.0	-8.2	(10Hz) 2480MHz, EUT On Side
4810.628	45.9	10.2	1.9	157.0	-10.8	0.0	Horz	AV	0.0	45.3	54.0	-8.7	(10Hz) 2405MHz, EUT Vert
4809.083	54.8	10.2	1.9	157.0	0.0	0.0	Horz	PK	0.0	65.0	74.0	-9.0	2405MHz, EUT Vert
4810.620	45.2	10.2	3.2	201.0	-10.8	0.0	Vert	AV	0.0	44.6	54.0	-9.4	(10Hz) 2405MHz, EUT Vert
4808.948	54.3	10.2	3.2	201.0	0.0	0.0	Vert	PK	0.0	64.5	74.0	-9.5	2405MHz, EUT Vert
7333.641	36.0	19.1	1.1	85.0	-10.8	0.0	Vert	AV	0.0	44.3	54.0	-9.7	(10Hz) 2445MHz, EUT Vert
4889.572	43.2	10.5	1.9	162.0	-10.8	0.0	Horz	AV	0.0	42.9	54.0	-11.1	(10Hz) 2445MHz, EUT Vert
4889.199	52.3	10.5	1.9	162.0	0.0	0.0	Horz	PK	0.0	62.8	74.0	-11.2	2445MHz, EUT Vert
4890.779	52.0	10.5	1.0	36.0	0.0	0.0	Vert	PK	0.0	62.5	74.0	-11.5	2445MHz, EUT Vert
4889.543	42.8	10.5	1.0	36.0	-10.8	0.0	Vert	AV	0.0	42.5	54.0	-11.5	(10Hz) 2445MHz, EUT Vert
4958.935	51.1	10.7	1.9	158.0	0.0	0.0	Horz	PK	0.0	61.8	74.0	-12.2	2480MHz, EUT Vert
4959.558	41.7	10.7	1.9	44.0	-10.8	0.0	Vert	AV	0.0	41.6	54.0	-12.4	(10Hz) 2480MHz, EUT Vert
4959.611	41.6	10.7	1.9	158.0	-10.8	0.0	Horz	AV	0.0	41.5	54.0	-12.5	(10Hz) 2480MHz, EUT Vert

Freq (MHz)	Amplitude (dBuV)	Factor (dB)	Antenna Height (meters)	Azimuth (degrees)	Duty Cycle Correction Factor (dB)	External Attenuation (dB)	Polarity/ Transducer Type	Detector	Distance Adjustment (dB)	Adjusted (dBuV/m)	Spec. Limit (dBuV/m)	Compared to Spec. (dB)	Comments
4958.904	50.8	10.7	1.9	44.0	0.0	0.0	Vert	PK	0.0	61.5	74.0	-12.5	2480MHz, EUT Vert
14428.510	39.7	17.9	1.4	216.0	0.0	0.0	Vert	PK	0.0	57.6	74.0	-16.4	2480MHz, EUT Vert
14431.840	39.3	17.9	1.0	41.0	0.0	0.0	Horz	PK	0.0	57.2	74.0	-16.8	2480MHz, EUT Vert
14429.210	26.3	17.9	1.4	216.0	-10.8	0.0	Vert	AV	0.0	33.4	54.0	-20.6	(10Hz) 2405MHz, EUT Vert
14430.530	26.2	17.9	1.0	41.0	-10.8	0.0	Horz	AV	0.0	33.3	54.0	-20.7	(10Hz) 2405MHz, EUT Vert
12225.230	43.5	-3.8	1.0	331.0	0.0	0.0	Vert	PK	0.0	39.7	74.0	-34.3	2445MHz, EUT Vert
12223.130	43.4	-3.8	1.0	249.0	0.0	0.0	Horz	PK	0.0	39.6	74.0	-34.4	2445MHz, EUT Vert
12025.700	43.8	-4.2	1.0	33.0	0.0	0.0	Vert	PK	0.0	39.6	74.0	-34.4	2405MHz, EUT Vert
12025.130	43.6	-4.2	1.0	212.0	0.0	0.0	Horz	PK	0.0	39.4	74.0	-34.6	2405MHz, EUT Vert
12398.410	42.3	-3.4	1.0	293.0	0.0	0.0	Horz	PK	0.0	38.9	74.0	-35.1	2480MHz, EUT Vert
12400.340	41.9	-3.4	1.0	260.0	0.0	0.0	Vert	PK	0.0	38.5	74.0	-35.5	2480MHz, EUT Vert
12024.080	31.0	-4.2	1.0	212.0	-10.8	0.0	Horz	AV	0.0	16.0	54.0	-38.0	(10Hz) 2405MHz, EUT Vert
12023.850	31.0	-4.2	1.0	33.0	-10.8	0.0	Vert	AV	0.0	16.0	54.0	-38.0	(10Hz) 2405MHz, EUT Vert
12223.010	30.4	-3.8	1.0	331.0	-10.8	0.0	Vert	AV	0.0	15.8	54.0	-38.2	(10Hz) 2445MHz, EUT Vert
12223.230	30.3	-3.8	1.0	249.0	-10.8	0.0	Horz	AV	0.0	15.7	54.0	-38.3	(10Hz) 2445MHz, EUT Vert
12398.000	29.5	-3.4	1.0	260.0	-10.8	0.0	Vert	AV	0.0	15.3	54.0	-38.7	(10Hz) 2480MHz, EUT Vert
12398.000	29.5	-3.4	1.0	293.0	-10.8	0.0	Horz	AV	0.0	15.3	54.0	-38.7	(10Hz) 2480MHz, EUT Vert

Work Order:	CARE0015	Date:	04/18/13	
Project:	None	Temperature:	23.8 °C	
Job Site:	EV01	Humidity:	31.3% RH	
Serial Number:	001D40000031002F	Barometric Pres.:	1032 mbar	Tested by: Brandon Hobbs
EUT:	Coordinator			
Configuration:	1			
Customer:	Care Innovations			
Attendees:	Bill Morse Stan Telson			
EUT Power:	5VDC			
Operating Mode:	On transmitting 802.11 Zigbee			
Deviations:	None			
Comments:	Please reference the data comments for frequency and EUT orientation. 10Hz video averaging was used for all average test points. The EUT was operating at 100% duty cycle while under test.			



Test Specifications	Test Method
FCC 15.247:2013	ANSI C63.10:2009

Run #	17	Test Distance (m)	3	Antenna Height(s)	1-4m	Results	Pass
-------	----	-------------------	---	-------------------	------	---------	------




Freq (MHz)	Amplitude (dBuV)	Factor (dB)	Antenna Height (meters)	Azimuth (degrees)	Duty Cycle Correction Factor (dB)	External Attenuation (dB)	Polarity/Transducer Type	Detector	Distance Adjustment (dB)	Adjusted (dBuV/m)	Spec. Limit (dBuV/m)	Compared to Spec. (dB)	Comments
2483.500	42.2	1.9	1.2	195.0	-10.8	20.0	Vert	AV	0.0	53.3	54.0	-0.7	Band Edge (10Hz), EUT Vert
2483.500	41.5	1.9	1.0	270.0	-10.8	20.0	Horz	AV	0.0	52.6	54.0	-1.4	Band Edge (10Hz), EUT On Side
2483.500	41.1	1.9	1.0	268.0	-10.8	20.0	Horz	AV	0.0	52.2	54.0	-1.8	Band Edge (10Hz), EUT On End
2483.500	39.9	1.9	1.0	175.0	-10.8	20.0	Horz	AV	0.0	51.0	54.0	-3.0	Band Edge (10Hz), EUT Vert
2483.540	47.5	1.9	2.0	205.0	0.0	20.0	Vert	PK	0.0	69.4	74.0	-4.6	Band Edge, EUT On Side
2483.620	45.5	1.9	1.0	213.0	0.0	20.0	Vert	PK	0.0	67.4	74.0	-6.6	Band Edge, EUT On End
2483.500	35.6	1.9	2.0	205.0	-10.8	20.0	Vert	AV	0.0	46.7	54.0	-7.3	Band Edge (10Hz), EUT On Side
2483.843	43.0	1.9	1.0	175.0	0.0	20.0	Horz	PK	0.0	64.9	74.0	-9.1	Band Edge, EUT Vert
2483.500	33.7	1.9	1.0	213.0	-10.8	20.0	Vert	AV	0.0	44.8	54.0	-9.2	Band Edge (10Hz), EUT On End
2388.623	37.3	1.6	2.6	77.0	0.0	20.0	Vert	PK	0.0	58.9	74.0	-15.1	Band Edge EUT Vert
2388.017	37.1	1.6	1.0	181.0	0.0	20.0	Horz	PK	0.0	58.7	74.0	-15.3	Band Edge EUT Horz
2388.960	23.1	1.5	1.0	181.0	-10.8	20.0	Horz	AV	0.0	33.8	54.0	-20.2	Band Edge (10Hz)EUT Horz
2389.790	23.1	1.5	2.6	77.0	-10.8	20.0	Vert	AV	0.0	33.8	54.0	-20.2	Band Edge (10Hz)EUT Vert
2483.517	53.7	1.9	1.2	195.0	0.0	20.0	Vert	PK	0.0	102.7			Fundamental, EUT Vert, Marker Delta Method
2483.500				195.0	0.0	20.0	Vert	PK	0.0	73.3	74.0	-0.7	102.7 - 29.39 = 73.31 dB
2483.563	53.0	1.9	1.0	270.0	0.0	20.0	Horz	PK	0.0	101.8			Fundamental, EUT On Side, Marker Delta Method
2483.500			1.0	270.0	0.0	20.0	Horz	PK	0.0	71.2	74.0	-2.8	101.8 - 30.63 = 71.17 dB
2483.560	52.8	1.9	1.0	268.0	0.0	20.0	Horz	PK	0.0	100.0			Fundamental, EUT On End, Marker Delta Method
2483.500			1.0	268.0	0.0	20.0	Horz	PK	0.0	68.8	74.0	-5.2	100 - 31.22 = 68.78 dB

Spurious Radiated Emissions



Spurious Radiated Emissions



AC POWERLINE CONDUCTED EMISSIONS

TEST DESCRIPTION

Using the mode of operation and configuration noted within this report, conducted emissions tests were performed. The frequency range investigated (scanned), is also noted in this report. Conducted power line measurements are made, unless otherwise specified, over the frequency range from 150 kHz to 30 MHz to determine the line-to-ground radio-noise voltage that is conducted from the EUT power-input terminals that are directly (or indirectly via separate transformer or power supplies) connected to a public power network. Equipment is tested with power cords that are normally used or that have electrical or shielding characteristics that are the same as those cords normally used. Typically those measurements are made using a LISN (Line Impedance Stabilization Network), the 50 Ω measuring port is terminated by a 50 Ω EMI meter or a 50 Ω resistive load. All 50 Ω measuring ports of the LISN are terminated by 50 Ω .

TEST EQUIPMENT

Description	Manufacturer	Model	ID	Last Cal.	Interval
Receiver	Rohde & Schwarz	ESCI	ARH	01/24/2013	12 mo
Attenuator	Coaxicom	66702 2910-20	RBR	04/25/2013	12 mo
High Pass Filter	TTE	H97-100K-50-720B	HHD	02/01/2012	24 mo
LISN	Solar	9252-50-R-24-BNC	LIN	03/11/2013	12 mo
LISN	Solar	9252-50-R-24-BNC	LIP	04/08/2013	12 mo

MEASUREMENT UNCERTAINTY

Description		
Expanded k=2	2.94 dB	-2.94 dB

CONFIGURATIONS INVESTIGATED

CARE0016-2

MODES INVESTIGATED

Low Channel 2405 MHz
Mid Channel 244 MHz
High Channel 2480 MHz

AC POWERLINE CONDUCTED EMISSIONS

EUT:	Quiet Care 3.0 Coordinator	Work Order:	CARE0016
Serial Number:	001D40000031002F	Date:	05/01/2013
Customer:	Care Innovations	Temperature:	22.4°C
Attendees:	Bill Morse	Relative Humidity:	28.2%
Customer Project:	None	Bar. Pressure:	1033 mb
Tested By:	Cole Ghizzone	Job Site:	EV07
Power:	110VAC/60Hz	Configuration:	CARE0016-2

TEST SPECIFICATIONS

Specification:	Method:
FCC 15.207:2013	ANSI C63.10:2009

TEST PARAMETERS

Run #:	23	Line:	Neutral	Ext. Attenuation (dB):	20
--------	----	-------	---------	------------------------	----

COMMENTS

Coordinator plugged into Facility Server, Facility Server connected to laptop and UPS, transmitting, large ferrites added to laptop USB, ferrite added to AC power cable on Facility Server's power supply.


EUT OPERATING MODES

Low Channel 2405 MHz


DEVIATIONS FROM TEST STANDARD

None

Quasi Peak Data - vs - Quasi Peak Limit



Average Data - vs - Average Limit



AC POWERLINE CONDUCTED EMISSIONS

RESULTS - Run #23

Quasi Peak Data - vs - Quasi Peak Limit

Freq (MHz)	Amp. (dBuV)	Factor (dB)	Adjusted (dBuV)	Spec. Limit (dBuV)	Margin (dB)
14.056	31.7	21.1	52.8	60.0	-7.2
14.120	31.2	21.1	52.3	60.0	-7.7
13.996	31.2	21.1	52.3	60.0	-7.7
13.742	31.0	21.1	52.1	60.0	-7.9
14.306	30.8	21.2	52.0	60.0	-8.0
14.182	30.8	21.1	51.9	60.0	-8.1
13.680	30.5	21.1	51.6	60.0	-8.4
14.430	30.2	21.2	51.4	60.0	-8.6
13.870	30.0	21.1	51.1	60.0	-8.9
13.556	30.0	21.1	51.1	60.0	-8.9
14.492	29.9	21.2	51.1	60.0	-8.9
14.620	29.3	21.2	50.5	60.0	-9.5
14.366	29.3	21.2	50.5	60.0	-9.5
13.432	28.8	21.1	49.9	60.0	-10.1
14.742	28.6	21.2	49.8	60.0	-10.2
14.928	27.8	21.2	49.0	60.0	-11.0
15.054	27.3	21.2	48.5	60.0	-11.5
13.116	24.9	21.1	46.0	60.0	-14.0
29.232	21.2	22.0	43.2	60.0	-16.8

Average Data - vs - Average Limit

Freq (MHz)	Amp. (dBuV)	Factor (dB)	Adjusted (dBuV)	Spec. Limit (dBuV)	Margin (dB)
14.056	26.0	21.1	47.1	50.0	-2.9
14.306	25.6	21.2	46.8	50.0	-3.2
14.120	25.3	21.1	46.4	50.0	-3.6
14.182	25.2	21.1	46.3	50.0	-3.7
13.996	25.2	21.1	46.3	50.0	-3.7
14.430	25.0	21.2	46.2	50.0	-3.8
13.742	24.9	21.1	46.0	50.0	-4.0
14.492	24.8	21.2	46.0	50.0	-4.0
13.680	24.5	21.1	45.6	50.0	-4.4
13.870	24.1	21.1	45.2	50.0	-4.8
14.620	24.0	21.2	45.2	50.0	-4.8
14.366	23.7	21.2	44.9	50.0	-5.1
13.556	23.7	21.1	44.8	50.0	-5.2
14.742	23.6	21.2	44.8	50.0	-5.2
14.928	22.7	21.2	43.9	50.0	-6.1
15.054	22.4	21.2	43.6	50.0	-6.4
13.432	22.3	21.1	43.4	50.0	-6.6
13.116	17.7	21.1	38.8	50.0	-11.2
29.232	-0.2	22.0	21.8	50.0	-28.2

CONCLUSION

Pass



Tested By

AC POWERLINE CONDUCTED EMISSIONS

EUT:	Quiet Care 3.0 Coordinator	Work Order:	CARE0016
Serial Number:	001D40000031002F	Date:	05/01/2013
Customer:	Care Innovations	Temperature:	22.4°C
Attendees:	Bill Morse	Relative Humidity:	28.2%
Customer Project:	None	Bar. Pressure:	1033 mb
Tested By:	Cole Ghizzone	Job Site:	EV07
Power:	110VAC/60Hz	Configuration:	CARE0016-2

TEST SPECIFICATIONS

Specification:	Method:
FCC 15.207:2013	ANSI C63.10:2009

TEST PARAMETERS

Run #:	24	Line:	High Line	Ext. Attenuation (dB):	20
--------	----	-------	-----------	------------------------	----

COMMENTS

Coordinator plugged into Facility Server, Facility Server connected to laptop and UPS, transmitting, large ferrites added to laptop USB, ferrite added to AC power cable on Facility Server's power supply.


EUT OPERATING MODES

Low Channel 2405 MHz


DEVIATIONS FROM TEST STANDARD

None

Quasi Peak Data - vs - Quasi Peak Limit



Average Data - vs - Average Limit



AC POWERLINE CONDUCTED EMISSIONS

RESULTS - Run #24

Quasi Peak Data - vs - Quasi Peak Limit

Freq (MHz)	Amp. (dBuV)	Factor (dB)	Adjusted (dBuV)	Spec. Limit (dBuV)	Margin (dB)
13.992	31.3	21.1	52.4	60.0	-7.6
14.116	31.1	21.1	52.2	60.0	-7.8
14.056	31.0	21.1	52.1	60.0	-7.9
13.866	31.0	21.1	52.1	60.0	-7.9
13.742	31.0	21.1	52.1	60.0	-7.9
14.178	30.8	21.1	51.9	60.0	-8.1
13.680	30.5	21.1	51.6	60.0	-8.4
14.366	30.3	21.2	51.5	60.0	-8.5
14.306	30.3	21.2	51.5	60.0	-8.5
14.430	30.1	21.2	51.3	60.0	-8.7
13.616	30.1	21.1	51.2	60.0	-8.8
14.492	29.7	21.2	50.9	60.0	-9.1
13.556	29.7	21.1	50.8	60.0	-9.2
14.616	29.0	21.2	50.2	60.0	-9.8
14.742	28.6	21.2	49.8	60.0	-10.2
14.866	28.0	21.2	49.2	60.0	-10.8

Average Data - vs - Average Limit

Freq (MHz)	Amp. (dBuV)	Factor (dB)	Adjusted (dBuV)	Spec. Limit (dBuV)	Margin (dB)
13.992	25.7	21.1	46.8	50.0	-3.2
14.116	25.6	21.1	46.7	50.0	-3.3
14.178	25.4	21.1	46.5	50.0	-3.5
13.866	25.2	21.1	46.3	50.0	-3.7
14.366	25.1	21.2	46.3	50.0	-3.7
14.056	25.0	21.1	46.1	50.0	-3.9
13.742	25.0	21.1	46.1	50.0	-3.9
14.306	24.9	21.2	46.1	50.0	-3.9
14.430	24.8	21.2	46.0	50.0	-4.0
14.492	24.6	21.2	45.8	50.0	-4.2
13.680	24.5	21.1	45.6	50.0	-4.4
14.616	23.9	21.2	45.1	50.0	-4.9
13.616	23.8	21.1	44.9	50.0	-5.1
14.742	23.6	21.2	44.8	50.0	-5.2
13.556	23.2	21.1	44.3	50.0	-5.7
14.866	23.1	21.2	44.3	50.0	-5.7

CONCLUSION

Pass



Tested By

AC POWERLINE CONDUCTED EMISSIONS

EUT:	Quiet Care 3.0 Coordinator	Work Order:	CARE0016
Serial Number:	001D40000031002F	Date:	05/01/2013
Customer:	Care Innovations	Temperature:	22.4°C
Attendees:	Bill Morse	Relative Humidity:	28.2%
Customer Project:	None	Bar. Pressure:	1033 mb
Tested By:	Cole Ghizzone	Job Site:	EV07
Power:	110VAC/60Hz	Configuration:	CARE0016-2

TEST SPECIFICATIONS

Specification:	Method:
FCC 15.207:2013	ANSI C63.10:2009

TEST PARAMETERS

Run #:	25	Line:	High Line	Ext. Attenuation (dB):	20
--------	----	-------	-----------	------------------------	----

COMMENTS

Coordinator plugged into Facility Server, Facility Server connected to laptop and UPS, transmitting, large ferrites added to laptop USB, ferrite added to AC power cable on Facility Server's power supply.


EUT OPERATING MODES

Mid Channel 2445 MHz


DEVIATIONS FROM TEST STANDARD

None

Quasi Peak Data - vs - Quasi Peak Limit



Average Data - vs - Average Limit



AC POWERLINE CONDUCTED EMISSIONS

RESULTS - Run #25

Quasi Peak Data - vs - Quasi Peak Limit

Freq (MHz)	Amp. (dBuV)	Factor (dB)	Adjusted (dBuV)	Spec. Limit (dBuV)	Margin (dB)
14.116	31.5	21.1	52.6	60.0	-7.4
13.992	31.5	21.1	52.6	60.0	-7.4
14.240	31.0	21.2	52.2	60.0	-7.8
13.802	30.9	21.1	52.0	60.0	-8.0
14.302	30.6	21.2	51.8	60.0	-8.2
13.680	30.5	21.1	51.6	60.0	-8.4
13.366	27.8	21.1	48.9	60.0	-11.1
29.234	24.9	22.0	46.9	60.0	-13.1

Average Data - vs - Average Limit

Freq (MHz)	Amp. (dBuV)	Factor (dB)	Adjusted (dBuV)	Spec. Limit (dBuV)	Margin (dB)
14.116	26.0	21.1	47.1	50.0	-2.9
13.992	25.9	21.1	47.0	50.0	-3.0
14.240	25.6	21.2	46.8	50.0	-3.2
14.302	25.1	21.2	46.3	50.0	-3.7
13.802	24.7	21.1	45.8	50.0	-4.2
13.680	24.4	21.1	45.5	50.0	-4.5
13.366	21.4	21.1	42.5	50.0	-7.5
29.234	3.9	22.0	25.9	50.0	-24.1

CONCLUSION

Pass

Tested By

AC POWERLINE CONDUCTED EMISSIONS

EUT:	Quiet Care 3.0 Coordinator	Work Order:	CARE0016
Serial Number:	001D40000031002F	Date:	05/01/2013
Customer:	Care Innovations	Temperature:	22.4°C
Attendees:	Bill Morse	Relative Humidity:	28.2%
Customer Project:	None	Bar. Pressure:	1033 mb
Tested By:	Cole Ghizzone	Job Site:	EV07
Power:	110VAC/60Hz	Configuration:	CARE0016-2

TEST SPECIFICATIONS

Specification:	Method:
FCC 15.207:2013	ANSI C63.10:2009

TEST PARAMETERS

Run #:	26	Line:	Neutral	Ext. Attenuation (dB):	20
--------	----	-------	---------	------------------------	----

COMMENTS

Coordinator plugged into Facility Server, Facility Server connected to laptop and UPS, transmitting, large ferrites added to laptop USB, ferrite added to AC power cable on Facility Server's power supply.


EUT OPERATING MODES

Mid Channel 2445 MHz


DEVIATIONS FROM TEST STANDARD

None

Quasi Peak Data - vs - Quasi Peak Limit



Average Data - vs - Average Limit



AC POWERLINE CONDUCTED EMISSIONS

RESULTS - Run #26

Quasi Peak Data - vs - Quasi Peak Limit

Freq (MHz)	Amp. (dBuV)	Factor (dB)	Adjusted (dBuV)	Spec. Limit (dBuV)	Margin (dB)
13.930	31.6	21.1	52.7	60.0	-7.3
14.052	31.4	21.1	52.5	60.0	-7.5
14.180	31.3	21.1	52.4	60.0	-7.6
13.740	31.0	21.1	52.1	60.0	-7.9
14.366	30.9	21.2	52.1	60.0	-7.9
14.492	30.3	21.2	51.5	60.0	-8.5
14.802	28.9	21.2	50.1	60.0	-9.9
29.658	26.0	22.1	48.1	60.0	-11.9

Average Data - vs - Average Limit

Freq (MHz)	Amp. (dBuV)	Factor (dB)	Adjusted (dBuV)	Spec. Limit (dBuV)	Margin (dB)
14.366	25.8	21.2	47.0	50.0	-3.0
13.930	25.8	21.1	46.9	50.0	-3.1
14.052	25.7	21.1	46.8	50.0	-3.2
14.180	25.6	21.1	46.7	50.0	-3.3
14.492	25.0	21.2	46.2	50.0	-3.8
13.740	24.8	21.1	45.9	50.0	-4.1
14.802	23.8	21.2	45.0	50.0	-5.0
29.658	8.4	22.1	30.5	50.0	-19.5

CONCLUSION

Pass

Tested By

AC POWERLINE CONDUCTED EMISSIONS

EUT:	Quiet Care 3.0 Coordinator	Work Order:	CARE0016
Serial Number:	001D40000031002F	Date:	05/01/2013
Customer:	Care Innovations	Temperature:	22.4°C
Attendees:	Bill Morse	Relative Humidity:	28.2%
Customer Project:	None	Bar. Pressure:	1033 mb
Tested By:	Cole Ghizzone	Job Site:	EV07
Power:	110VAC/60Hz	Configuration:	CARE0016-2

TEST SPECIFICATIONS

Specification:	Method:
FCC 15.207:2013	ANSI C63.10:2009

TEST PARAMETERS

Run #:	27	Line:	Neutral	Ext. Attenuation (dB):	20
--------	----	-------	---------	------------------------	----

COMMENTS

Coordinator plugged into Facility Server, Facility Server connected to laptop and UPS, transmitting, large ferrites added to laptop USB, ferrite added to AC power cable on Facility Server's power supply.


EUT OPERATING MODES

High Channel 2480 MHz


DEVIATIONS FROM TEST STANDARD

None

Quasi Peak Data - vs - Quasi Peak Limit



Average Data - vs - Average Limit



AC POWERLINE CONDUCTED EMISSIONS

RESULTS - Run #27

Quasi Peak Data - vs - Quasi Peak Limit

Freq (MHz)	Amp. (dBuV)	Factor (dB)	Adjusted (dBuV)	Spec. Limit (dBuV)	Margin (dB)
13.992	31.9	21.1	53.0	60.0	-7.0
14.116	31.7	21.1	52.8	60.0	-7.2
14.302	31.2	21.2	52.4	60.0	-7.6
13.680	30.9	21.1	52.0	60.0	-8.0
14.430	30.5	21.2	51.7	60.0	-8.3
13.556	29.7	21.1	50.8	60.0	-9.2
14.740	29.3	21.2	50.5	60.0	-9.5
29.656	26.0	22.1	48.1	60.0	-11.9

Average Data - vs - Average Limit

Freq (MHz)	Amp. (dBuV)	Factor (dB)	Adjusted (dBuV)	Spec. Limit (dBuV)	Margin (dB)
13.992	26.3	21.1	47.4	50.0	-2.6
14.116	26.2	21.1	47.3	50.0	-2.7
14.302	25.8	21.2	47.0	50.0	-3.0
14.430	25.0	21.2	46.2	50.0	-3.8
13.680	24.7	21.1	45.8	50.0	-4.2
14.740	24.2	21.2	45.4	50.0	-4.6
13.556	23.1	21.1	44.2	50.0	-5.8
29.656	8.5	22.1	30.6	50.0	-19.4

CONCLUSION

Pass



Tested By

AC POWERLINE CONDUCTED EMISSIONS

EUT:	Quiet Care 3.0 Coordinator	Work Order:	CARE0016
Serial Number:	001D40000031002F	Date:	05/01/2013
Customer:	Care Innovations	Temperature:	22.4°C
Attendees:	Bill Morse	Relative Humidity:	28.2%
Customer Project:	None	Bar. Pressure:	1033 mb
Tested By:	Cole Ghizzone	Job Site:	EV07
Power:	110VAC/60Hz	Configuration:	CARE0016-2

TEST SPECIFICATIONS

Specification:	Method:
FCC 15.207:2013	ANSI C63.10:2009

TEST PARAMETERS

Run #:	28	Line:	High Line	Ext. Attenuation (dB):	20
--------	----	-------	-----------	------------------------	----

COMMENTS

Coordinator plugged into Facility Server, Facility Server connected to laptop and UPS, transmitting, large ferrites added to laptop USB, ferrite added to AC power cable on Facility Server's power supply.


EUT OPERATING MODES

High Channel 2480 MHz


DEVIATIONS FROM TEST STANDARD

None

Quasi Peak Data - vs - Quasi Peak Limit



Average Data - vs - Average Limit



AC POWERLINE CONDUCTED EMISSIONS

RESULTS - Run #28

Quasi Peak Data - vs - Quasi Peak Limit

Freq (MHz)	Amp. (dBuV)	Factor (dB)	Adjusted (dBuV)	Spec. Limit (dBuV)	Margin (dB)
14.054	31.6	21.1	52.7	60.0	-7.3
13.930	31.4	21.1	52.5	60.0	-7.5
13.740	31.1	21.1	52.2	60.0	-7.8
14.240	31.0	21.2	52.2	60.0	-7.8
14.116	31.0	21.1	52.1	60.0	-7.9
14.366	30.8	21.2	52.0	60.0	-8.0
13.802	30.8	21.1	51.9	60.0	-8.1
13.866	30.5	21.1	51.6	60.0	-8.4
13.616	30.5	21.1	51.6	60.0	-8.4
14.490	30.1	21.2	51.3	60.0	-8.7
14.552	29.9	21.2	51.1	60.0	-8.9
14.180	29.7	21.1	50.8	60.0	-9.2
13.490	29.5	21.1	50.6	60.0	-9.4
14.680	29.1	21.2	50.3	60.0	-9.7
14.802	28.6	21.2	49.8	60.0	-10.2
15.114	27.4	21.2	48.6	60.0	-11.4
15.240	26.6	21.2	47.8	60.0	-12.2
13.304	26.4	21.1	47.5	60.0	-12.5
15.364	24.6	21.2	45.8	60.0	-14.2
29.038	23.5	22.0	45.5	60.0	-14.5
14.934	-0.9	21.2	20.3	60.0	-39.7

Average Data - vs - Average Limit

Freq (MHz)	Amp. (dBuV)	Factor (dB)	Adjusted (dBuV)	Spec. Limit (dBuV)	Margin (dB)
14.054	26.0	21.1	47.1	50.0	-2.9
14.240	25.7	21.2	46.9	50.0	-3.1
14.116	25.6	21.1	46.7	50.0	-3.3
13.930	25.5	21.1	46.6	50.0	-3.4
14.366	25.4	21.2	46.6	50.0	-3.4
13.740	25.0	21.1	46.1	50.0	-3.9
14.490	24.8	21.2	46.0	50.0	-4.0
13.866	24.8	21.1	45.9	50.0	-4.1
14.552	24.6	21.2	45.8	50.0	-4.2
13.802	24.6	21.1	45.7	50.0	-4.3
13.616	24.3	21.1	45.4	50.0	-4.6
14.180	24.1	21.1	45.2	50.0	-4.8
14.680	23.6	21.2	44.8	50.0	-5.2
14.802	23.4	21.2	44.6	50.0	-5.4
13.490	22.9	21.1	44.0	50.0	-6.0
15.114	22.2	21.2	43.4	50.0	-6.6
15.240	21.6	21.2	42.8	50.0	-7.2
13.304	19.7	21.1	40.8	50.0	-9.2
15.364	19.6	21.2	40.8	50.0	-9.2
29.038	1.4	22.0	23.4	50.0	-26.6
14.934	-5.4	21.2	15.8	50.0	-34.2

CONCLUSION

Pass



Tested By