



603 Queensbury Ave.
Queensbury, NY 12804 USA
Phone 800-772-6446
Fax 518-798-3625

Model: **HVP01**

SN

CONFORMS TO UL STD 60601-1
CERTIFIED TO CSA STD C22.2 NO. 601.1

THIS DEVICE CONTAINS AN RFID
DEVICE WITH FCC ID: YHS-600-104443
IC:4690B-600104443

Power Supply:
100-230V~ 50/60HZ 420 Watt

Fuse Type:
2 fuses 5 A 250 V - 5 x 20mm

CAUTION: FOR CONTINUED PROTECTION AGAINST RISK OF FIRE REPLACE
ONLY WITH THE SAME TYPE AND RATINGS OF FUSES

ATTENTION: POUR NE PAS COMPROMETTRE LA PROTECTION CONTRE LES
RISQUES D'INCENDIE, REPLACER PAR UN FUSIBLE DE MEME TYPE ET
MEMES CARACTERISTIQUES NOMINALES

THIS DEVICE DOES NOT CONTAIN USER SERVICEABLE PARTS.

**WARNING: HIGH VOLTAGE. ELECTROCUTION HAZARD! SERVICING MUST
BE DONE BY QUALIFIED PERSONNEL ONLY.**

WARNING: DO NOT USE IN THE PRESENCE OF FLAMMABLE GASSES



ETL CLASSIFIED



Intertek
3176792

