

**MERCURY INSTRUMENTS
MODEL CNI2**

FCC ID: YIFCNI2G24

TEMP CODE T3C

**FOR USE IN HAZARDOUS LOCATIONS
CLASS I, DIV 2, GROUPS C & D**

**CAUTION: BONDING BETWEEN CONDUIT
CONNECTIONS IS NOT AUTOMATIC AND
MUST BE PROVIDED AS A PART OF THE
INSTALLATION.**

**WARNING: SUBSTITUTION OF COMPONENTS
MAY IMPAIR SUITABILITY FOR CLASS I, DIV 2**

**USE ONLY BATTERY PACK NO.
40-3503, 40-3444, 40-5170, 40-2596**

