


KAP Medical Auto-fowler transmitter label



NOTES: UNLESS OTHERWISE SPECIFIED

1. MATERIAL NOTE:
5 MIL POLYESTER, BRIGHT SILVER, WITH
PRESSURE-SENSITIVE ACRYLIC ADHESIVE,
2 MIL CLEAR/GLOSS POLYESTER.
2. 2 COLOR SILK-SCREEN (GOLD + BLACK)
THERMAL PRINT SERIAL NUMBER

REVISIONS			
ECO	REV	DESCRIPTION	DATE
	NEW	Initial Release	
08-0003	A	KAP LOGO DELETED	1/15/08
09-0028	B	ADDED DEGREE BOX OPTION	11/13/09
10-0208	C	UPDATE TITLE BLOCK, TEXT & GRAPHIC	2010/11/24



APPROVALS DRAWN: <i>[Signature]</i> CHECKED: <i>[Signature]</i> ENGINEERING: <i>[Signature]</i> QUALITY: <i>[Signature]</i>		DATE 2010/11/24 2010/11/29	UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES TOLERANCES ON: .X DECIMAL .06 .X DECIMAL .03 .XX DECIMAL .015 .XXX DECIMAL .005 ANGLES .35 SURFACE FINISH 125/ BREAK SHARP EDGES TO .015 R OR CHAMFER FILLET R .015 MAX	KAP MEDICAL 1395 Pico Street Corona, CA 92881 Ph: 951 340 4360 Fax: 951 340 4361
TITLE: LABEL, AUTO FOWLER SENSOR, K-150		SIZE: A		PART NUMBER: 400194
SCALE: NTS		SHEET 1 of 2		REV: C

PROPRIETARY NOTICE

THE DATA CONTAINED HEREIN CONSTITUTES PROPRIETARY INFORMATION OF KAP MEDICAL, AND SHALL NOT BE USED, DUPLICATED OR DISCLOSED IN WHOLE OR IN PART WITHOUT THE PRIOR WRITTEN AUTHORIZATION OF KAP MEDICAL.

This device complies with part 15 of the FCC rules. Operation is subject to the following two conditions:

- (1) This device may not cause harmful interference, and
- (2) This device must accept harmful interference received, including interference that may cause undesired operation.

P/N 400863 Rev. New

NOTES: UNLESS OTHERWISE SPECIFIED

1. MATERIAL NOTE:
5 MIL POLYESTER, BRIGHT SILVER, WITH
PRESSURE-SENSITIVE ACRYLIC ADHESIVE-(3M7941)
2. 2 MIL CLEAR/GLOSS POLYESTER.
2 COLOR SILK-SCREEN (SILVER + BLACK)
THERMAL PRINT SERIAL NUMBER
3. MUST BE ROHS COMPLIANT

REVISIONS				DATE
ECO	REV	DESCRIPTION		
15-0238	NEW	INITIAL RELEASE		

This device complies with part 15 of the FCC rules. Operation is subject to the following two conditions:
 (1) This device may not cause harmful interference, and
 (2) This device must accept harmful interference received, including interference that may cause undesired operation.

R.125 TYP.

PRELIMINARY

Handwritten signature and date: 6/9/15

APPROVALS		DATE	
DESIGNED			
CHECKED			
DRAWN			
SURFACE FINISH 12 $\frac{1}{2}$ /		SURFACE SHARP EDGES TO .015 R OR CHAMFER	
FILED 8.013 MAX			
TITLE		1395 PICO STREET CORONA, CA 92601 P/N 400863	
LABEL, AUTOFOWLER, FCC RULES			
SIZE	PRINT NUMBER	NEW	
A	400863		
SCALE	SHEET	1 OF 1	
2:1			

PROPRIETARY NOTICE
 THE DATA CONTAINED HEREIN IS THE PROPERTY OF AUTOFOWLER, INC. AND IS NOT TO BE REPRODUCED OR TRANSMITTED IN ANY FORM OR BY ANY MEANS, ELECTRONIC OR MECHANICAL, INCLUDING PHOTOCOPYING, RECORDING, OR BY ANY INFORMATION STORAGE AND RETRIEVAL SYSTEM, WITHOUT THE WRITTEN PERMISSION OF AUTOFOWLER, INC.