

Test Equipment List with Calibration Data

Seite 1 von 1

Test report number(s):
5010010752-02418

Date of test:
11/2010

Type	Inv.-No.	Type Designation	Serial Number	Manufacturer	Calibration Organization	Date of Calibration	
						Last	Next
EMI test receiver	1028	ESHS10	860043/016	Rohde & Schwarz	Rohde & Schwarz	10/2010	04/2012
EMI test receiver	1569	ESMI	839379/013	Rohde & Schwarz	Rohde & Schwarz	10/2009	01/2012
EMI test receiver	2044	ESU8	100232	Rohde & Schwarz	Rohde & Schwarz	05/2009	11/2010
Spectrum analyser	1666	FSP30	100063	Rohde & Schwarz	Rohde & Schwarz	11/2009	05/2011
Preamplifier	1484	ACO/180-3530	32641	CTT	TÜV SÜD SENTON	03/2009	03/2011
Preamplifier	1651	CPA9231A	3393	Schaffner Electrotest	TÜV SÜD SENTON	10/2010	10/2012
Preamplifier	1684	AFS3-00100800-32-LN	847743	MITEQ	TÜV SÜD SENTON	04/2010	10/2011
V-network	1059	ESH3-Z5	894785/005	Rohde & Schwarz	Rohde & Schwarz	05/2009	05/2011
Loop antenna	1016	HFH2-Z2	882964/0001	Rohde & Schwarz	Rohde & Schwarz	09/2009	03/2011
Double ridged waveguide horn antenna	1516	3115	9508-4553	EMCO Elektronik	Seibersdorf Laboratories	10/2010	10/2012
TRILOG Broadband Antenna	1802	VULB 9163	9163-214	Schwarzbeck	Schwarzbeck	11/2009	05/2011
TRILOG Broadband Antenna	2058	VULB 9163	9163-408	Schwarzbeck	Schwarzbeck	11/2009	05/2011
Horn antenna	1012	3160-05	9112-1001	EMCO Elektronik			Note 2
Horn antenna	1013	3160-06	9112-1001	EMCO Elektronik			Note 2
Horn antenna	1014	3160-07	9112-1008	EMCO Elektronik			Note 2
Horn antenna		3160-08	9112-1002	EMCO Elektronik			Note 2
Horn antenna		3160-09	9403-1025	EMCO Elektronik			Note 2

Note 1: Date of next calibration contains maximum tolerance if applicable.

Note 2: According to manufacturer's specifications no calibration is required but mechanical check only. This check is performed before every test sequence.