Client

TA

Certificate No: D2450V2-SN712 Jul10

### **CALIBRATION CERTIFICATE**

Object

D2450V2 - SN: 712

Calibration Procedure(s)

TMC-XZ-01-027

Calibration procedure for dipole validation kits

Calibration date:

July 15, 2010

Condition of the calibrated item

In Tolerance

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature ( $22\pm3$ )  $^{\circ}$ C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards      | ID# C      | al Date(Calibrated by, Certificate No.) | Scheduled Calibr | ation |
|------------------------|------------|---|------------------|-------|
| Power Meter NRVD       | 101253     | 18-Jun-10 (TMC, No.JZ10-248)            | Jun-11           |       |
| Power sensor NRV-Z5    | 100333     | 18-Jun-10 (TMC, No. JZ10-248)           | Jun-11           |       |
| Reference Probe ES3DV3 | SN 3149    | 25-Sep-09(SPEAG, No.ES3-3149_Sep09)     | Sep-10           |       |
| DAE4                   | SN 777     | 09-Jul-10(TMC, No.DAE4-777_Jul10)       | Jul-11           |       |
| RF generator E4438C    | MY4509287  | 9 17-Jun-10(TMC, No.JZ10-302)           | Jun-11           |       |
| Network Analyzer 8753E | US38433212 | 02-Aug-09(TMC, No.JZ09-056)             | Aug-10           |       |

|                | Name        | Function                          | Signature |
|----------------|-------------|-----------------------------------|-----------|
| Calibrated by: | Lin Hao     | SAR Test Engineer                 | 献卷        |
| Reviewed by:   | Qi Dianyuan | SAR Project Leader                | 2000      |
| Approved by:   | Xiao Li     | Deputy Director of the laboratory | ART.      |

Issued: July 15, 2010

This calibration certificate shall not be reproduced except in full without written approval of the laboratoty.

Certificate No: D2450V2-712\_Jul10

Page 1 of 9

Glossary:

TSL

tissue simulating liquid

ConvF N/A sensitivity in TSL / NORMx,y,z not applicable or not measured

Calibration is Performed According to the Following Standards:
a) IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003

b) IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) For hand-held devices used in close proximity to the ear (frequency range of 300MHz to 3GHz)", February 2005

c) Federal Communications Commission Office of Engineering & Technology (FCC OET), "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields; Additional Information for Evaluating Compliance of Mobile and Portable Devices with FCC Limits for Human Exposure to Radiofrequency Emissions", Supplement C (Edition 01-01) to Bulletin 65

#### Additional Documentation:

d) DASY System Handbook

Certificate No: D2450V2-712\_Jul10

#### Methods Applied and Interpretation of Parameters:

 Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.

Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point
exactly below the center marking of the flat phantom section, with the arms oriented parallel to
the body axis.

Feed Point Impedance and Return Loss: These parameters are measured with the dipole
positioned under the liquid filled phantom. The impedance stated is transformed from the
measurement at the SMA connector to the feed point. The Return Loss ensures low reflected
power. No uncertainty required.

 Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.

SAR measured: SAR measured at the stated antenna input power.

SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna

 SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

# Measurement Conditions DASY system configuration, as

| DASY5                  | V5.0  |
|------------------------|---|
| Advanced Extrapolation |   |
| 2mm Oval Phantom ELI4  |   |
| 10 mm                  | with Spacer   |
| dx, dy, dz = 5 mm      |   |
| 2450 MHz ± 1 MHz       |   |
|                        | Advanced Extrapolation  2mm Oval Phantom EL14  10 mm  dx, dy, dz = 5 mm |

# Head TSL parameters The following parameters a

|                                  | Temperature     | Permittivity | Conductivity     |
|----------------------------------|-----------------|--------------|------------------|
| Nominal Head TSL parameters      | 22.0 °C         | 39.2         | 1.80 mhc/m       |
| Measured Head TSL parameters     | (22.0 ± 0.2) °C | 40.0 ± 6 %   | 1.78 mho/m ± 6 % |
| Head TSL temperature during test | (21.9 ± 0.2) °C |              |                  |

### SAR result with Head TSL

| SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL | Condition          | 1                         |
|---|--------------------|---------------------------|
| SAR measured  | 250 mW input power | 13.3 mW/g                 |
| SAR normalized  | normalized to 1W   | 53.2 mW/g                 |
| SAR for nominal Head TSL parameters 1                 | normalized to 1W   | 53.7 mW /g ± 17.0 % (k=2) |

| SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL | Condition          |                           |
|---|--------------------|---------------------------|
| SAR measured =  | 250 mW input power | 6.22 mW/g                 |
| SAR normalized  | normalized to 1W   | 24.9 mW/g                 |
| SAR for nominal Head TSL parameters 1                   | normalized to 1W   | 25.0 mW /g ± 16.5 % (k=2) |

<sup>\*</sup> Correction to nominal TSL parameters according to d), chapter "SAR Sensitvities"

**Body TSL parameters** 

The following parameters and calculations were applied.

|                                  | Temperature     | Permittivity | Conductivity     |
|----------------------------------|-----------------|--------------|------------------|
| Nominal Body TSL parameters      | 22.0 °C         | 52.7         | 1,95 mho/m       |
| Measured Body TSL parameters     | (22.0 ± 0.2) °C | 52.5 ± 6%    | 2.02 mho/m ± 6 % |
| Body TSL temperature during test | (21.8 ± 0.2) °C | 244          |                  |

## SAR result with Body TSL

| SAR averaged over 1 cm <sup>3</sup> (1 g) of Body TSL | Condition          |                           |
|---|--------------------|---------------------------|
| SAR measured  | 250 mW input power | 13.1 mW/g                 |
| SAR normalized  | normalized to 1W   | 52.4 mW / g               |
| SAR for nominal Body TSL parameters <sup>2</sup>      | normalized to 1W   | 52.6 mW /g ± 17.0 % (k=2) |

| SAR averaged over 10 cm <sup>3</sup> (10 g) of Body TSL | Condition          |                           |
|---|--------------------|---------------------------|
| SAR measured  | 250 mW input power | 6.18 mW/g                 |
| SAR normalized  | normalized to 1W   | 24.7 mW / g               |
| SAR for nominal Body TSL parameters <sup>2</sup>        | normalized to 1W   | 24.8 mW /g ± 16.5 % (k=2) |

<sup>&</sup>lt;sup>2</sup> Correction to nominal TSL parameters according to d), chapter "SAR Sensitivities"

#### **Appendix**

#### Antenna Parameters with Head TSL

| Impedance, transformed to feed point | 53.8Ω + 2.0 jΩ |  |
|--------------------------------------|----------------|--|
| Return Loss                          | - 26.8 dB      |  |

## Antenna Parameters with Body TSL

| Impedance, transformed to feed point | 50.3Ω + 5.1 jΩ |  |
|--------------------------------------|----------------|--|
| Return Loss                          | - 25.4 dB      |  |

#### General Antenna Parameters and Design

| Electrical Delay (one direction) | 1,155 ns |
|----------------------------------|----------|
|                                  |          |

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

#### **Additional EUT Data**

| Manufactured by | SPEAG         |
|-----------------|---------------|
| Manufactured on | July 05, 2002 |

#### **DASY5 Validation Report for Head TSL**

Date/Time: 2010-7-15 9:15:30

Test Laboratory: TMC, Beijing, China

DUT: Dipole 2450 MHz; Type: D2450V2; Serial: SN: 712

Communication System: CW Frequency: 1800 MHz Duty Cycle: 1:1 Medium: Head 2450MHz

Medium parameters used: f = 2450 MHz;  $\sigma = 1.78 \text{ mho/m}$ ;  $\epsilon_r = 40.0$ ;  $\rho = 1000 \text{ kg/m}^3$ 

Phantom section: Flat Section

#### DASY5 Configuration:

Probe: ES3DV3 - SN3149; ConvF(5.18, 5.18, 5.18); Calibrated: 25.09.09

Electronics: DAE4 Sn777; Calibration: 09.07.10

Phanton: 2mm Oval Phantom ELI4; Type: QDOVA001BB

Measurement SW: DASY5, V5.0 Build 119.9; Postprocessing SW: SEMCAD, V13.2 Build 87


#### Pin=250mW; d=10mm/Zoom Scan (7x7x7)/Cube 0:

Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 98.1 V/m; Power Drift = -0.057 dB


Peak SAR (extrapolated) = 28.8 W/kg

SAR(1 g) = 13.3 mW/g; SAR(10 g) = 6.22 mW/gMaximum value of SAR (measured) = 17.1 mW/g



Certificate No: D2450V2-712\_Jul10

### Impedance Measurement Plot for Head TSL



## DASY5 Validation Report for Body TSL

Date/Time: 2010-7-15 10:37:31

Test Laboratory: TMC, Beijing, China

DUT: Dipole 2450 MHz; Type: D2450V2; Serial: SN: 712

Communication System: CW Frequency: 2450 MHz Duty Cycle: 1:1

Medium: Body 1800MHz

Medium parameters used; f = 2450 MHz;  $\sigma$  = 2.02 mho/m;  $\epsilon$  , = 52.5;  $\rho$  = 1000 kg/m³

Phantom section: Flat Section

#### DASY5 Configuration:

Probe: ES3DV3 - SN3149; ConvF(4.97, 4.97, 4.97); Calibrated: 25.09.09

Electronics: DAE4 Sn777; Calibration: 09.07.10

Phanton: 2mm Oval Phantom ELI4; Type: QDOVA001BB

Measurement SW: DASY5, V5.0 Build 119.9; Postprocessing SW: SEMCAD, V13.2 Build 87


# Pin=250mW; d=10mm/Zoom Scan (7x7x7)/Cube 0:

Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 94.8 V/m; Power Drift = -0.009 dB

Peak SAR (extrapolated) =29.1 W/kg

 $SAR(1 g) = 13.1 \ \text{mW/g}; \ SAR(10 g) = 6.18 \ \text{mW/g}$  Maximum value of SAR (measured) =  $16.5 \ \text{mW/g}$ 



Certificate No: D2450V2-712\_Jul10

Page 8 of 9

# Impedance Measurement Plot for Body TSL

