

RF Exposure Statement:
Page 1 of 1

Client: ANDON HEALTH CO.,LTD

No. 3 Jinping Road, Ya An Street Nankai District Tianjin P.R. China

Test item: iHealth Wireless Smart Gluco-Monitoring System

Identification:BG5FCC ID:ZRYBG5

FCC Requirement

According to FCC 2.1091, mobile equipment must comply with the following applicable limit for maximum permissible exposure (MPE) specified in FCC 1.1310:

Equipment Use	Frequency Range	Power Density [mW/cm ²]	Average Time [min]
General Population /	1.5 – 100GHz	1	6
Uncontrolled Exposure			

IC Requirement

According to RSS-102 (Issue 4), clause 2.5.2, no routine RF exposure evaluation is required if the transmitter power (e.i.r.p.) is below the following threshold:

Frequency Range	SAR Limitation [W]	
Above 1.5GHz	5	

Measurement Result

The maximum measured transmitter power is the following:

Conducted Output Power P _{out} [mW]	Maximum Antenna Gain [dBi]	P _{out} EIRP [mW]	Power Density at 20cm [mW/cm ²]
1.05	2.45	0.24	0.00036

Note

The power density S in mW/cm² is calculated according to the Friis formula: $S = (P_{out} \cdot G) / (4\pi \cdot D^2)$, where $S = power density in mW/cm^2$

P_{out} = antenna conducted output power in mW

G = antenna gain in linear scale (here: 2.45dBi=10log(G))

D = distance between observation point and radiating structure in cm (here: 20cm)

Conclusion

The device complies with the FCC and IC RF exposure requirements since the maximum transmitter power density is below the FCC limit and the e.i.r.p. power is below the IC RF exposure evaluation exemption threshold.

Refer to test report 15058197 001 for more details.