



Certificate of Analysis

1.06035.1000 Methanol hypergrade for LC-MS LiChrosolv®
Batch I0931035

Batch Values		
Purity (GC)	≥ 99.9	%
Identity (IR)	conforms	
Evaporation residue	≤ 1.0	mg/l
Water	≤ 0.01	%
Acidity	≤ 0.0002	meq/g
Alkalinity	≤ 0.0002	meq/g
Al (Aluminium)	≤ 10	ppb
Ca (Calcium)	≤ 10	ppb
Fe (Iron)	≤ 10	ppb
K (Potassium)	≤ 5	ppb
Mg (Magnesium)	≤ 10	ppb
Na (Sodium)	≤ 100	ppb
Any further metal (ICP-MS)	≤ 5	ppb
Gradient grade (at 220 nm)	≤ 2.0	mAU
Gradient grade (at 235 nm)	≤ 1.0	mAU
Fluorescence (as quinine at 254 nm)	≤ 1.0	ppb
Fluorescence (as quinine at 365 nm)	≤ 0.5	ppb
Transmission (at 210 nm)	≥ 35	%
Transmission (at 220 nm)	≥ 60	%
Transmission (at 230 nm)	≥ 75	%
Transmission (from 260 nm)	≥ 98	%
Suitable for LC-MS (tested with ion trap MS); (Intensity of background mass peak based on reserpine (APCI/ESI positiv))	≤ 2	ppb
Suitable for LC-MS (tested with ion trap MS); Intensity of background mass peak based on reserpine (APCI/ESI negative)	≤ 20	ppb

Suitable for PAH-analysis.

Filtered by 0.2 µm filter.

Suitable for UPLC / UHPLC / Ultra HPLC - instruments.

Suitable for Q-TOF LC-MS.

Date of release (DD.MM.YYYY) 19.01.2018

Minimum shelf life (DD.MM.YYYY) 31.01.2021

Jeannette David

Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.