

PDF document of the Fridge-tag® 2

Identification number: 130400016798
 Date and time of report creation: 12.01.2022 09:25h
 Activation date: 13.08.2020 06:44h
 Upper alarm limit: Above +8.0°C for 10h
 Lower alarm limit: Below -0.5°C for 1h
 Measurement interval:¹⁾ 1min (fixed)

No.	Date (dd.MM.yyyy)	Events ²⁾	Average temp.	Lower alarm limit				Upper alarm limit				Signature / notes Action taken
				Status	Min. temp.	Cumulative daily time below the limit	Alarm trigger time	Status	Max. temp.	Cumulative daily time above the limit	Alarm trigger time	
1	Today		+2.9°C	In progress	+2.8°C	0min		In progress	+3.1°C	0min		
2	11.01.2022		+2.4°C	ok	+2.2°C	0min		ok	+2.8°C	0min		
3	10.01.2022		+2.6°C	ok	+2.3°C	0min		ok	+3.6°C	0min		
4	09.01.2022		+3.1°C	ok	+3.0°C	0min		ok	+3.6°C	0min		
5	08.01.2022		+2.8°C	ok	+2.5°C	0min		ok	+3.0°C	0min		
6	07.01.2022		+2.9°C	ok	+2.7°C	0min		ok	+3.0°C	0min		
7	06.01.2022		+3.1°C	ok	+2.6°C	0min		ok	+5.5°C	0min		
8	05.01.2022		+3.3°C	ok	+2.8°C	0min		ok	+7.1°C	0min		
9	04.01.2022		+3.4°C	ok	+2.8°C	0min		ok	+3.7°C	0min		
10	03.01.2022		+3.0°C	ok	+2.8°C	0min		ok	+3.4°C	0min		
11	02.01.2022		+2.7°C	ok	+2.5°C	0min		ok	+3.2°C	0min		
12	01.01.2022		+3.4°C	ok	+3.1°C	0min		ok	+3.8°C	0min		
13	31.12.2021		+3.1°C	ok	+3.1°C	0min		ok	+3.5°C	0min		
14	30.12.2021		+2.8°C	ok	+2.7°C	0min		ok	+3.1°C	0min		
15	29.12.2021		+2.4°C	ok	+2.3°C	0min		ok	+2.7°C	0min		
16	28.12.2021		+3.0°C	ok	+2.3°C	0min		ok	+4.8°C	0min		
17	27.12.2021		+2.8°C	ok	+2.6°C	0min		ok	+3.4°C	0min		
18	26.12.2021		+2.3°C	ok	+2.1°C	0min		ok	+2.7°C	0min		
19	25.12.2021		+3.0°C	ok	+2.3°C	0min		ok	+3.8°C	0min		
20	24.12.2021		+2.9°C	ok	+2.7°C	0min		ok	+3.3°C	0min		
21	23.12.2021		+2.5°C	ok	+2.4°C	0min		ok	+3.0°C	0min		
22	22.12.2021		+3.0°C	ok	+2.6°C	0min		ok	+3.8°C	0min		
23	21.12.2021		+3.1°C	ok	+3.0°C	0min		ok	+3.7°C	0min		
24	20.12.2021		+2.6°C	ok	+2.5°C	0min		ok	+3.0°C	0min		
25	19.12.2021		+2.9°C	ok	+2.4°C	0min		ok	+3.7°C	0min		
26	18.12.2021		+2.6°C	ok	+2.2°C	0min		ok	+3.2°C	0min		
27	17.12.2021		+2.1°C	ok	+2.0°C	0min		ok	+2.3°C	0min		
28	16.12.2021		+2.4°C	ok	+2.3°C	0min		ok	+3.2°C	0min		
29	15.12.2021		+3.2°C	ok	+3.0°C	0min		ok	+3.6°C	0min		
30	14.12.2021		+2.7°C	ok	+2.5°C	0min		ok	+3.3°C	0min		

1) Sampling and data analysis every minute
 2) t = time / date changed

Date and place: _____

Signature: _____

PDF document of the Fridge-tag® 2

Identification number: 130400016798
 Date and time of report creation: 12.01.2022 09:25h
 Activation date: 13.08.2020 06:44h
 Upper alarm limit: Above +8.0°C for 10h
 Lower alarm limit: Below -0.5°C for 1h
 Measurement interval:¹⁾ 1min (fixed)

No.	Date (dd.MM.yyyy)	Events ²⁾	Average temp.	Lower alarm limit				Upper alarm limit				Signature / notes Action taken
				Status	Min. temp.	Cumulative daily time below the limit	Alarm trigger time	Status	Max. temp.	Cumulative daily time above the limit	Alarm trigger time	
31	13.12.2021		+3.4°C	ok	+3.2°C	0min		ok	+3.8°C	0min		
32	12.12.2021		+3.0°C	ok	+2.9°C	0min		ok	+3.3°C	0min		
33	11.12.2021		+2.7°C	ok	+2.4°C	0min		ok	+5.0°C	0min		
34	10.12.2021		+3.2°C	ok	+2.4°C	0min		ok	+3.6°C	0min		
35	09.12.2021		+2.8°C	ok	+2.6°C	0min		ok	+3.1°C	0min		
36	08.12.2021		+2.4°C	ok	+2.3°C	0min		ok	+3.2°C	0min		
37	07.12.2021		+3.2°C	ok	+3.0°C	0min		ok	+3.6°C	0min		
38	06.12.2021		+2.8°C	ok	+2.6°C	0min		ok	+3.3°C	0min		
39	05.12.2021		+2.5°C	ok	+2.3°C	0min		ok	+3.3°C	0min		
40	04.12.2021		+3.1°C	ok	+2.8°C	0min		ok	+3.8°C	0min		
41	03.12.2021		+2.9°C	ok	+2.9°C	0min		ok	+3.1°C	0min		
42	02.12.2021		+3.0°C	ok	+2.8°C	0min		ok	+3.4°C	0min		
43	01.12.2021		+3.0°C	ok	+2.5°C	0min		ok	+4.0°C	0min		
44	30.11.2021		+3.1°C	ok	+2.6°C	0min		ok	+3.9°C	0min		
45	29.11.2021		+3.3°C	ok	+2.9°C	0min		ok	+4.5°C	0min		
46	28.11.2021		+2.8°C	ok	+2.5°C	0min		ok	+3.3°C	0min		
47	27.11.2021		+3.4°C	ok	+3.2°C	0min		ok	+3.8°C	0min		
48	26.11.2021		+2.9°C	ok	+2.8°C	0min		ok	+3.2°C	0min		
49	25.11.2021		+3.0°C	ok	+2.5°C	0min		ok	+4.1°C	0min		
50	24.11.2021		+3.1°C	ok	+3.0°C	0min		ok	+3.8°C	0min		
51	23.11.2021		+3.1°C	ok	+2.8°C	0min		ok	+4.1°C	0min		
52	22.11.2021		+3.0°C	ok	+2.7°C	0min		ok	+4.1°C	0min		
53	21.11.2021		+3.0°C	ok	+2.5°C	0min		ok	+3.6°C	0min		
54	20.11.2021		+3.1°C	ok	+2.6°C	0min		ok	+4.0°C	0min		
55	19.11.2021		+3.0°C	ok	+2.8°C	0min		ok	+3.6°C	0min		
56	18.11.2021		+2.8°C	ok	+2.6°C	0min		ok	+3.8°C	0min		
57	17.11.2021		+3.4°C	ok	+2.9°C	0min		ok	+3.8°C	0min		
58	16.11.2021		+3.2°C	ok	+3.0°C	0min		ok	+3.6°C	0min		
59	15.11.2021		+2.7°C	ok	+2.6°C	0min		ok	+3.3°C	0min		
60	14.11.2021		+3.1°C	ok	+2.8°C	0min		ok	+3.6°C	0min		

1) Sampling and data analysis every minute
 2) t = time / date changed

Date and place: _____

Signature: _____