

UNITED ACCREDITATION FOUNDATION



Document Name	ACCREDITATION REQUIREMENTS FOR CERTIFICATION OF MEDICAL DEVICES - QUALITY MANAGEMENT SYSTEMS (MDQMS)
Document Number	UAF-CAB-MDQMS
Applicable For	APPLICABLE TO ALL CERTIFICATION BODIES APPLYING FOR OR HOLDING ACCREDITATION FOR MDQMS
Revision Number	V02
Revision Date	MAR-10-2026
Effective Date	MAR-10-2026



**ACCREDITATION REQUIREMENTS FOR
CERTIFICATION OF MEDICAL DEVICES-
QUALITY MANAGEMENT SYSTEMS (MDQMS)**

Doc. No.: UAF-CAB-MDQMS

Rev. No.: V02

Effective Dt: MAR-10-2026

Contents

1.	ACCREDITATION REQUIREMENTS	3
2.	ADDITIONAL DOCUMENTS	3
3.	UAF DOCUMENTS APPLICABLE FOR ACCREDITATION OF ISO 13485 CERTIFICATION	3
4.	CERTIFICATION DOCUMENTS	4
5.	SCOPE OF ACCREDITATION	4
6.	REQUIREMENTS	4
7.	RESOURCE REQUIREMENTS	4
8.	INFORMATION REQUIREMENTS	5
9.	PROCESS REQUIREMENTS	5
10.	MANAGEMENT SYSTEM REQUIREMENTS FOR CERTIFICATION BODIES	5
11.	ACCREDITATION ASSESSMENTS	5
12.	ASSESSMENTS	5
13.	INITIAL ASSESSMENTS	6
14.	WITNESS ASSESSMENTS	6
15.	SURVEILLANCE AND REASSESSMENTS	7
16.	SCOPE EXTENSION	8
17.	ANNEX A: SCOPES AS PER IAF MD 8:2020	8

1. ACCREDITATION REQUIREMENTS

This scheme document establishes additional scheme-specific accreditation requirements applicable to certification bodies performing certification against the applicable management system standard covered under this scheme. These requirements supplement the applicable accreditation standards, including UAF-GEN-CAB-01 — General Accreditation Requirements and UAF-GEN-CAB-02 — General Rules and Conditions for Use of UAF Accreditation Mark, together with other applicable international mandatory documents and normative references identified in this document.

Accreditation of Certification Bodies under this scheme shall be granted and maintained in accordance with the following requirements:

- ISO/IEC 17021-1 — Conformity assessment — Requirements for bodies providing audit and certification of management systems.
- UAF-GEN-CAB-01 — General Accreditation Requirements.
- UAF-GEN-CAB-02 — General Rules and Conditions for Use of UAF Accreditation Mark.
- Applicable IAF/GACI Mandatory Documents and other normative documents referenced in this scheme document.

In case of any ambiguity, difference in interpretation, or lack of clarity between this scheme document and other applicable UAF accreditation documents and policies, clarification shall be obtained from UAF. The interpretation or guidance provided by UAF shall be considered final for the purposes of accreditation.

2. ADDITIONAL DOCUMENTS

ISO/IAF/GACI documents that are applicable for accreditation of ISO 13485 certification:

- IAF MD 1 IAF Mandatory Document for the Audit and Certification of a Management System Operated by a Multi-Site Organization
- IAF MD 2 IAF Mandatory Document for the Transfer of Accredited Certification of Management Systems.
- IAF MD 4 IAF Mandatory Document for the Use of Information and Communication Technology (ICT) for Auditing/Assessment Purposes
- IAF MD 9 Application of ISO/IEC 17021 in Medical Device Quality Management Systems (ISO 13485)
- IAF MD 11 IAF Mandatory Document for Application of ISO/IEC 17021 for Audits of Integrated Management Systems (IMS)
- IAF MD 23 Control of Entities Operating on Behalf of Accredited Management Systems Certification Bodies
- IAF MD 28 IAF Mandatory Document for the Upload and Maintenance of Data on IAF Database

3. UAF DOCUMENTS APPLICABLE FOR ACCREDITATION OF ISO 13485 CERTIFICATION

- UAF-GEN-CAB-01-General Accreditation requirements
- UAF-GEN-CAB-02-Conditions for the use of UAF Symbol
- UAF-PL:2014-01: Policy on Providing Certification Activities in Foreign Countries Where Regulatory Approval or Authorization is Required.

- UAF F-31A-Annexure A- Guidance for addition of countries, location(s), critical location(s).
- UAF-PL:2016-06: UAF Policy for the Collection of Data to Provide Indicators of Management System Certification Bodies' Performance as per IAF MD 15:2014.
- UAF-PL:2018-04: UAF Policy on handling misconduct & unethical practices in the certification process
- UAF-PL:2019-07-01: UAF Policy on accreditation scopes & witnessing activities for accreditation of management system CB
- UAF-PL:2019-07-02: UAF Policy regarding general administration
- UAF Transition Policies as applicable.

4. CERTIFICATION DOCUMENTS

Certification bodies (CB) certify against ISO 13485; Medical devices — Quality management systems — Requirements for regulatory purposes.

5. SCOPE OF ACCREDITATION

The Scope of accreditation for ISO 13485 is defined in accordance with the requirements of IAF MD 9. It is granted as per the technical area corresponding to the directives concerning medical devices. These have been listed in Annex A. UAF may further restrict Scope to specific Technical Areas.

6. REQUIREMENTS

6.1 General

6.1.1 Legal And Contractual Matters

The requirements of ISO/IEC 17021-1:2015, Clause 5.1 apply. In addition, the MDQMS specific requirements and guidance apply as detailed in IAF MD9.

6.1.2 Management Of Impartiality

The requirements of ISO/IEC 17021-1:2015, Clause 5.2 apply. In addition, the MDQMS specific requirements and guidance apply as detailed in IAF MD9.

6.1.3 Liability And Financing

The requirements of ISO/IEC 17021-1:2015, Clause 5.3 apply.

6.2 Structural Requirements

The requirements of ISO/IEC 17021-1:2015, Clause 6 apply.

7. RESOURCE REQUIREMENTS

7.1 Competence of management and personnel

The requirements of ISO/IEC 17021-1:2015, Clause 7.1 apply and in addition, the MDQMS specific requirements and guidance apply as detailed in IAF MD9.

7.2 Personnel involved in the certification activities

The requirements of ISO/IEC 17021-1:2015, Clause 7.2 apply and in addition, the MDQMS specific requirements and guidance apply as detailed in IAF MD9.

7.3 Use of individual external auditors and external technical experts

The requirements of ISO/IEC 17021-1:2015, Clause 7.2 apply and in addition, the MDQMS specific requirements and guidance apply as detailed in IAF MD9.

8. INFORMATION REQUIREMENTS

8.1 Publicly accessible information

The requirements of ISO/IEC 17021-1:2015, Clause 8.1 apply and in addition, the MDQMS specific requirements and guidance apply as detailed in IAF MD9.

8.2 Certification documents

The requirements of ISO/IEC 17021-1:2015, Clause 8.2 apply and in addition, the MDQMS specific requirements and guidance apply as detailed in IAF MD9.

8.3 Reference to certification and use of marks

The requirements of ISO/IEC 17021-1:2015, Clause 8.3 apply.

8.4 Confidentiality

The requirements of ISO/IEC 17021-1:2015, Clause 8.4 apply.

8.5 Information exchange between a certification body and its clients

The requirements of ISO/IEC 17021-1:2015, Clause 8.5 apply.

9. PROCESS REQUIREMENTS

The requirements of ISO/IEC 17021-1:2015, Clause 9 apply. In addition, the MDQMS specific requirements and guidance apply as detailed in IAF MD9.

10. MANAGEMENT SYSTEM REQUIREMENTS FOR CERTIFICATION BODIES

The requirements of ISO/IEC 17021-1:2015, Clause 10 apply. In addition, the MDQMS specific requirements and guidance apply as detailed in IAF MD9.

11. ACCREDITATION ASSESSMENTS

The accreditation assessment depends upon the following factors:

- a. The applied technical areas.
- b. Key locations and other operational non-key locations.
- c. Countries in which UAF-accredited certificates may be issued.
- d. Existing accreditation by other Accreditation Bodies.
- e. Other certification or conformity assessment activities carried out by the Certification Body
- f. Other relevant details provided in the application.

12. ASSESSMENTS

- a. Document Review
- b. Initial Office Assessments
- c. Witness Assessments.
- d. Remote Assessments

- e. Scope Verifications during extraordinary circumstances.

13. INITIAL ASSESSMENTS

The assessment extent and content depend on the requested scope of accreditation, the other activities for which the body is accredited or requests accreditation and the performance of the body at previous assessments.

The assessment consists of the following activities:

- Document Review
- Office assessment
- Witness Assessments
- Scope Verifications during extraordinary circumstances.

14. WITNESS ASSESSMENTS

For the minimum number of audits to be witnessed, the requirements of IAF MD 8, Clause 7.4.5 shall apply.

The following may be considered in deciding the number of witnesses and selection of witnesses during surveillance and reassessment:

- a. The CB's overall performance.
- b. IAF MD 8
- c. Factors such as process complexity or legislation, etc., influence the certified organization's ability to demonstrate its ability to meet the intended outcome of the MS.
- d. Feedback from interested parties, including complaints about certified organizations.
- e. The results of the CB's internal audit(s).
- f. Scheme owner requirements, etc.
- g. Change(s) in CB work patterns – growth of work within a specific region or technical area.
- h. A number of clients within the CB's Scope of accreditation.
- i. Confidence in the CB's auditor evaluation and approval process.
- j. Previous or other office or witnessing assessment results, etc.

In addition to the general criteria for selection of audits to be witnessed, UAF may also consider the following:

- a. UAF will normally avoid witnessing auditors who have already been witnessed for the same scheme.
- b. UAF will normally avoid witnessing audits at the same organization.
- c. UAF will review the CB's audit report as part of the witness assessment.
- d. To enable selection of audits for witnessing, the CB shall, upon request, provide audit planning information for a defined period. This information shall include, as a minimum:
 - Type of audit (initial, surveillance, or recertification)
 - Name and address of the auditee
 - Audit standard(s)
 - Scope of certification
 - Names of auditor(s) and technical expert(s)
 - Date(s) of the audit

15. SURVEILLANCE AND REASSESSMENTS

The surveillance assessments of the Certification Body's ISO 13485 certification activities shall normally be conducted annually or as defined in the assessment program, or earlier where justified by the assessment team based on identified risks, changes, or performance concerns. The duration and scope of surveillance assessments shall be determined through a risk-based assessment program, considering the volume, complexity, and maturity of the ISO 13485 certification activities.

The number and type of witness assessments during surveillance shall be determined in accordance with the assessment program in line with IAF MD 8, Clause 7.9.3 and UAF-PL:2019-07-01 — UAF Policy on accreditation scopes & witnessing activities for accreditation of management system CB. Witnessing decisions and number of witnesses may vary considering the outcomes of certification file reviews, sector complexity, use of technical experts, and changes to personnel or processes.

During each surveillance assessment, UAF shall verify the continued effective implementation of the ISO 13485 certification scheme, including compliance with IAF MD9 and applicable accreditation requirements.

Certification files reviewed during surveillance and reassessment shall be selected using a risk-based sampling methodology. The sampling approach may consider:

- Number of ISO 13485 certificates issued since the previous assessment
- Organization size, complexity and other factors
- Risk Classification of Medical device
- Use of ICT or remote assessment techniques
- Previous nonconformities or complaints

The witnessing program shall ensure, as a minimum, that one audit from each of the Main Technical Areas (shown in Annex A) under the Scope of accreditation within an accreditation cycle (surveillances and reassessment) is conducted before the expiry of accreditation. The sampling for witnessing shall give priority to higher-risk technical areas. Witness assessments should avoid the repeated witnessing of the same CAB client organization. The AB shall take into account previous results of witnessing to establish its witness strategy. The number of files to be reviewed for each assessment is calculated based on the number of certificates issued since the last assessment. This is approximately one-tenth of the square root of the number of certificates, with a maximum of six files and a minimum of one file. The number of files to be reviewed may increase depending on risk-based parameters detailed in the assessment program.

Where applicable, witness assessments may be planned to cover multiple schemes (e.g. combined audits with other standards), provided that it ensures adequate coverage of ISO 13485-specific competence and assessment objectives.

Applying the above guidelines shall consider whether witness assessments may serve multiple schemes (e.g. through witnessing combined audits) and how such witness assessments are performed in other schemes. The accreditation cycle will be for four years,

and the reaccreditation assessment should be carried out at least four months before the accreditation expiry. The assessment plan will be listed in the assessment program and

In applying the above guidelines, it shall be considered whether witness assessments may serve for multiple schemes (e.g., by witnessing combined audits) and how witness assessments are performed in other schemes.

16. SCOPE EXTENSION

The assessment by UAF in case of an application for an additional scope (Main Technical Area) consists of one or more of the following assessment methods:

- Documents and records review.
- Verification of documents and records and interview of relevant staff during a visit to the office of the CB.
- A witness assessment in the main technical cluster per UAF's witness policy document is in line with IAF MD 8 requirements.
- At least the defined requirements for competence in that technical cluster, records of the qualification process of auditors for that cluster, and a complete certification dossier in that cluster will be verified in the scope extension assessment.

17. ANNEX A: SCOPES AS PER IAF MD 8:2020

Main areas	technical	Technical areas	Product categories covered by the technical areas
Table A.1.1 - NON-ACTIVE MEDICAL DEVICES			
Non-active Medical Devices		General non-active, non-implantable medical devices	- Non-active devices for anesthesia, emergency and intensive care - Non-active devices for injection, infusion, transfusion and dialysis - Non-active orthopedic and rehabilitation devices - Non-active medical devices with measuring function - Non-active ophthalmologic devices - Non-active instruments - Contraceptive medical devices - Non-active medical devices for disinfecting, cleaning, rinsing - Non-active devices for in vitro fertilisation (IVF) and assisted reproductive technologies (ART) - Non-active medical devices for ingestion
		Non-active implants	- Non-active cardiovascular implants - Non-active orthopedic implants - Non-active functional implants - Non-active soft tissue implants
		Devices for wound care	Bandages and wound dressings

Main areas	technical	Technical areas	Product categories covered by the technical areas
			• Suture material and clamps • Other medical devices for wound care
		Non-active dental devices and accessories	• Non-active dental devices/equipment and instruments • Dental materials • Dental implants
		Non-active medical devices other than specified above	
Table A.1.2 - ACTIVE (NON-IMPLANTABLE) MEDICAL DEVICES			
Active Devices (Non-implantable)	Medical (Non-implantable)	General active medical devices	• Devices for extra-corporal circulation, infusion and haemopheresis • Respiratory devices, devices including hyperbaric chambers for oxygen therapy, inhalation anaesthesia • Devices for stimulation or inhibition • Active surgical devices • Active ophthalmologic devices • Active dental devices • Active devices for disinfection and sterilization • Active rehabilitation devices and active prostheses • Active devices for patient positioning and transport • Active devices for in vitro fertilisation (IVF) and assisted reproductive technologies (ART) • Software, including software design for medical devices • Medical gas supply systems and parts thereof
		Devices for imaging	• Devices utilizing ionizing radiation • Devices utilizing non-ionizing radiation
		Monitoring devices	• Monitoring devices of non-vital physiological parameters • Monitoring devices of vital physiological parameters
		Devices for radiation therapy and thermo therapy	• Devices utilising ionizing radiation • Devices utilising non-ionizing radiation • Devices for hyperthermia / hypothermia • Devices for (extracorporal) shock-wave therapy (lithotripsy)
		Active (non-implantable) medical devices other than specified above	

Main areas	technical	Technical areas	Product categories covered by the technical areas
Table A.1.3 - ACTIVE IMPLANTABLE MEDICAL DEVICES			
Active Implantable Medical Devices	General active implantable medical devices	- Active implantable medical devices for stimulation / inhibition - Active implantable medical devices delivering drugs or other substances - Active implantable medical devices substituting or replacing organ functions	
	Implantable medical devices other than specified above		
Table A.1.4 - IN VITRO DIAGNOSTIC MEDICAL DEVICES			
In Vitro Diagnostic Medical Devices (IVD) Reagents	Reagents and reagent products, calibrators and control materials for: - Clinical Chemistry - Immunochemistry (Immunology) - Haematology/Haemostasis/ - Immunohematology - Microbiology - Infectious Immunology - Histology/Cytology - Genetic Testing		
	IVD Instruments and software		
	IVD medical devices other than specified above		
Table A.1.5 – STERILIZATION METHODS FOR MEDICAL DEVICES			
Sterilization Method for Medical Devices	Ethylene oxide gas sterilization (EOG)		
	Moist heat		
	Aseptic processing		
	Radiation sterilization (e.g. gamma, x-ray, electron beam)		
	Low temperature steam and formaldehyde sterilization		
	Thermic sterilization with dry heat		
	Sterilization with hydrogen peroxide		
	Sterilization method other than specified above		
Table A1.6 – DEVICES INCORPORATING / UTILIZING SPECIFIC SUBSTANCES / TECHNOLOGIES			
Devices incorporating/utilizing	Medical devices incorporating medicinal substances		

Main technical areas	Technical areas	Product categories covered by the technical areas
specific substances/ technologies		
	Medical devices utilizing tissues of animal origin	
	Medical devices incorporating derivatives of human blood	
	Medical devices utilizing micromechanics	
	Medical devices utilizing nanomaterials	
	Medical devices utilizing biological active coatings and/or materials or being wholly or mainly absorbed	
	Medical devices incorporating or utilizing specific substances / technologies / elements, other than specified above	
Table A1.7 – PARTS AND SERVICES		
Parts or services	Raw materials	Raw metals, plastic, wood, ceramic
	Components	Electrical components, fasteners, shaped raw materials, machined raw materials, and molded plastic
	Subassemblies	Electronic subassemblies mechanical subassemblies, made to drawings and/or work instructions
	Calibration services*	Verification/confirmation services for measuring instruments, tools, or test fixtures
	Distribution services	Distributors providing storage and delivery of medical devices, not acting as a 'legal manufacturer' for medical devices.
	Maintenance services	Electrical or mechanical repair services, facility cleaning and maintenance services, uniform cleaning and testing of ESD smocks.
	Transportation services	Trucking, shipping, air transportation service in general.
	Other services	Consulting services related to medical devices, packaging services, etc.