Technical Specification – Portable Suction Apparatus

1. Item Details

- **Item:** Portable Suction Apparatus
- **Category:** Medical Equipment & Devices
- **Quantity Required:** 2 units
- **Intended Department:** Medical & Surgical Equipment
- **Purpose:** Clinical suction during surgical procedures, emergency care, and patient transfers.

2. Minimum Technical Requirements

- Electric portable suction system
- Adjustable vacuum pressure
- Suitable for hospital clinical use
- Rechargeable battery backup
- Collection container for aspirated fluids
- Overflow protection mechanism
- Easy-to-clean and disinfect surfaces
- Compact and portable design
- Clear vacuum/pressure indication
- Appropriate safety protection for clinical operation

3. Performance Requirements

- Stable suction performance during continuous operation
- Adjustable suction level according to clinical requirements
- Reliable operation during patient transfer
- Battery operation available when mains power is unavailable
- Easy replacement/emptying of collection container

4. Safety Requirements

- Electrical safety protection
- Overflow protection
- Secure collection container connection
- Audible/visual alarm where applicable
- Equipment should comply with applicable medical-device safety standards

5. Accessories

The supplied equipment should include, as applicable:

- Collection container
- Suction tubing
- Patient suction accessories
- Power cable/adapter
- Rechargeable battery
- User/operation manual

6. Warranty & Service

- Minimum **1-year warranty**
- Supplier should provide installation/setup guidance where required
- Supplier should provide basic user training
- After-sales technical support should be available
- Replacement parts should be available during the supported service period

7. Supplier Documentation

Supplier should provide:

- Product technical specification
- User manual
- Warranty information
- Regulatory/compliance documentation
- Service and maintenance information