



PT CAHAYA MAS CEMERLANG

Office: Komplek Majapahit Permai Blok A No.110 & C No.105, Jl. Majapahit No.18-20-22, Jakarta Pusat 10160
Factory: Jalan Rawa Gelam II No.3, Kawasan Industri Pulogadung, Jakarta Timur
Telp: 021-344 3456 (Hunting) Fax: 021-460 2340/460 2344
E-mail: info@cahayamascemerlang.com | Website: www.cahayamascemerlang.com

CALIBRATION REPORT

Report No: Stasiun Dummy Kalibrasi/CR-2026/VIII/OMS-CMC/001

Station Name	: Stasiun Dummy Kalibrasi	Calibration Date	: 10 Agustus 2026
Address	: Jalan Dummy No. 1, Jakarta	Coordinate	: -

1. SENSOR PARAMETER CALIBRATION & REGISTER ADJUSTMENT

Parameter	Standart / CRM	Hasil Pembacaan	Internal Coeff (K / B)	Result
DO Calibration	-	Solution 1 Solution 2 CRM Reference: 5.5100	0.0000 99.8400 3.14 mg/L Status: PASS	K: 0.99911 B: -0.000024
Turbidity Calibration	-	Solution 1 Solution 2 CRM Reference: 4.6700	3.7000 19.6300 4.64 NTU Status: PASS	K: 0.265294 B: -5.97067
TDS Calibration	-	Solution 1 Solution 2 CRM Reference: 200.0000	1.2880 12.7900 201.29 mg/L Status: PASS	K: 1.067077 B: -0.185172
COD Calibration	-	Solution 1 Solution 2 CRM Reference: 14.8000	10.6100 95.2400 13.24 mg/L Status: PASS	K: 1.070413 B: -8.484534
BOD Calibration	-	Solution 1 Solution 2 CRM Reference: 9.1800	10.6100 95.2400 Status: PASS	K: 1.070413 B: -8.484534
pH Calibration	-	Solution 1 Solution 2 Solution 3	3.9700 Unit 6.8300 Unit 10.0000 Unit Status: PASS	K1: -58.77783 K2: -58.77783 K3: -21.16125 K4: -58.77783 K5: -58.77783 K6: -15.39816
TSS Calibration	-	Solution 1 CRM Reference: 50.0000	49.9100 Status: PASS	K: 0.01409858 B: 0
Amonia Calibration	-	Solution 1 CRM Reference: 0.8920	1.0100 Status: PASS	K: 1 B: 0.238715
Nitrit Calibration	-	Solution 1 Solution 2 CRM Reference: 1.8200	10.1000 99.2000 Status: PASS	-
Nitrat Calibration	-	Solution 1 Solution 2 CRM Reference: 5.4400	10.0200 99.5000 Status: PASS	*Stored internally in probe EEPROM

2. WATER SAMPLE MEASUREMENT & BLANK TEST

Sample Type	Temp (°C)	DO (mg/L)	TDS (mg/L)	Turb (NTU)	pH Unit	COD (mg/L)	BOD (mg/L)	TSS (mg/L)	NH3-N (mg/L)	NO3-N (mg/L)
Aquades (Blank)	25.0000	8.7700	0.0000	0.0000	6.9600	0.0000	0.0000	0.0000	0.0000	0.0000
Water Sample (River)	25.2000	6.8500	145.2000	12.4000	7.1500	18.5000	4.2000	15.0000	0.1200	2.1000

Notes: Dummy calibration record for a single station verification flow.

Place / Date: 11 Agustus 2026
Calibration Officer:



dummy_cal_officer
PT Cahaya Mas Cemerlang