


TiO₂ pad, saline diluted 3× and the biases moved off the clamp

scope_20260829_103015.csv, one continuous 1 kHz record, I_{na_ref} = 1.157 V. Two injections into the droplet: both are graded, both recover, nothing saturates.


1 — two injections

graded, and the baseline comes back both times



2 — first injection

the smaller aliquot: -206 mV, no residual offset



3 — second injection

the larger aliquot: -398 mV, no residual offset

