

INSTRUMENT TEST REPORT

Instrument:	Timbangan	No. Kontrol:	2QC-TMB-010
Kategori Sampel:	RUAH	Sampel:	LPRG5
Lokasi:	R. Timbang 1	QC Analyst:	Ayu Maharani (qc_analyst_oj_stabtest)
PIC:	admin	Cal. Expiry:	21/01/2027
Start Time:	21/09/26 13:28	End Time:	21/09/26 13:32
Duration:	3 min	Total Items:	3 items

Initial Condition:

[v] Timbangan Bersih: true
[v] Timbangan masih dalam ED Kalibrasi: true

Final Condition:

[v] Timbangan kembali ke 0.000 g: true

SUMMARY: Total: 3 Success: 3 (100.0%) Failed: 0 Re-reads: 0

DETAILED TEST RESULTS

Item	Batch	Result	Time
#1	JJ009_A	5.4826	21/09/2026 13:29
#2	JJ009_T	5.4809	21/09/2026 13:30
#3	JJ009_B	5.4846	21/09/2026 13:31

TESTED BY

Ayu Maharani
qc_analyst_oj_stabtest
21/09/2026