

INSTRUMENT TEST REPORT

Instrument:	pH Meter	No. Kontrol:	2QC-PHM-007
Kategori Sampel:	RUAH	Sampel:	LEJO1
Lokasi:	R. Instrument	QC Analyst:	Barkah Haren Sumonegoro (qc_analyst_oj_stabtest)
PIC:	admin	Cal. Expiry:	03/11/2026
Start Time:	15/09/26 16:39	End Time:	15/09/26 16:41
Duration:	2 min	Total Items:	3 items
Initial Condition:			
[v] Kondisi Awal: true			
Final Condition:			
[v] Kondisi Akhir: true			
SUMMARY:			
Total: 3		Success: 3 (100.0%)	Failed: 0
		Re-reads: 0	

DETAILED TEST RESULTS

Item	Batch	pH (pH)	Temperature (°C)	Time
#1	JJ093_A	3.44	23.7	15/09/2026 16:40
#2	JJ093_T	3.44	23.8	15/09/2026 16:40
#3	JJ093_B	3.44	23.8	15/09/2026 16:41

TESTED BY

Barkah Haren Sumonegoro

qc_analyst_oj_stabtest

15/09/2026