

INSTRUMENT TEST REPORT

|                                                                                                                 |                |              |                                                 |
|-----------------------------------------------------------------------------------------------------------------|----------------|--------------|-------------------------------------------------|
| Instrument:                                                                                                     | pH Meter       | No. Kontrol: | 2QC-PHM-007                                     |
| Kategori Sampel:                                                                                                | RUAH           | Sampel:      | LEJS1 - Extra Joss Sport Liq                    |
| Lokasi:                                                                                                         | R. Instrument  | QC Analyst:  | Thiara Juliana Rizdani (qc_analyst_oj_stabtest) |
| PIC:                                                                                                            | admin          | Cal. Expiry: | 03/11/2026                                      |
| Start Time:                                                                                                     | 07/09/26 14:52 | End Time:    | 07/09/26 14:54                                  |
| Duration:                                                                                                       | 2 min          | Total Items: | 3 items                                         |
| Initial Condition:                                                                                              |                |              |                                                 |
| [v] Kondisi Awal: true                                                                                          |                |              |                                                 |
| Final Condition:                                                                                                |                |              |                                                 |
| [v] Kondisi Akhir: true                                                                                         |                |              |                                                 |
| SUMMARY:      Total: 3                      Success: 3 (100.0%)      Failed: 0                      Re-reads: 0 |                |              |                                                 |

DETAILED TEST RESULTS

| Item | Batch   | pH (pH) | Temperature (Â°C) | Time        |
|------|---------|---------|-------------------|-------------|
| #1   | JJ218_A | 4.00    | 23.4              | 07/09 14:53 |
| #2   | JJ218_T | 3.99    | 23.5              | 07/09 14:54 |
| #3   | JJ218_B | 3.99    | 23.4              | 07/09 14:54 |

TESTED BY

Thiara Juliana Rizdani

qc\_analyst\_oj\_stabtest

07/09/2026