

INSTRUMENT TEST REPORT

Instrument:	Timbangan	No. Kontrol:	2QC-TMB-001
Kategori Sampel:	FINISHED_GOOD	Sampel:	LKMCA - Komix Cough Liq /30Sch
Lokasi:	R. Instrument	QC Analyst:	Desep Saepul Trio (qc_analyst_ipc)
PIC:	admin	Cal. Expiry:	20/12/2026
Start Time:	01/07/26 09:24	End Time:	01/07/26 09:27
Duration:	2 min	Total Items:	10 items

Initial Condition:

[v] Waterpass di tengah: true
[v] Timbangan Bersih: true
[v] Timbangan masih dalam ED Kalibrasi: true

Final Condition:

[v] Timbangan kembali ke 0.000 g: true

SUMMARY: Total: 10 Success: 10 (100.0%) Failed: 0 Re-reads: 0

DETAILED TEST RESULTS

Item	Batch	Weight	Time
#1	JF907	8.5461	01/07 09:25
#2	JF907	8.6436	01/07 09:25
#3	JF907	8.5146	01/07 09:26
#4	JF907	8.6574	01/07 09:26
#5	JF907	8.6915	01/07 09:26
#6	JF907	8.7605	01/07 09:26
#7	JF907	8.6586	01/07 09:26
#8	JF907	8.6959	01/07 09:26
#9	JF907	8.7759	01/07 09:26
#10	JF907	8.658	01/07 09:26

TESTED BY

Desep Saepul Trio
qc_analyst_ipc
01/07/2026