

INSTRUMENT TEST REPORT

Instrument:	Polarimeter	No. Kontrol:	2QC-POR-001
Kategori Sampel:	RM	Sampel:	Sweetness Enhancer F-6896
Lokasi:	R. Instrument	QC Analyst:	admin (superadmin)
PIC:	ayu	Cal. Expiry:	20/01/2027
Start Time:	07/04/26 13:27	End Time:	07/04/26 13:27
Duration:	1 min	Total Items:	4 items
Initial Condition:			
[v] Kondisi Awal: true			

SUMMARY: Total: 4 Success: 4 (100.0%) Failed: 0 Re-reads: 0

DETAILED TEST RESULTS

Item	Batch	SR (Å°)	AR (Å°)	TEMP (Å°C)	ISS (Å°Z)	POL ()	Time
#1	2R2603701	-	+70.248	26.9	+52.748	-	07/04 13:26
#2		-	+67.823	26.9	+52.748	-	07/04 13:26
#3		-	+67.823	26.9	+52.748	-	07/04 13:26
#4		-	+67.823	26.9	+50.927	-	07/04 13:26

VERIFICATION

TESTED BY

admin
superadmin
07/04/2026

VERIFIED BY

QC Supervisor
Date: