

INSTRUMENT TEST REPORT

Equipment Testing Results
Connection: Bridge Mode

DOCUMENT INFORMATION

Usage ID:	USG-47	Instrument:	Refractometer
Instrument Code:	2QC-RFK-002	Test Date:	15 January 2026
Operator:	admin	Total Items Tested:	32 items

Instrument Details

Instrument Name:	Refractometer
Instrument Code:	2QC-RFK-002
Type:	Equipment
Serial Number:	230801XN
Brand/Manufacturer:	Atago
Location:	R. Instrument
PIC Instrument:	ayu

Calibration Status

Calibration Date:	06 January 2026
Calibration Expiry:	06 January 2027
Status:	Valid (expires in 355 days)

Test Information

Usage ID:	USG-47
Kategori Sampel:	RM
Sampel:	Sampel RM 2
Operator:	admin (superadmin)
Start Time:	15 Jan 2026, 10:28:52
End Time:	15 Jan 2026, 10:29:50
Duration:	1 minutes
Total Items:	32 items

Initial Condition

â€¢ item_1: map[note: ok:true value:true]

Final Condition

â€œ OK

Results Summary

Total Items:	32
Successful Reads:	32
Failed Reads:	0
Re-reads:	0
Success Rate:	100.0%
Data Source:	Bridge

Detailed Results

Item #	Batch Number	Value	Unit	Read At	PC ID
#1	hk001	Brix= 79.50% nD=1.48941 t=20.0	-	15 Jan, 10:28:10	PC-Clien
#2	hk001	Brix= 79.79% nD=1.49016 t=20.0	-	15 Jan, 10:28:10	PC-Clien
#3	hk001	Brix= 77.07% nD=1.48314 t=20.0	-	15 Jan, 10:28:10	PC-Clien
#4	hk001	Brix= 76.95% nD=1.48283 t=20.0	-	15 Jan, 10:28:10	PC-Clien
#5	hk001	Brix= 79.88% nD=1.49040 t=20.0	-	15 Jan, 10:28:10	PC-Clien
#6	hk001	Brix= 79.68% nD=1.48988 t=20.0	-	15 Jan, 10:28:11	PC-Clien
#7	hk001	Brix= 79.48% nD=1.48936 t=20.0	-	15 Jan, 10:28:11	PC-Clien
#8	hk001	Brix= 76.65% nD=1.48205 t=20.0	-	15 Jan, 10:28:27	PC-Clien
#9	hk001	Brix= 77.01% nD=1.48298 t=20.0	-	15 Jan, 10:28:27	PC-Clien
#10	hk001	Brix= 76.85% nD=1.48257 t=20.0	-	15 Jan, 10:28:27	PC-Clien
#11	hk001	Brix= 76.65% nD=1.48206 t=20.0	-	15 Jan, 10:28:27	PC-Clien
#12	hk001	Brix= 77.06% nD=1.48309 t=20.0	-	15 Jan, 10:28:27	PC-Clien
#13	hk001	Brix= 76.91% nD=1.48272 t=20.0	-	15 Jan, 10:28:28	PC-Clien
#14	hk001	Brix= 76.78% nD=1.48239 t=20.0	-	15 Jan, 10:28:28	PC-Clien
#15	hk001	Brix= 76.61% nD=1.48196 t=20.0	-	15 Jan, 10:28:28	PC-Clien
#16	hk001	Brix= 76.65% nD=1.48206 t=20.0	-	15 Jan, 10:28:28	PC-Clien
#17	hk001	Brix= 67.36% nD=1.45908 t=20.0	-	15 Jan, 10:28:50	PC-Clien
#18	hk001	Brix= 65.77% nD=1.45530 t=20.0	-	15 Jan, 10:28:51	PC-Clien
#19	hk001	Brix= 65.91% nD=1.45562 t=20.0	-	15 Jan, 10:28:51	PC-Clien

TEST RESULTS (Continued)

Item #	Batch Number	Value	Unit	Read At	PC ID
#20	hk001	Brix= 65.94% nD=1.45571 t=20.0	-	15 Jan, 10:28:51	PC-Clien
#21	hk001	Brix= 65.89% nD=1.45559 t=20.0	-	15 Jan, 10:28:51	PC-Clien
#22	hk001	Brix= 65.80% nD=1.45536 t=20.0	-	15 Jan, 10:28:51	PC-Clien
#23	hk001	Brix= 65.84% nD=1.45547 t=20.0	-	15 Jan, 10:28:51	PC-Clien
#24	hk001	Brix= 49.59% nD=1.41924 t=20.0	-	15 Jan, 10:28:51	PC-Clien
#25	hk001	Brix= 49.50% nD=1.41903 t=20.0	-	15 Jan, 10:28:52	PC-Clien
#26	hk001	Brix= 79.50% nD=1.48941 t=20.0	-	15 Jan, 10:29:04	PC-Clien
#27	hk001	Brix= 79.79% nD=1.49016 t=20.0	-	15 Jan, 10:29:04	PC-Clien
#28	hk001	Brix= 77.07% nD=1.48314 t=20.0	-	15 Jan, 10:29:05	PC-Clien
#29	hk001	Brix= 76.95% nD=1.48283 t=20.0	-	15 Jan, 10:29:05	PC-Clien
#30	hk001	Brix= 79.88% nD=1.49040 t=20.0	-	15 Jan, 10:29:05	PC-Clien
#31		Brix= 79.68% nD=1.48988 t=20.0	-	15 Jan, 10:29:05	PC-Clien
#32		Brix= 79.48% nD=1.48936 t=20.0	-	15 Jan, 10:29:05	PC-Clien

Legend: (R) = Re-read | Yellow background = Re-read | Red background = Error

Document Verification

This document certifies that the instrument testing has been conducted in accordance with Good Laboratory Practice (GLP) and ALCOA principles (Attributable, Legible, Contemporaneous, Original, Accurate)

TESTED BY

admin
superadmin

15 January 2026

VERIFIED BY

QC Supervisor

Date:

ALCOA Compliance Statement

- “ Attributable: All data is attributed to the operator who performed the test
- “ Legible: All data is clearly readable and permanently recorded
- “ Contemporaneous: Data was recorded at the time of observation
- “ Original: This is the original record or a certified true copy
- “ Accurate: All data has been verified for accuracy and completeness

