

INSTRUMENT TEST REPORT

Instrument:

Timbangan

No. Kontrol:

2QC-TMB-001

Kategori Sampel:

STABTEST

Sampel:

POSKC - Zaiki Pwd

Lokasi:

R. Instrument

QC Analyst:

admin (superadmin)

PIC:

ayu

Cal. Expiry:

20/12/2026

Start Time:

25/05/26 11:18

End Time:

25/05/26 11:19

Duration:

1 min

Total Items:

3 items

Initial Condition:

[v] Waterpass di tengah: true

[v] Timbangan Bersih: true

[v] Timbangan masih dalam ED Kalibrasi: true

Final Condition:

[v] Timbangan kembali ke 0.000 g: true

SUMMARY:

Total: 3

Success: 3 (100.0%)

Failed: 0

Re-reads: 0

DETAILED TEST RESULTS

Item	Batch	Weight	Time
#1	TEST	0.0758	25/05 11:18
#2	TEST	0.0758	25/05 11:19
#3	TEST	0.0756	25/05 11:19

TESTED BY

admin

superadmin

25/05/2026

VERIFIED BY

QC Supervisor

Date: