

INSTRUMENT TEST REPORT

Instrument:

Timbangan

No. Kontrol:

2QC-TMB-001

Kategori Sampel:

STABTEST

Sampel:

POSKC - Zaiki Pwd

Lokasi:

R. Instrument

QC Analyst:

admin (superadmin)

PIC:

admin

Cal. Expiry:

20/12/2026

Start Time:

12/06/26 09:25

End Time:

12/06/26 09:27

Duration:

2 min

Total Items:

5 items

Initial Condition:

[v] Waterpass di tengah: true

[v] Timbangan Bersih: true

[v] Timbangan masih dalam ED Kalibrasi: true

Final Condition:

[v] Timbangan kembali ke 0.000 g: true

SUMMARY: Total: 5 Success: 5 (100.0%) Failed: 0 Re-reads: 0

DETAILED TEST RESULTS

Item	Batch	Weight	Time
#1	TEST_A	0.2755	12/06 09:26
#2	TEST_A	0.1263	12/06 09:27
#3	TEST_B	0.0466	12/06 09:27
#4	TEST_B	0.0467	12/06 09:27
#5	TEST_B	0.0467	12/06 09:27

TESTED BY

admin
superadmin
12/06/2026

VERIFIED BY

QC Supervisor
Date: