

INSTRUMENT TEST REPORT

Instrument:	pH Meter	No. Kontrol:	2QC-PHM-007
Kategori Sampel:	RUAH	Sampel:	LBMAT
Lokasi:	R. Instrument	QC Analyst:	Thiara Juliana Rizdani (qc_analyst_oj_stabtest)
PIC:	admin	Cal. Expiry:	03/09/2026
Start Time:	17/07/26 14:21	End Time:	17/07/26 14:25
Duration:	4 min	Total Items:	2 items
Initial Condition:			
[v] Kondisi Awal: true			
Final Condition:			
[v] Kondisi Akhir: true			

SUMMARY: Total: 2 Success: 2 (100.0%) Failed: 0 Re-reads: 0

DETAILED TEST RESULTS

Item	Batch	pH (pH)	Temperature (Â°C)	Time
#1	JG880_A	4.37	22.1	17/07 14:24
#2	JG880_B	4.36	22.1	17/07 14:25

TESTED BY

Thiara Juliana Rizdani
qc_analyst_oj_stabtest
17/07/2026