

INSTRUMENT TEST REPORT

Instrument:	pH Meter	No. Kontrol:	2QC-PHM-007
Kategori Sampel:	RUAH	Sampel:	LKMC1 - Komix Cough Liq
Lokasi:	R. Instrument	QC Analyst:	Meiriza Zamri (qc_analyst_oj_stabtest)
PIC:	admin	Cal. Expiry:	03/09/2026
Start Time:	04/08/26 14:32	End Time:	04/08/26 14:34
Duration:	3 min	Total Items:	2 items
Initial Condition: [v] Kondisi Awal: true			
Final Condition: [v] Kondisi Akhir: true			

SUMMARY: Total: 2 Success: 2 (100.0%) Failed: 0 Re-reads: 0

DETAILED TEST RESULTS

Item	Batch	pH (pH)	Temperature (Â°C)	Time
#1	JH919_A	5.34	24.3	04/08 14:33
#2	JH919_B	5.36	24.9	04/08 14:33

TESTED BY

Meiriza Zamri
qc_analyst_oj_stabtest
04/08/2026