

## **Detailed Certificate of Calibration**

Ms. Ute Pfeffer

**GERMANY** 

SCHUNK GmbH & Co. KG

Lauffen/Neckar, D-74348

ute.pfeffer@de.schunk.com

Bahnhofstraße 106 - 134

**Customer Information** 

**Equipment Calibrated** 

Six-Axis Force/Torque Sensor ATI Industrial Automation

Manufacturer: ATI Indust Serial Number: FT13240 Model: Delta Calibration: SI-660-60

Electronics: DAQ Output Range: ±10V

Description:

Gain Multiplier: 100%

Equipment Condition and Notes: Factory new.

Calibration Results: Passed

Offset: Normal Gain: Normal

Calibrated Ranges (±):

Fx	Fy	Fz	Tx	Ту	Tz
660 N	660 N	1980 N	60 N-m	60 N-m	60 N-m

Measurement Uncertainty (95% confidence level, percent of full-scale load):

Fx	Fy	Fz	Tx	Ту	Tz
1.25%	1.25%	1.50%	1.00%	1.25%	1.75%

The above Measurement Uncertainty values are the maximum amount of error for each axis expressed as a percentage of its full-scale load.

Calibration Temperature:  $22.2^{\circ}\pm1.1^{\circ} \text{ C } (72^{\circ}\pm2^{\circ} \text{ F})$ 

Temperature Compensation: hardware

Calibration Method: WI-FTP-026, DAQ Calibration Instructions

Date of Calibration: 07 Jan 2013

Certificate Date: 07 Jan 2013

Calibrated by: Melvin Wells, Calibration Technician

Page 1 of 6 Certificate Number: FT13240-20130107

El E. Well



This calibration is traceable to the National Institute of Standards and Technology (NIST). Calibration standards (not all of the listed standards are used for this calibration):

Calibration Equipment						
Description	Identification					
Calibration Tool	FTT-1927					
DAQ, 200KS/S, 16-BIT, 16 AN	CNTLR-2002-1					
DAQ, 200KS/S, 16-BIT, 16 AN	CNTLR-2015-3					
DAQ, 200KS/S, 16-BIT, 16 AN	CNTLR-2025-2					
3X STND-ALN CONTROLLER	CNTLR-2040					
3X STND-ALN CONTROLLER	CNTLR-2042					
3X STND-ALN CONTROLLER	CNTLR-2050					
PCI-6034E	CNTLR-2060-4					
30K LOAD CELL	FTT-1990					
9105-NETB BOX #2	FTT-1999					
SS Hook Weights	WHT-1700					
Weights Type II	WHT-1706					
SS Set of 13 Weights	WHT-1707					
SS Set of 13 Weights	WHT-1708					
Brass Set of 13 Weights	WHT-1709					
SS Set of 13 Weights	WHT-1720					
SS Set of 13 Weights	WHT-1730					

ATI Industrial Automation (ATI) certifies that the above product was calibrated in accordance with applicable ATI procedures. These procedures are compliant with the ISO 9001 standard to ensure that the above product is within ATI specifications. To meet this level of accuracy any loads must be correctly aligned to the transducer origin and the transducer must be mounted to a sufficiently strong surface.

To ensure the transducer measurement uncertainties listed on page 1 are met, more-conservative limits are used as calibration targets during the calibration process. If any of the calibration targets are exceeded the calibration will not pass. It is possible for a transducer to exceed these calibration targets while meeting the page 1 measurement uncertainties. The following calibration targets were used for this transducer: Fx=1.00%, Fy=1.00%, Fz=0.75%, Tx=0.75%, Ty=1.00%, and Tz=1.00%.

Note: If this is a recalibration of a legacy transducer that does not have precision locating features (such as dowel holes), there could be additional error in Tx and Ty due to inexact mounting location. Precision locating features are highly recommended for best accuracy and can be added by ATI.

This certificate shall not be reproduced except in full without written approval from ATI. This certificate only applies to the items listed and does not include unlisted ancillary items such as data acquisition equipment.

For questions or comments, please contact your ATI representative.



## Calibration Accuracy Section Sensor System FT13240, Delta/SI-660-60 Force units: N; Torque units: N-m

Calibrated Ranges (±)							
Fx Fy Fz Tx Ty Tz							
660	660	1980	60	60	60		

			Applied	Loads		
	Fx	Fy	Fz	Tx	Ту	Tz
1	0.000	422.581	0.000	-48.284	0.000	0.000
2	-422.581	0.000	0.000	0.000	-48.284	0.000
3	0.000	-422.581	0.000	48.284	0.000	0.000
4	422.581	0.000	0.000	0.000	48.284	0.000
5	0.000	556.028	0.000	-21.142	0.000	0.000
6	-556.028	0.000	0.000	0.000	-21.142	0.000
7	0.000	-556.028	0.000	21.142	0.000	0.000
8	556.028	0.000	0.000	0.000	21.142	0.000
9	0.000	311.376	0.000	-3.954	0.000	-35.578
10	0.000	311.376	0.000	-3.954	0.000	35.580
11	-311.376	0.000	0.000	0.000	-3.954	-35.597
12	-311.376	0.000	0.000	0.000	-3.954	35.568
13	0.000	-311.376	0.000	3.954	0.000	-35.580
14	0.000	-311.376	0.000	3.954	0.000	35.578
15	311.376	0.000	0.000	0.000	3.954	-35.568
16	311.376	0.000	0.000	0.000	3.954	35.597
17	0.000	0.000	556.028	-42.440	0.000	0.000
18	0.000	0.000	556.028	0.000	-42.379	0.000
19	0.000	0.000	556.028	42.350	0.000	0.000
20	0.000	0.000	556.028	0.000	42.366	0.000
21	0.000	0.000	1112.055	0.000	0.000	0.000
22	0.000	0.000	-1112.055	0.000	0.000	0.000
23	0.000	0.000	-556.028	42.440	0.000	0.000
24	0.000	0.000	-556.028	0.000	42.379	0.000
25	0.000	0.000	-556.028	-42.350	0.000	0.000
26	0.000	0.000	-556.028	0.000	-42.366	0.000

Refer to page 6 for important information on regarding this report.



	Full-Scale Error							
	Fx	Fy	Fz	Tx	Ту	Tz		
1	-0.13%	0.06%	0.16%	0.08%	0.08%	-0.26%		
2	0.20%	-0.01%	0.25%	0.06%	0.06%	-0.27%		
3	-0.14%	-0.12%	0.17%	-0.10%	0.08%	-0.28%		
4	0.37%	0.01%	0.14%	0.02%	-0.21%	-0.18%		
5	-0.10%	0.08%	0.16%	-0.09%	0.06%	-0.07%		
6	0.05%	-0.03%	0.15%	-0.01%	-0.11%	-0.24%		
7	-0.11%	-0.08%	0.17%	0.06%	0.04%	-0.13%		
8	0.25%	-0.05%	0.21%	0.01%	0.10%	-0.22%		
9	-0.04%	-0.32%	0.06%	-0.02%	0.19%	0.01%		
10	-0.04%	-0.09%	0.07%	-0.03%	-0.20%	-0.06%		
11	0.17%	-0.01%	0.04%	-0.25%	0.01%	0.03%		
12	0.23%	-0.15%	0.06%	0.27%	-0.04%	0.02%		
13	-0.03%	0.03%	0.05%	0.03%	-0.22%	0.01%		
14	-0.01%	0.09%	0.04%	-0.10%	0.25%	0.07%		
15	-0.07%	0.01%	0.08%	0.21%	0.00%	-0.06%		
16	-0.15%	-0.14%	0.07%	-0.18%	-0.04%	-0.04%		
17	-0.05%	0.06%	-0.06%	0.06%	-0.04%	-0.04%		
18	0.30%	0.10%	-0.07%	-0.02%	0.03%	0.11%		
19	-0.01%	0.19%	-0.08%	0.02%	-0.03%	-0.02%		
20	0.03%	0.10%	0.00%	-0.08%	0.01%	-0.11%		
21	-0.07%	0.21%	0.13%	-0.04%	-0.02%	0.01%		
22	0.01%	0.21%	-0.11%	0.03%	0.06%	-0.01%		
23	-0.02%	0.13%	0.06%	0.02%	-0.01%	-0.02%		
24	0.08%	0.16%	0.09%	0.03%	0.02%	0.07%		
25	0.01%	0.02%	0.06%	-0.15%	0.00%	-0.05%		
26	0.06%	0.16%	0.07%	-0.06%	-0.20%	-0.05%		

Refer to page 6 for important information on regarding this report.

	Offset Report						
	Fx	Fy	Fz	Tx	Ту	Tz	
F/T Offset	-0.9026	-0.3505	4.5917	-0.0203	-0.1809	0.0157	
	SG0	SG1	SG2	SG3	SG4	SG5	
SG Offset	0.0143	0.0042	0.0278	0.0048	0.0241	0.0080	
±SG Limit	0.4883	0.4883	<i>0.4</i> 883	<i>0.4</i> 883	0.4883	0.4883	

Offsets are measured in a unique configuration not available to the user.

Refer to page 6 for important information on regarding this report.



	Gain-Check Report					
	SG0	SG1	SG2	SG3	SG4	SG5
Lower Limit	0.7000	0.7000	0.7000	0.7000	0.7000	0.7000
Lower Output	0.8482	0.8586	0.8421	0.8218	0.8373	0.8549
Upper Output	0.8522	0.8829	0.8440	0.8452	0.8422	0.8795
Upper Limit	1.0000	1.0000	1.0000	1.0000	1.0000	1.0000

Gain readings are measured in a unique loading configuration. Refer to page 6 for important information on regarding this report.



As part of our commitment to quality, each ATI force/torque transducer undergoes rigorous accuracy testing. This process, which involves applying and verifying a rich set of loading cases designed to cover the transducer's entire six-axis calibrated range, is designed to ensure that your transducer meets the measurement uncertainties listed in this Certificate of Calibration.

Our transducers often exceed our quality standards for accuracy. Often, transducers perform exceptionally well in certain loading situations. This report summarizes the performance of your ATI F/T transducer in our factory tests. It can be thought of as a 'best-case scenario' snapshot of your transducer's performance under laboratory conditions, in a variety of loading situations. You can expect the accuracy of your transducer measurements to fall somewhere between its performance during testing and the measurement uncertainties listed on its calibration certificate.

The Calibration Accuracy Section contains several tables of data. The Calibrated Ranges (±) table lists the transducer's rated range for each axis. The Applied Loads table lists the loads applied during calibration and testing. The Full-Scale Error table shows the sensor system's measurement error as a percentage of full scale for each axis in each loading case. The Offset Report table shows transducer readings during offset adjustment and associated control limits. And the Gain-Check Report table shows verification of the transducer's sensitivity and associated control limits. If included, the Before and After Report table shows a loading case relating the transducer's performance as received to its performance after recalibration.

For best accuracy, be sure to use your transducer's precision location features, and mount your transducer to a stiff surface. If an ongoing guarantee of sensor accuracy is important to you, we recommend that your sensor be tested annually. Contact your ATI Industrial Automation distributor to schedule recalibrations.