

FibriChecker

Throws events

- onFingerDetected ()
- onFingerRemoved ()
- onHeartBeat (int *value*)
- onPulseDetected ()
- onPulseDetectionTimeExpired()
- onFingerDetectionTimeExpired()
- onMeasurementFinised ()
- onMovementDetected()
- onCalibrationReady()
- onMeasurementProcessed (MeasurementData *measurementData*)
- onMeasurementStart ()
- onSampleReady (double *ppg*, double *raw*)
- timeRemaining (int seconds)

Provides measurement after recording:

Measurement:

- **List<List<YuvList>** quadrants
- **MotionData** acc
- **MotionData** gryo
- **MotionData** rotation
- **MotionData** grav
- **List<Integer>** time
- **String** version
- **Int** heartrate
- **List<Float>** ppg

MotionData

- **List<Float>** x
- **List<Float>** y
- **List<Float>** z

YuvList

- **List<Double>** y
- **List<Double>** u
- **List<Double>** v

Initialization

[FibriChecker new];

- isFlashEnabled **default:** true
- isMovementDetectionEnabled **default:** true
- waitForStartRecordingSignal **default:** false
- movementDetectionUpperLimit **default** 14
- movementDetectionLowerLimit **default** 6
- quadrantColSize **default** 4
- quadrantRowSize **default** 4
- sampleTime **default** 60
- pulseDetectionExpiryTime **default** 10
- fingerDetectionExpiryTime **default** -1
- isGyroEnabled **default** false
- isAccelerometerEnabled **default** false
- isRotationEnabled **default** false
- isGravimetricEnabled **default** false