

Job ID: #72211

Patient Name
Steven Courtney

Patient ID
P-6235

Uploaded CT Scan File Name
CT_Scan_2021_09_01.zip

Stent Bifurcation
ETBF 28 13 C 124 E

Iliac Limb Sizing
ETLW 16 16 C 124 E

Proximal Neck Dimension
20mm

Insights

Within Green Landing Zones
Maximum Aortic Neck Dilatation

7%

IFU Violations

0 Violations

Endoleak Risk

Low Risk of Endoleaks within 6 months

Parameters of Interest

Max Displacement Force at Neck

Does not exceed Safety Factor of 10


Max Resistance Locations

[0, -400], [25, -425], [-80, -330]

Max Aorta Displacement

6.7mm at [-25, -440]

Imaging Results



Landing Zone Legend

Green

Minimal angulation, resistance and displacement for this SG configuration.

Orange

Acceptable angulation, resistance and displacement for this SG configuration.

Red

Not acceptable angulation, resistance and displacement for this SG configuration. Incompatible geometry.