



Rules Of Registration, Terms and Conditions of Contract

认证合同及条款

Legally binding agreement between ACM International Certification Limited (ACM) and the Client. Valid from the point that the Client's application for Registration is accepted by ACM.

1. Introduction

ACM is an independent certification body and these rules have been drawn up in accordance with the requirements placed on ACM by Accreditation Bodies. ACM shall abide by these requirements as specified in ISO17021 or ISO17065 and other related documents and standards. The scope of accreditation can be found on the web site of Accreditation Body. If a sector is outside the scope of accreditation then an unaccredited certificate may be offered at the discretion of ACM.

Although not a statement guaranteeing that the product or services actually meet specified requirements, accredited certification of a management system is a measure of compliance, at the time of audit, with the appropriate international standards.

2. Definitions

For the purpose of these Terms and Conditions:

Applicant; means an individual, body corporate or body unincorporated applying for a Certificate of Registration.

Audit/Assessment; means a verification of the effectiveness of the relevant management systems and processes operated by an Applicant through the examination of materials, finished product, methods of test, records, systems, environmental and other activities established by the Applicant within its Management System.

Certificate of Registration; means a document issued by ACM in recognition that the management system operated by the Applicant meets the requirements of the applicable Standard or Scheme and these Terms and Conditions.

Client; Companies, organisations or parts thereof that are audited and certified by ACM.

Major non-conformance; where there is a lack of evidence to demonstrate that a clause within a standard has been met. For example, no evidence, records or documents can be found or non-conformities raised at a previous audit have not been properly closed.

Minor non-conformance; where there is a lack of evidence to demonstrate that a clause within a standard has been fully met. For example, some evidence, records or documents are available.

Observer/Witness; means person who accompanies the audit team but does not act as an auditor. An observer can be a regulator such as the SIA or Accreditation Body such as UKAS or other interested party. In addition a technical expert may accompany an auditor but does not act as an auditor in the audit team unless qualified to do so.

Registration; The process by which Clients gain and maintain certification against a standard. Registration covers all certification activities, including all activities required outside the certification process required to maintain Registration. For example, payment terms and conditions, regulatory requirements.

Registered Company; means an individual, body corporate or body unincorporated which has been granted a Certificate of Registration.

Services; as set out in the quotation

Scope of Registration; means the range of products, services, activities covered by the management system of a registered company specified on the Certificate of Registration.

Standard; means the Standards or other requirements for which the applicant organisation is seeking certification.

3. Scope of the Rules of Registration

These rules of registration are applicable to all audits undertaken by ACM (or its local representatives) within the current scope of accreditation as detailed. In fulfilling this scope, the Client agrees to supply all necessary information to ACM so that a full and fair assessment can be undertaken. ACM will give due notice of any changes to its requirements and will verify that each certified Client complies with the new requirements. This agreement is applicable to all sites as listed on the certificate or attached schedule.

ISO 45001 requirements: - This contract requires that the certified Client informs the Certification Body, without delay, of the occurrence of a serious incident or breach of regulation necessitating the involvement of the competent regulatory authority.

4. Personnel, Impartiality and Confidentiality

ACM undertakes to provide suitably qualified and technically competent personnel for all audit and surveillance activities using their own staff or suitable qualified and competent associates. All members of ACM (full-time employees, associates or local representatives) are required to sign confidentiality agreements concerning all confidential information to which they may be exposed. They are required to declare any conflict of interest prior to an audit or as soon as they become aware of a threat to the impartiality of the audit.

The Client will be notified in advance of the lead auditor's name and who will be attending.

All information acquired by ACM about an applicant or a registered company shall be confidential and, except as required by an accreditation body or regulator or for input to required industry databases, shall not be disclosed to a third party without prior written agreement of the company concerned. ACM does reserve the right to inform the relevant authorities if breaches of legislation are found as part of the audit process.

All certification activities and the management of impartiality are overseen by the Directors of ACM. Risks to impartiality are reviewed by an Independent Board.

5. Access and Safety

The Client is responsible for ensuring that when ACM visits your site ACM has adequate protective equipment for the working environment to be assessed. Where specialist training is required this shall be disclosed to ACM at the outset. Clients shall raise these issues with ACM in advance of a visit.

To ensure an effective audit, Clients are required to make all necessary arrangements for the conduct of the audit, including access to all processes and areas within the organisation, records and personnel. This policy also applies to follow up audits following a complaint received from an interested party, for example, about the products/services provided by the certified organisation.

6. Application for Initial Registration

On receiving a completed Application Form, ACM or its local representative will prepare a quotation detailing audit cost which will be forwarded to the Client. You agree to accept these terms and conditions by returning your quotation.

On receipt of the acceptance of the quote, ACM or its local representative will issue an invoice to the Client who in turn will make payment to ACM or its local representative. The audit will then be planned and carried out in accordance with ACM accredited management system processes.

The process for initial registration normally involves an on-site stage 1 audit of the management system and a subsequent on-site stage 2 audit of the implementation of that system.

It is the responsibility of the applicant to satisfy themselves that the proposed scope of registration meets their requirements. Changes to the scope following acceptance of the quotation could result in amendment to the fees.

7. Changes to your Organisation



Rules Of Registration, Terms and Conditions of Contract

认证合同及条款

It is a condition of this agreement that you inform ACM of any substantive changes that take place or are planned to take place in or to your structure, strategy, management or culture (a "change") which could affect your certificate.

Without limitation, a change may include any merges, acquisitions, takeovers, new brand using existing people, cases of significant growth or reduction in employee numbers and changes in name or ownership. You must inform ACM prior to the change taking place or, if this is not possible, as soon as is reasonably practicable following the change.

ACM reserves the right to investigate the nature of any change and its impact on your certification. Wherever possible, this investigation shall be timed to take place with a routine assessment of your certification. There may be a requirement to re-quote in line with the new changes and you will be notified of this prior to the assessment taking place.

8. Audit Method

The first stage of the audit requires the Auditor to conduct an on-site readiness review of the Client's management system to assess the documentation and if the implementation of the management system is at a level sufficient to progress to the Stage two audit.

When satisfied with the compliance of the documentation and system readiness the Auditor will produce a report and will agree a date with the Client for the Stage 2 audit. The Stage 2 audit will then be conducted in accordance with ACM accredited management system processes. If further visits are required, due to non-compliances found, these will be undertaken and the Client will be liable for any extra charges incurred based on the current audit day rate in the quotation. The audit will be carried out against agreed audit criteria.

Once registration has been obtained the Client will be under a duty to notify ACM of any changes that significantly affect the registration. ACM may re-audit if necessary (short notice) due to the significant changes or as a result of any complaint, or follow up on suspended clients. If any additional charges are incurred ACM reserve the right to pass these additional charges on to the Client.

ISO 45001 requirements: If the Client provides services at another organisation's premises, we will verify that the client's OH&SMS covers these offsite activities (notwithstanding the OH&SMS obligations of the other organization). In determining the time to be spent for audit, we shall consider to audit periodically any organization site where these employees work. Whether all sites shall be audited will depend on various factors such as OH&S risks associated with the activities therein performed, contract agreements, being certified by another accredited, internal audit system, statistics on accidents and near misses.

Audit methodology will include interviews, observation of activities, electronic and hard copy document and record review. Conclusions will be based on the evidence obtained during the audit. The auditor(s) will use sampling techniques to obtain the evidence and no guarantee can be given that a different conclusion may have been reached previously or if a different sample had been taken.

9. Certification

On completion of the audit the Auditor will submit the report with a recommendation for either granting or refusing registration. A suitably qualified, independent and authorised member of the ACM team will make the certification decision and authorise the issue of the certificate.

Any non-conformities raised during an audit must be corrected and closed prior to a certification decision being made. The Client agrees to meet the extra visit or documentation review costs involved in closing out the non-conformities.

Certificates will not be issued unless payment in full has been received. The certificate remains the property of ACM and is valid for three years, providing the client maintains the management system to the required standard.

10. Annual Registration, Surveillance, Re-certification and Special / Short Notice Visits

After the issue of a certificate, planned surveillance visits will be carried out by an Auditor at the Client's premises at least annually in order to maintain registration. If areas of concern are identified, more visits may be carried out at the discretion of the ACM Management. Major non-conformities identified at a surveillance visit must be corrected by the Client within the timescales agreed with the auditor and closed out by ACM either by document / record review or additional site visit. The Client agrees to meet the extra costs relating to any additional close out work required.

Minor non-conformities identified at a surveillance must be corrected in the agreed timescale. The action taken will be checked at the next scheduled audit. ACM may require the Client to send evidence to confirm that effective action has been taken. Failure to address minor non-conformities may result in escalation to a major non-conformance and in such cases the Client agrees to meet extra costs relating to additional close out work.

Following a Stage 2 audit Clients must undertake their first surveillance before the first anniversary date of their certificate. If this is not completed on time for any reason the certificate will be suspended. Following suspension, failure to complete within three months will result in certificate withdrawal.

Re- Certification: A certificate will normally be valid for a period of three years (subject to ongoing compliance as detailed in these terms and conditions) A re-certification audit is required to maintain registration. The Client will notify ACM without delay of any matters incurred that may affect the capability of the management system to fulfil requirements.

The re-certification audit must take place prior to the expiry of the certificate. Should the audit identify any non-conformities these must be corrected by the Client within the agreed timescale agreed with the auditor and subsequently closed out by ACM.

Special Visits: Work required to close out improvements needs or non-conformities on, for example, additional assessments may require additional revisit audits. The Client agrees to meet the additional costs and these will be quoted by ACM in advance.

In addition, if a complaint is made about a certified Client, or we have received notification of any changes we reserve the right to undertake a special / revisit audit to investigate whether the company is still meeting the requirements of the appropriate standard they are approved to. The client agrees to meet these additional costs.

Independently from the involvement of the competent regulatory authority, a special / revisit audit may be necessary in the event that ACM becomes aware that there has been a serious incident related to occupational health and safety, for example, a serious accident, or a serious breach of regulation, in order to investigate if the management system has not been compromised and did function effectively. ACM shall document the outcome of its investigation.

Information on incidents such as a serious accident, or a serious breach of regulation necessitating the involvement of the competent regulatory authority, provided by the certified Client or directly gathered by the audit team during the special / revisit audit, shall provide grounds for ACM to decide on the actions to be taken, including a suspension or withdrawal of the certification, in cases where it can be demonstrated that the system seriously failed to meet the OH&S certification requirements.

11. Extension or Revision to the Scope of Registration

This may be applied for in the same way as the initial audit, indicating the increased scope of registration required. The audit will be carried out in the areas not previously audited. If successful, a new certificate indicating the new scope will be issued by ACM. There will be a charge for extensions to scope and the re-issuing of the certificate. Should a certified scope be reduced, the Client will amend all advertising matter.



Rules Of Registration, Terms and Conditions of Contract

认证合同及条款

If a Client is already registered to a particular standard, for example ISO 45001 and subsequently wishes to add an additional standard such as ISO 9001, this will be treated as a new application and the Rules of Registration relating to initial audits shall be followed.

12. Publicity & Privacy

When a certificate has been issued, the Client has the right to publish the fact. The relevant accreditation marks and logos can be used on stationery and websites relating only to the audited scope of registration and standard. ACM will make available on request, the certification status of any client.

The Client must not make or permit any misleading statement regarding its certification, the scope and/or standards covered, the Client's locations covered by certification. Any references to a Client's management system must not imply that ACM certifies a product, service or process.

The Client must not apply certification marks or accreditation marks to laboratory test, calibration or inspection reports as such reports are deemed to be products.

ACM will contact the Client in relation to organising their certification and to inform/provide information or facts of relevance to their certification.

13. Logo and Certificate Misuse

ACM will take all necessary steps to ensure that there is no misuse of the accreditation marks and logos or the certificate by the Client. The Client undertakes not to misuse or misrepresent the marks and logos or the certificate in any way.

14. Fees

ACM will always quote for any work in advance of delivery and ensure Clients are fully aware of their obligations.

All fees are strictly non-refundable and will be paid in accordance with the terms indicated on the invoice.

All fees for audits and annual registration are reviewed annually and are available on request. All fees are subject to VAT and are strictly non-refundable.

Annual fees are payable in advance of audit dates and are to maintain registration.

Fees include mileage and travel but not accommodation which will be arranged separately when required unless included as part of the quotation and noted as included.

Fees include the issue of 1 certificate per standard for new certifications or re-certification cycles. For extra copies required for multiple sites, or at surveillance stages for transfers, an additional cost is incurred.

Additional fees are payable where ACM is required to close out non conformities raised during an audit. These can be additional document reviews, special visits or on-site audits.

Fees for re-instatement of Registration following a period of suspension or withdrawal will be incurred. ACM will specify the fees required in advance along with any conditions.

Certificates will not be issued to Clients unless payment has been received.

Instalments are only permissible if the previous year's fees have been paid in full and on time.

Certificates re-issued as a result of changes to the Client's circumstance like change of address, company name. etc., at the client request, shall attract a nominal administration fee per certificate.

The fee payable for cancellation will be on a sliding scale based on a percentage of ACM's current day rate for certification days quoted for the job. ACM may cancel a planned audit which may also result in Registration being suspended or withdrawn if the Client fails to make payment in full on the due date.

15. Certification Suspension, Withdrawal or Restoring Certification

On completion of the audit, to the appropriate standards or specifications, the certificate:

May be suspended due to:

- Continued misuse of certification marks and / or ACM logos.
- Failure to apply corrective action as a result of non-conformities found at during an audit or surveillance visits.
- Failure to allow an audit to be conducted as planned, e.g. surveillance.
- Breaches in legislation relevant to the scope of Registration.
- Failure to pay on time.
- Any other breach of ACM's Rules of registration and Terms and Conditions of Contract.

Or withdrawn due to:

- Failure to respond to requests made by ACM after suspension of certificate.
- At the Client's request.
- Failure to pay.
- Ceasing to carry on business or threatening to do so, application for insolvency or bankruptcy petition presented against it, entry into voluntary or compulsory liquidation or having a receiver, administrator appointed over its assets.

Following either the suspension or withdrawal of a certificate the client will discontinue its use and return the original certificate to ACM and discontinue to claim accredited management system registration.

If the Client continues to claim certification following withdrawal then ACM maintains the right to report the Client to the relevant legal authority and to take appropriate legal action.

Certificates cannot be extended beyond the expiry date and if it does expire the consequences will be explained. Following withdrawal, the certificate can be restored if the Client re-applies within the first 6 months of withdrawal, subject to technical review and completion of the re-certification audit process. After 6 months the Client will need to re-apply and at least be subject to a stage 2 audit. The stage 1 audit can be waived if a technical review confirms no significant changes have occurred since the last audit.

ISO 45001 requirements:- Information on incidents such as a serious accident, or a serious breach of regulation necessitating the involvement of the competent regulatory authority, provided by the certified Client or directly gathered by the audit team during the special audit, shall provide grounds for the Certification Body to decide on the actions to be taken, including a suspension or withdrawal of the certification, in cases where it can be demonstrated that the system seriously failed to meet the OH&S certification requirements.

16. Cancellation of Planned Audits

ACM is committing resource when agreeing audit dates. Future audit dates are normally agreed between the auditor and Client during an audit and will be recorded in the report. Also, ACM and the Client will agree dates which will be confirmed by ACM in writing. It is not our policy to issue reminders and it is the Client's responsibility to ensure that the audit can be delivered in accordance with the plan.

Where ACM incurs a cost for technical experts, accreditation body observers as result of a Client postponement or cancellation these fees will be charged at cost to the Client.

17. Appeals Procedure



Rules Of Registration, Terms and Conditions of Contract

认证合同及条款

If for any reason a Client disagrees with the Auditor's verdict they are at liberty to lodge an appeal with ACM's Managing Director. All appeals will be held in the presence of an Independent Board. The Board will hear evidence from the Client's representative and the relevant Auditor. The decision of the Independent Board is final and binding on both the Client and ACM. No counter claims will be allowed by either party. No costs, for whatever reason, will be allowed for either party as a result of an appeal. Expenses of the appeal will be met in full by the party who has the decision against them.

18. Client Complaints against ACM Personnel

If a Client has reason for a complaint this should be submitted in writing to ACM. Documented procedures for handling complaints can be found on the ACM website. All complaints will be responded to in line with the timelines set out in our procedures and will be investigated thoroughly. Results will be sent in writing to the Client. Following investigation, if the Client is not satisfied with the outcome it will be referred to a Director or the Client may be asked to lodge an appeal.

19. Complaints against the Registered Company

The Client agrees to make available to ACM or its representative all information pertaining to complaints received by the client from customers, regulators and any other interested parties. In the event of an alleged breach of relevant legislative or regulatory requirement the ACM Client/registered company must inform ACM as soon as possible.

20. Audit Team

ACM will provide an appropriately qualified, competent and impartial audit team or individual auditor to deliver the agreed audit plan. In addition, the auditor / audit team may be accompanied by a technical expert, translator or interpreter. The Client has the right to object to any individual auditor but must do so immediately upon notification of the auditor/audit team. ACM reserves the right to change the assigned auditor(s) or add additional auditors to meet operational requirements.

The Client does not have the right to require that a specific named auditor conducts a particular audit.

To maintain impartiality ACM will review from time to time the frequency of audits undertaken by an auditor with a particular Client. If ACM considers that impartiality is threatened due to familiarity then it may be necessary to change the auditor.

From time to time the auditor / audit team may be accompanied by a trainee. A trainee has no status on the audit and will at all times be under the supervision of a lead auditor. Clients will not be charged for a trainee and advance permission for attendance will always be sought from the Client.

The Client in agreeing the Rules of Registration also agrees that no pressure, intimidation or inducement will be offered to ACM auditors or staff designed to change or alter a decision. All instances will be reported to ACM.

21. Witnessed Audits by Accreditation Bodies

It is a condition of these Rules of Registration that all ACM certificated Clients should, if requested, allow representatives of accreditation or regulatory bodies such as UKAS to witness ACM staff carrying out audits. Failure to allow this could jeopardise the Client's registration. The presence of any observers will be agreed prior to any audit. Observers will have no influence over the outcome of the audit.

22. Terms of Payment

Charges for services will be those agreed in advance with the Client in the quotation and any additional charges as outlined in the Rules of Registration. A quotation will only be deemed to be valid when ACM receives the acceptance agreeing to these Rules and Contract Terms.

All fees are payable in advance of audits for initial and re-certification. Annual registration fees will be invoiced prior to anniversary date of the stage two assessment and should be paid in full within 30 days from receipt of invoice or earlier if the anniversary date is within this period.

ACM reserves the right to carry out an annual review of charges and shall endeavour to notify the Client of any resulting changes to the charges at least 30 days prior to implementation.

The Client may cancel this contract giving notice no later than one month prior to delivery of service. The Rules of Registration for suspension or withdrawal will apply.

Cancellation of this contract, the audit or surveillance dates by the Client within twenty working days of the agreed dates will result in a cancellation fee. This will be on a sliding scale, based on a percentage of the quoted days.

ACM will be unable to deliver services if the Client fails to pay in accordance with the invoice terms and Rules of Registration.

23. Liability

Neither ACM nor any of its servants or agents shall be liable for any loss, expense or damage however so sustained by any company, Client or person due to any act whatsoever taken by ACM or its servants or agents, save to the extent that any attempted exclusion or liability would be contrary to law. All work and information obtained during an audit will be treated as confidential and not disclosed to any third party unless as required by law or under the terms of our Accreditation Bodies.

All warranties and conditions and other terms implied by statute or common law, are to the fullest extent permitted by law, excluded by this agreement.

The Client acknowledges and accepts that where services are required to change to comply with Accreditation Bodies guidelines or any other legal requirements, ACM shall not be liable for breach under this Agreement.

Nothing in this agreement limits or excludes the liability of ACM for:

- Death or personal injury resulting from negligence; or
- Any damage or liability incurred by the Client as a result of fraud, or fraudulent misinterpretation by ACM.

ACM shall not be liable for:

- Loss of profit;
- Loss of business;
- Depletion of goodwill and/or similar losses;
- Loss of contract
- Loss or corruption of data or information or
- Any special indirect, consequential or pure economic loss, costs, damages, charges or expenses;

ACM's total liability in contract, (including negligence or breach of statutory duty), misinterpretation, restitution, or otherwise arising in connection with the performance of this agreement shall be limited to the Services Fee.

24. Indemnity

The Client will indemnify and keep indemnified ACM against all and any claims or losses, proceedings, lost profits, damages, awards, expenses, claims and costs suffered by ACM as a result of misuse or misrepresentation by the Client of any certification marks, logos, approval or registration given to the client by ACM under these Rules of Registration Terms and Conditions.

25. Entire Agreement

This Contract constitutes the whole agreement between the parties and supersedes all previous agreements between the parties relating to its subject matter.



Rules Of Registration, Terms and Conditions of Contract

认证合同及条款

Each Party acknowledges that in entering into this Contract it has not relied upon any representation, inducement promise or agreement, whether oral, written or otherwise made by or on behalf of the other Party and which has not been incorporated within or specifically referred to in this Contract, and that no other agreement, statement or promise not contained in this Contract shall be valid or binding.

26. Governed Law and Jurisdiction

This Contract shall be governed and construed in accordance with the Laws of England and Wales. The Parties hereby submit to the exclusive jurisdiction and procedure of the English Courts.

27. Notice and Communication

Any notice to be given under this Agreement shall be in writing and shall be deemed to have been duly given if left at or sent by first class post, registered post or facsimile or other electronic media (email is deemed an acceptable form of communication) to a party at the address or relevant telecommunication number for such party or such other address as the party may from time to time designate by written notice to the other.

Any notice or other document shall be deemed to have been received by the addressee two working days following the date of despatch of the notice or other document by post or, where the notice or other document is sent by hand or is given by facsimile or other electronic media simultaneously with the delivery or transmission.

28. Changes to Terms and Conditions

No variation of the contract shall be valid unless it is in writing and signed by duly authorised representatives of both parties.

A person who is not a party to this contract shall not have any rights under or in connection with it.

Both parties shall comply, and shall ensure that each of their sub-contractors, agents and personnel comply with any relevant and applicable anti bribery and corruption laws, regulations and/or legislation related to delivery and receipt of the services.

29. Anti- Bribery and Corruption

The Client warrants and represents to ACM that it complies with the Bribery Act 2010 and that it has not and shall not, in connection with the services contracted, make, promise or offer to make any offer of payment or transfer of anything of value, directly or indirectly; i) to any government official or to any intermediary for payment to a government official, or ii) to any political party for the purpose of influencing any act or decision of such official or securing an improper advantage to assist ACM in obtaining or retaining business. Failure by the Client to comply with this clause shall constitute a breach of contract

30. Confidentiality

Both parties shall keep in strict confidence and treat the other parties confidential information as confidential and use it only for the purposes of the contract except in so far as may be necessary for the performance of any obligations of the contract or to the extent that such information is generally available to the public or to the extent that disclosure of information is required to be made by law.

Each party agrees that this obligation shall continue in force without limitation in point of time irrespective of the termination of the contract for any reason but shall cease to apply to information once it enters into the public domain. It shall also cease to apply to information which is received independently from another source without the imposition of any duty of confidence.

31. Force Majeure

Neither party shall have any liability to the other if it is prevented from, or delayed in performing, its obligations under the contract, or from carrying on its business by any event(s) or combination of events where such event(s) arises from, or is attributable to acts, events, omissions or accidents beyond the reasonable control of the relevant party including, but not limited to, acts of God, terrorism, war or flood (force majeure event). In such circumstances the time for performance shall be extended by a period equivalent to the period during which performance of the obligation has been delayed or failed to be performed due to the force majeure event.

32. Contact Permission

By agreeing to these terms and conditions you give permission for ACM to contact you by email or telephone for the purpose of certification or certification training.

Company 组织名称	Hebei Xiong'an Shanglong Medical Technology Co.,Ltd.	Date 日期	2026-05-27
-----------------	--	------------	------------

一阶段审核记录 Stage 1 Audit Record

Requirements 要求	Audit Record 审核记录
组织是否形成文件的管理手册，包括：管理体系的范围，任何删减和不适用的详细说明和理由？ Has the organization developed a documented management manual, including the scope of the management system, detailed explanations and reasons for any deletions or inapplicability?	☒ 是 yes ☐ 否 no 说明 / description:
组织管理体系运行是否超过三个月？ Has the management system been in operation for more than three months?	☒ 是 yes ☐ 否 no 说明 / description:
组织是否编制了标准中明确要求形成的文件？ Has the organization written documents that explicitly require the formation of standards?	☒ 有 exist ☐ 没有或不全 None or incomplete 说明 / description:
体系文件与受审核组织实际情况是否相符，特别是体系文件中描述的产品和服务、部门设置和职责与权限、生产和服务过程等 Whether the system documents match the actual situation of the audited organization, especially the products and services described in the system documents, department settings and responsibilities and authorities, production and service processes, etc	☒ 相符 consistent ☐ 不相符 inconsistent 说明 / description:
体系文件是否受控？ Are management system documents under control?	☒ 是 yes ☐ 否 no 说明 / description:
是否明确了组织的架构与职能？ Has the organizational structure and functions been clearly defined?	☒ 是 yes ☐ 否 no 说明 / description:
组织地址是否和申报的一致？	☒ 相符 consistent ☐ 不相符 inconsistent

Requirements 要求	Audit Record 审核记录
Is the organizational address consistent with the declared address?	说明 / description:
组织是否有多场所，多场所是否纳入管理体系范围？ Does the organization have multiple sites, and are multiple sites included in the scope of the management system?	☒ 没有 None ☐ 有 exist，已纳入管理体系范围，见多现场清单 Already included in the scope of the management system, see multiple sites lists
组织人数是否和申报的一致？ Is the number of organizational members consistent with the declared number?	☒ 相符 consistent ☐ 不相符 inconsistent 说明 / description:
是否充分识别了适用的法规和标准要求？ Has the applicable regulatory requirements been fully identified?	☒ 是 yes ☐ 否 no 说明 / description:
行政许可证、资质证书、强制性认证证书 Administrative license, qualification certificate, mandatory certification certificate	☐ 不需要 N/R ☒ 需要并已提供 need/already provided: 第一类医疗器械生产备案凭证，备案编号：冀雄安药监生产备 20260001 号，生产范围：2002 分类目录； 2017 分类目录：19-03-助行器械；备案部门：河北雄安新区管理委员会，备案日期：04/02/2026； 第一类医疗器械备案信息表，备案编号：冀雄安械备 20260001，产品名称：手动轮椅车，备案部门：河北雄安新区管理委员会，备案日期：19/01/2026； ☐ 需要但未提供 Required/not provided 说明 / description:
是否建立了设计开发控制程序，并按照规定要求保留各阶段的设计开发资料？	☐ 不需要 N/R ☒ 是 yes ☐ 否 no 说明/description:
是否编制了医疗器械文档？ Have medical device documents been prepared?	☒ 是 yes ☐ 否 no 说明/description:

Requirements 要求	Audit Record 审核记录
是否依据标准对产品建立了风险管理过程，包括风险管理计划、风险管理报告、风险控制措施、综合剩余风险等？	☒ 是 yes ☐ 否 no 说明/description:
是否建立了上市后监管/警戒系统程序，并按要求履行向监管机构报告的义务？包括产品警戒、忠告性通知/召回、不良事件报告等？	☒ 不需要 N/R ☐ 是 yes ☐ 否 no 说明/description:
组织的现场（或多个现场）位置、布局和条件，请概述 Please provide an overview of the location(s), layout, and operational conditions of the organization's site(s).	企业成立于 2007 年成立，注册地位于河北省雄安新区雄县张青口村一村 1 组 141-2 号，企业自有厂房（见村委证明），总面积约 6000 平方米， 公司为产品实现及质量管理体系运行提供了必要的资源，满足管理体系的有效运行和产品的实现。
产品或服务的策划：包含生产/服务流程，外包过程，关键过程，需要确认的过程 Planning of products or services: including production/service processes, outsourcing processes, critical processes, and processes that require confirmation	产品生产工艺流程： 关键过程：成品组装 特殊过程：无 外包过程：无
人力资源概况 Overview of Human Resources	公司组织架构配置了适用的人力资源，如配置了各部门经理及生产作业人员、产品检验人员和相应的办公室人员，人力资源配置满足管理体系的有效运行和产品的实现。
设施设备概况 Overview of Facilities and Equipment	基础设施：办公环境、生产车间、搬运设施、消防设施等。 生产设备：风批、电钻、调试台、空压机、自动捆包机等。 检测设备：百分表、数显卡尺、钢卷尺、数显角度尺等，满足来料、过程及成品检验需求。 生产车间道路畅通、物品堆放整齐、区域明确、照明通风良好。现场配备了灭火器等消防设施，厂区大院严禁吸烟，工作环境基本符合规定要求。 未提供储气罐压力表和安全阀的有效校验证据。
最高管理层是否展示了对管理体系的领导力和承诺，请概述 Has the top management demonstrated leadership and commitment to the management system? Please overview	公司高层主要通过以下方面充分展示了领导力和承诺： 1、 制定和明确了公司方针、目标，并确保与组织的环境和战略方向一致； 2、 确保在公司内部充分有效实施管理体系，将管理体系融入到公司的各管理环节； 3、 在公司内部积极沟通和宣导管理体系的有效实施的重要性； 4、 为管理体系的有效实施提供必须的资源； 5、 建立公司的组织架构明确公司各管理层的职责和权限等。

Requirements 要求	Audit Record 审核记录
方针、目标策划与概况 Policy, Goal Planning and Overview	公司高层依据标准要求策划明确了公司方针和目标。 组织的质量方针：满足顾客需求是我们长期的质量方针；持续质量改进是公司发展的根本动力； 公司质量目标： 1) 顾客满意度≥95 分； 2) 客户投诉率≤2%； 3) 产品一次检验合格率≥98%； 各部门目标： 综合部：员工培训计划完成率 100%；账物一致率 100%等； 质量部：检测工作按时完成率≥95%；监视和测量设备检定/校准及时率 100%等； 生产部：设备维护保养完成率 100%；生产计划完成率 100%等； 供销部：采购计划按时完成率 100%；顾客满意度≥95 分等； 公司目标的制定与方针保持了一致、可测量，并在公司各职能层次进行了分解和沟通，定期考核和评价。 公司通过内审和管理评审对方针和目标的适宜性、有效性进行评审。
组织的架构与职能概况 Overview of Organizational Structure and Functions	公司高层依据标准要求和公司有效管理的需求建立了有效组织架构，并明确了各职能层次的职责和权限，包括管理层供销部、生产部、质量部、综合部等，并形成了文件。
是否进行了内审并提供了内部审核资料（如内审计划、首末次会议签到记录、现场审核记录、不符合报告、内审报告）,请概述： Have internal audits been conducted and internal audit documents (such as the audit plan, sign-in records of the opening and closing meetings, on-site audit records, nonconformity reports, and the internal audit report) been provided? Please provide a summary:	公司策划编制了《内部审核程序》，明确了内审实施的具体要求，包括内审的频次（每年一次）、内审计划、审核组的选择、内审报告编制、不符合报告等要求，与最高管理者面谈沟通和查阅相关证据获取管理评审信息： 1、内审时间/审核组：2026.05.15-16，组长：尚月永，组员：王倩（生产部经理身份）； 2、覆盖范围/内审方式：ISO 13485:2016 标准及相关法律法规所覆盖的所有部门，本次内审覆盖了组织管理体系适用的全部条款和活动，采用集中式按部门审核。 3、内审组能力与公正性：内审组成员在实施内审前接受了内审实施要求的培训，并接受了审核员任命，内审组成员为公司内部正式员工，且审核组成员未审核本部门的工作。 4、本次审核期间核查的证据信息：查见《内部审核计划》、《内部审核检查表》、《内审不合格报告》和《内部质量审核报告》。 5.见内审计划明确了组内的分工并经管理代批准，查见内审签到表，总经理/管代和各部门负责人均参与了本次内审。 6.见生产部检查表，涵盖条款6.3/6.4/7.1/7.3/7.5.1/7.5.2...，内审检查表覆盖了部门所有条款 7.本次内审发现一项一般不符合，涉及条款 7.5.8 标识，不符合报告中明确了原因分析、纠正及纠正措施，并于 2026 年 5 月 17 日经验证关闭。 查内审报告，对本次的内审给出了整体评价，并经管理者代表审核、总经理批准下发至各部门及最高管理者作为管理评审输入的信息。 组织的内审实施符合标准要求。
是否进行了管理评审并提供了管理评审资料（如管理评是计划、会议签到记录、管理评审输入和输出信息、改进等）,请概述：	公司依据标准要求策划编制了《管理评审控制程序》，明确了管理评审实施频次（每年一次）、管理评审输入、管理评审输出及改时实施等要求。与最高管理者面谈沟通和查阅相关证据获取管理评审信息，如： 1、管理评审时间/主持者：2026 年 5 月 25 日，主持：尚月永 2、参会人员：查见签到表，参会人员包括：董晶晶、尚月兴、王倩、尚悦茂等 3、管理评审输入信息：查见《管理评审计划》，明确了管理评审输入信息，如：

Requirements 要求	Audit Record 审核记录
Has a management review been conducted and management review documents (such as the management review plan, meeting sign-in records, input and output information of the management review, and improvement actions) been provided? Please provide a summary:	a) 质量方针、目标的实施情况及适宜性评价； b) 公司的组织架构、职责分配、资源配置是否适宜； c) 本年度内部审核和外部审核的结果及其有效性； d) 纠正和预防措施的实施情况； e) 上次管理评审跟踪措施的完成情况； f) 投诉处置； g) 向监管机构的报告； 4、管理评审输出信息：查见《管理评审报告》，列出了本次管理评审输出信息，并提出了改进的问题和改进措施，如： a) 各级领导和员工的质量意识有待进一步提高，尤其是全员质量意识； b) 因人员短缺的原因，部分档案记录、管理不完善； c) 对体系运行过程相关信息收集不够； 管理评审结论：查见《管理评审报告》，批准：尚月永，对公司的管理体系的运行进行了综合的评价，并得出评价结论：质量管理体系是适宜的、其运行是有效的；本组织的质量方针、目标是能满足市场形势发展需要的；管理评审进一步完善了我公司的质量管理体系，在实际工作中有力地推动了我公司质量管理体系持续、有效运行。

Process Area 过程/区域	GM/MR	Responsible Person 负责人	Shang Yueyong 尚月永/Wang Qian 王倩
Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	Conformity and coherence of proposed certification scope (Containing description of medical device type with reference to Technical Areas in IAF MD9 Annex 1 or intended use) 拟认证范围的一致性和相关性 (包含参考 IAF MD9 附件 1 中的“技术领域”或预期用途的医疗设备类型说明)	A.1.1 :无源医疗器械 组织的认证范围为：手动轮椅车的生产和销售，产品为国内 I 类无源医疗器械，凭备案资质经营。组织备案的手动轮椅车由车架、前轮、后轮(含手圈)、座垫、靠垫、制动器、安全带、腿托带、手刹、扶手、护板、脚托架、脚托组成。以乘坐者手驱动或护理者手推为动力，具有四个车轮，组织的认证范围与实际一致。	
	Qualification and licensing status of the organization (including the name, number and validity period) 组织的资质、许可状态（包括名称、编号和有效期）？	第一类医疗器械生产备案凭证，备案编号：冀雄安药监生产备 20260001 号，生产范围：2002 分类目录；2017 分类目录：19-03-助行器械；备案部门：河北雄安新区管理委员会，备案日期：04/02/2026； 第一类医疗器械备案信息表，备案编号：冀雄安械备 20260001，产品名称：手动轮椅车，备案部门：河北雄安新区管理委员会，备案日期：19/01/2026；	
	Applicable Regulatory requirements (with reference to the Roles of the organization) List the requirements. Evidence of conformity to standard requirements to be filled in the checklist 适用的法规要求（参考组织的角色） 列出要求。符合标准要求的证据要填写在检查清单中	《医疗器械监督管理条例》； 《医疗器械生产监督管理办法》； 《医疗器械召回管理办法》； 《GB/T 42061-2022 医疗器械质量管理体系用于法规的要求》； 《医疗器械注册与备案管理办法》； 《GB/Z 13800-2021 手动轮椅车》；	
4.1	4.1General Requirements 总要求 4.1.1 The organization shall document a quality management system and maintain its effectiveness in accordance with the requirements of this International Standard and applicable regulatory requirements. The organization shall establish, implement and maintain any requirement, procedure, activity or arrangement required to be documented by this International Standard or applicable regulatory requirements. The organization shall document the role(s) undertaken by the organization under the applicable regulatory requirements. NOTE Roles undertaken by the organization can include manufacturer, authorized representative, importer or distributor. 组织应按本标准的要求和适用的法规要求，将质量管理体系形成文件并保持其有效性。 组织应建立、实施和保持本标准或适用法规所要求形成的文件的任何要求、程序、活动或安排。 组织应对在适用的法规要求下组织所承担的职能形成文件。 注：组织承担的职能包括生产商、授权代表、进口商或经销商	河北雄安尚龙医疗科技有限公司成立于 2007 年 4 月 29 日，注册地位于河北省雄安新区雄县张青口村一村 1 组 141-2 号 企业从事手动轮椅车的生产和销售，凭资质经营。企业建立了医疗器械质量管理体系的文件和记录。质量手册中，组织基于风险方法识别了手动轮椅车的使用过程方法对设计和生产活动进行监控，通过设定过程目标确保体系有效运行。 企业明确了组织需要应对的风险，通过和管理层交谈了解到，识别的风险如：生产能力，新技术、新设备、新工艺、市场容量、竞争力、价格风险，客户的需求等。风险控制措施如：质量部收集相关标准、拟定产品可行性方案进行产品风险分析，负责生产质量方面的信息，规避产品风险等。目前风险可控。 组织编制了《风险管理控制程序》，判断本公司的产品有关的危害，估计和评估风险，控制这些风险，并监控控制的有效性；提供《手动轮椅车风险管理报告》，针对手动轮椅车开展风险管理，识别各类潜在危害，评估风险等级并落实管控措施，对风险进行了分析和分级，对安全性特征进行了判断等，结论：风险管理计划已被适当的实施，全部剩余风险均处于可接受范围内，且收益超过风险等。 建立了《文件控制程序》，该程序文件对文件进行有效的控制，尤其是外来文件，特别是法律、法规等，文件控制程序加以管理。建立了记录控制程序，描述了如何确保记录是可读的，易于识别和跟踪。文件由综合部以纸制版发放并进行管理。 外包过程：无；	
	4.1.2 The organization shall: a) determine the processes needed for the quality management system		

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	and the application of these processes throughout the organization taking into account the roles undertaken by the organization; b) apply a risk-based approach to the control of the appropriate processes needed for the quality management system; c) determine the sequence and interaction of these processes. 组织应： a) 确定在所承担职能下质量管理体系所需的过程及其在整个组织的应用； b) 采用基于风险的方法控制质量管理体系所需的适当的过程。 c) 确定这些过程的顺序和相互作用。 4.1.3 For each quality management system process, the organization shall: a) determine criteria and methods needed to ensure that both the operation and control of these processes are effective; b) ensure the availability of resources and information necessary to support the operation and monitoring of these processes; c) implement actions necessary to achieve planned results and maintain the effectiveness of these processes; d) monitor, measure as appropriate, and analyze these processes; e) establish and maintain records needed to demonstrate conformance to this International Standard and compliance with applicable regulatory requirements (see 4.2.5). 对各质量管理体系过程，组织应： a) 确定为保证这些过程的有效运行和控制所需的准则和方法； b) 确保可以获得必要的资源和信息，以支持这些过程的运作和监视； c) 实施必要的措施，以实现对这些过程策划的结果并保持这些过程的有效性； d) 监视、测量（适用时）和分析这些过程； e) 建立并保持为证实符合本国际标准和适用的法规要求的记录（见4.2.5）。 4.1.4 The organization shall manage these quality management system processes in accordance with the requirements of this International Standard and applicable regulatory requirements. Changes to be made to these processes shall be: a) evaluated for their impact on the quality management system; b) evaluated for their impact on the medical devices produced under this quality management system; c) controlled in accordance with the requirements of this International Standard and applicable regulatory requirements. 组织应按本标准和适用的法规要求来管理这些质量管理体系过程。这些过程的变更应： a) 评价变更对质量管理体系的影响； b) 评价变更对依照本质量管理体系所生产的医疗器械的影响； c) 依据本标准和适用的法规要求进行控制。 4.1.5 When the organization chooses to outsource any process that affects product conformity to requirements, it shall monitor and ensure	关键过程：成品组装； 特殊过程：无； 组织明确其角色为手动轮椅车的生产及销售，企业采购零配件进行组装、调试、检验，完成产品交付。 建立了《计算机软件应用确认控制程序》，企业用于质量管理体系生产和服务过程的计算机软件如畅捷通 T+ V13.0 版；微打标签打印软件 2.0.2019.920 版等；	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	control over such processes. The organization shall retain responsibility for conformity to this International Standard and to customer and applicable regulatory requirements for outsourced processes. The controls shall be proportionate to the risk involved and the ability of the external party to meet the requirements in accordance with 7.4. The controls shall include written quality agreements. 当组织选择将影响产品符合要求的任何过程外包时，应监视和确保对这些过程的控制。组织应保留外包过程符合本标准、顾客要求和适用的法规要求的责任。控制应与所涉及的风险和外部方满足7.4要求的能力相适应。控制应包含书面的质量协议。 4.1.6 The organization shall document procedures for the validation of the application of computer software used in the quality management system. Such software applications shall be validated prior to initial use and, as appropriate, after changes to such software or its application. The specific approach and activities associated with software validation and re-validation shall be proportionate to the risk associated with the use of the software. Records of such activities shall be maintained (see 4.2.5). 组织应将用于质量管理体系的计算机软件应用的确认程序形成文件。在软件首次使用前应对软件应用进行确认，适当时，软件变更后或应用时进行确认。 与软件确认和再确认有关的特定方法和活动应与软件使用相关的风险相适应。应保留这些活动的记录。（见 4.2.5）。		
4.2.1	4.2 Documentation requirements 文件要求 4.2.1 General 总则 The quality management system documentation (see 4.2.4) shall include: a) documented statements of a quality policy and quality objectives; b) a quality manual; c) documented procedures and records required by this International Standard; d) documents, including records, determined by the organization to be necessary to ensure the effective planning, operation, and control of its processes; e) other documentation specified by applicable regulatory requirements. 质量管理体系文件（见 4.2.4）应包括： a) 形成文件的质量方针和质量目标， b) 质量手册； c) 本标准所要求形成文件的程序和记录； d) 组织确定的为确保其过程有效策划、运作和控制所需的文件，包括记录； e) 适用的法规要求规定的其他文件。	组织根据企业需求及标准要求，建立了医疗器械质量管理体系文件，覆盖标准标准要求的全部文件类型。具体包括： 1. 质量方针和质量目标（质量手册附件二）； 2. 《质量手册》（XASL-QM, A/0）； 3. 标准要求的程序文件，如：《文件控制程序》SBBJ-QP-001, A/0； 4. 确保过程有效运行所需的作业指导书等三级文件：如《工作环境管理制度》；《返工管理制度》；《手动轮椅车成品检验作业指导书》等； 5. 体系运行过程记录及外来文件。 上述文件均通过编号、版本等方式进行受控管理，文件清单完整可查。	
4.2.2	4.2.2 Quality manual 质量手册 The organization shall document a quality manual that includes: a) the scope of the quality management system, including details of and	企业按照标准的要求编制了质量手册，质量手册范围对质量管理体系的范围和任何删减的内容进行了详细的描述。并对质量管理体系形成的文件进行了说明。概述了质量管理体系的文件结构。	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	justification for any exclusion or non-application; b) the documented procedures for the quality management system, or reference to them; c) a description of the interaction between the processes of the quality management system. The quality manual shall outline the structure of the documentation used in the quality management system. 组织应形成文件的质量手册, 包括: a) 质量管理体系的范围, 包括任何删减和不适用的详细说明和理由 ; b) 为质量管理体系建立的形成文件的程序或对其引用 ; c) 质量管理体系过程之间的相互作用的表述。 质量手册应概述质量管理体系的文件结构。	业务范围: 手动轮椅车的生产和销售; 物理范围: 河北省雄安新区雄县张青口村一村 1 组 141-2 号; 删减条款如下: 1. 产品为组装后成品不存在安装活动, 简单折叠展开后即可使用, 删减 7.5.3 安装活动。 2. 产品未涉及到植入医疗器械, 删减 7.5.9.2 植入性医疗器械的专项要求。 3. 产品未涉及到需灭菌的医疗器械, 删减 7.5.5 无菌医疗器械的专用要求 4. 产品未涉及到需灭菌的医疗器械, 删减 7.5.7 无菌和无菌屏障系统的过程确认的专用要求 以上删减合理。	
4.2.3	4.2.3 Medical device file 医疗器械文档 For each medical device type or medical device family, the organization shall establish and maintain one or more files either containing or referencing documents generated to demonstrate conformity to the requirements of this International Standard and compliance with applicable regulatory requirements. The content of the file(s) shall include, but is not limited to: a) general description of the medical device, intended use/purpose, and labelling, including any instructions for use; b) specifications for product; c) specifications or procedures for manufacturing, packaging, storage, handling and distribution; d) procedures for measuring and monitoring; e) as appropriate, requirements for installation; f) as appropriate, procedures for servicing. 组织应对每个医疗器械类型或医疗器械族建立一个或多个文档, 文档包含或引用形成的文件以证明符合本标准要求 and 适用法规要求。 文件的内容应包括但不限于 : a) 医疗器械的概述、预期用途/预期目的和标记; 包括所有使用说明; b) 产品规范 ; c) 制造、包装、贮存、处置和流通的规范或程序; d) 测量和监视程序; e) 适当时, 安装要求; f) 适当时, 服务程序。	组织规定了医疗器械文件的建立和管理要求。针对手动轮椅车的生产和销售建立了完整的医疗器械文档, 具体包括: 1. 《风险管理控制程序》; 2. 《生产和服务提供控制程序》; 3. 《监视和测量装置控制程序》; 4. 《过程和产品监视测量程序》; 5. 《数据分析控制程序》; 6. 《纠正预防措施控制程序》; 经查, 上述文件均已受控, 与产品实际生产情况一致, 可满足产品质量保证和追溯的要求。	
5.1	5 MANAGEMENT RESPONSIBILITY 管理职责 5.1 Management commitment 管理承诺 Top management shall provide evidence of its commitment to the development and implementation of the quality management system and maintenance of its effectiveness by: a) communicating to the organization the importance of meeting customer as well as applicable regulatory requirements; b) establishing the quality policy; c) ensuring that quality objectives are established;	与最高管理者王倩 (副总经理) 沟通确认, 管理层通过质量策划证明了对质量管理体系建立和实施的承诺, 质量管理策划确保了质量管理体系的应用及适用的法规要求在组织中得到传达。 管理层为保持质量管理体系有效性的承诺, 开展了以下活动: 制定质量方针: 满足顾客需求是我们长期的质量方针; 持续质量改进是公司发展的根本动力。 1. 确保制定质量目标: 通过制定、统计和分析公司及各职能层次的质量目标, 确保目标可测量、与方针一致, 并定期检查达成情况;	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	d) conducting management reviews; e) ensuring the availability of resources. 最高管理者应通过以下活动，对其建立、实施质量管理体系并保持其有效性的承诺提供证据： a) 向组织传达满足顾客和法律法规要求的重要性， b) 制定质量方针， c) 确保制定质量目标； d) 进行管理评审， e) 确保资源的获得。	2. 建立质量管理体系：按照标准要求建立并实施质量管理体系，王倩任管理者代表负责体系的维护和管理。 3. 提供必要资源：包括包括人力资源（培训、再培训）、基础设施（厂房、组装工具、检验工具）、工作环境及信息资源等，以确保体系有效运行。 4. 组织管理评审：最高管理者按计划的时间间隔组织管理评审，对质量管理体系的持续适宜性、充分性和有效性进行评价，并保留评审记录。 5. 关注顾客需求：该组织通过实现质量目标来关注客户需求，如客户投诉率≤2%，顾客满意度≥95分，以增强顾客满意。	
5.2	5.2 Customer focus以顾客为关注焦点 Top management shall ensure that customer requirements and applicable regulatory requirements are determined and met with the aim of enhancing customer satisfaction 最高管理者应确保顾客要求和适用的法规要求得到确定并满足		
5.3	5.3 Quality policy质量方针 Top management shall ensure that the quality policy: a) is applicable / appropriate to the purpose of the organization; b) includes a commitment to comply with requirements and continually improve / maintain the effectiveness of the quality management system; c) provides a framework for establishing and reviewing quality objectives; d) is communicated and understood within the organization; e) is reviewed for continuing suitability. 最高管理者应确保品质方针： a) 与组织的宗旨相适应； b) 包括对满足要求和保持质量管理体系有效性的承诺； c) 提供制定和评审质量目标的框架， d) 在组织内得到沟通和理解； e) 在持续适宜性方面得到评审。		
5.4.1	5.4 Planning策划 5.4.1 Quality objectives质量目标 Top management shall ensure that quality objectives, including those needed to meet applicable regulatory requirements and requirements for products, are established at relevant functions and levels within the organization. The quality objectives shall be measurable and consistent with the quality policy. 最高管理者应确保在组织的相关职能和层次上建立质量目标，质量目标包括满足适用的法规要求和产品要求所需的内容。质量目标应是可测量的，并与质量方针保持一致。	与最高管理者王倩（副总经理）沟通确认，组织制定了适宜的质量目标，并在公司的各职能层次进行目标分解。查看组织形成文件的质量目标： 公司质量目标： 1) 顾客满意度≥95分； 2) 客户投诉率≤2%； 3) 产品一次检验合格率≥98%； 各部门目标： 综合部：员工培训计划完成率 100%；账物一致率 100%等； 质量部：检测工作按时完成率≥95%；监视和测量设备检定/校准及时率 100%等； 生产部：设备维护保养完成率 100%；生产计划完成率 100%等； 供销部：采购计划按时完成率 100%；顾客满意度≥95分等； 提供《2026 质量目标统计表》1-6 月，目标均达成，批准：尚月永，日期：2026 年 06 月 30 日； 组织在质量管理体系策划过程中，确保体系变更得到识别和控制。当质量目标或过程发生变更时，组织对变更进行评审和风险评估，确保体系的完整性不受影响。 与管理者代表沟通，组织在质量管理体系策划过程中，确保体系变更得到识别和控制。当质量目标或过程发生变更时，组织对变更进行评审和风险评估，确保体系的完整性不受影响。与管理者代表沟通，目前无重大体系变更。	
5.4.2	5.4.2 Quality management system planning质量管理体系策划 Top management shall ensure that: a) the planning of the quality management system is carried out in order to meet the requirements given in 4.1, as well as the quality objectives; b) the integrity of the quality management system is maintained when changes to the quality management system are planned and implemented. 最高管理者应确保： a) 对质量管理体系进行策划，以满足质量目标以及 4.1 的要求； b) 在对质量管理体系的变更进行策划和实施时，保持质量管理体系的完整		

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	性。		
5.5.1	5.5 Responsibility, authority and communication 职责、权限与沟通 5.5.1 Responsibility and authority 职责与权限 Top management shall ensure that responsibilities and authorities are defined, documented and communicated within the organization. Top management shall document the interrelation of all personnel who manage, perform and verify work affecting quality and shall ensure the independence and authority necessary to perform these tasks. 最高管理者应确保职责、权限得到规定、形成文件并在组织内沟通。 最高管理者应确定所有从事对质量有影响的管理、执行和验证工作的人员的相互关系形成文件，并确保其完成这些任务所必要的独立性和权限。	质量手册明确了公司的组织架构及各部门的职责与权限，涉及部门包括：生产部，供销部，质量部，综合部，经抽查组织策划的职责和权限，明确了上述部门的职责要求和权限，各部门主要职责例如：生产部负责： 1) 质量管理体系纠正预防措施的实施； 2) 配合生产车间及时处理生产过程中出现的技术问题； 3) 负责在贯彻执行公司的管理方针和管理目标； 4) 负责生产过程的控制等； 各岗位均制定了岗位说明书，明确了具体职责和任职要求。组织通过入职培训、会议等方式，确保各层级人员理解并履行其职责。	
5.5.2	5.5.2 Management representative 管理者代表 Top management shall appoint a member of management who, irrespective of other responsibilities, has responsibility and authority that includes: a) ensuring that processes needed for the quality management system are documented, implemented and maintained; b) reporting to top management on the effectiveness and performance of the quality management system and any need for improvement; c) ensuring the promotion of awareness of customer requirements, applicable regulatory requirements and quality management system requirements throughout the organization. 最高管理者应指定一名管理者，无论该成员在其他方面的职责如何，应具有以下方面的职责和权限： a) 确保质量管理体系所需的过程形成文件； b) 向最高管理者报告质量管理体系的有效性和任何改进的需求； c) 确保在整个组织内提高满足适用的法规要求和质量管理体系要求的意识。	质量手册任命管理者代表：王倩，其主要负责推动质量管理体系的有效运行，并对体系运行过程进行日常维护与持续改进；定期向总经理汇报质量管理体系的运行成效，并提出需要改进的方向和建议；通过培训、宣贯等方式，提升全体员工对客户要求和法规要求的认知与重视；代表公司就质量管理体系相关事项，与外部监管机构进行沟通联络。 管理者代表通过管理评审、专题会议等形式，定期向最高管理者汇报体系的运行状况。	
5.5.3	5.5.3 Internal communication 内部沟通 Top management shall ensure that appropriate communication processes are established within the organization and that communication takes place regarding the effectiveness of the quality management system. 最高管理者应确保在组织内建立适当的沟通过程，并确保对质量管理体系的有效性进行沟通。	组织建立了内部沟通机制，通过会议、通知、电话、网络、公告、培训及相关报表等多种方式，向内部各层级传达质量信息，确保质量管理体系的有效性得到充分沟通。	
5.6	5.6 Management review 管理评审 5.6.1 General 总则 The organization shall document procedures for management review. Top management shall review the organization's quality management system at documented planned intervals to ensure its continuing suitability, adequacy and effectiveness. The review shall include assessing opportunities for improvement and the need for changes to the quality management system, including the quality policy and quality objectives. Records from management reviews shall be maintained (see 4.2.5). 组织将管理评审形成文件。最高管理者应按形成文件的策划时间间隔评审组	公司依据标准要求策划编制了《管理评审控制程序》，明确了管理评审实施频次（每年一次）、管理评审输入、管理评审输出及改时实施等要求。与最高管理者面谈沟通和查阅相关证据获取管理评审信息，如： 1、管理评审时间/主持者：2026 年 5 月 25 日，主持：尚月永 2、参会人员：查见签到表，参会人员包括：董晶晶、尚月兴、王倩、尚悦茂等 3、管理评审输入信息：查见《管理评审计划》，明确了管理评审输入信息，如： a) 质量方针、目标的实施情况及适宜性评价； b) 公司的组织架构、职责分配、资源配置是否适宜； c) 本年度内部审核和外部审核的结果及其有效性；	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	织的质量管理体系，以确保其持续的适宜性、充分性和有效性。评审应包括评价质量管理体系改进的机会和变更的需要，包括质量方针和质量目标。应保持管理评审的记录（见 4.2.5）。 5.6.2 Review input 评审输入 The input to management review shall include, but is not limited to, information arising from: a) feedback; b) complaint handling; c) reporting to regulatory authorities; d) audits; e) monitoring and measurement of processes; f) monitoring and measurement of product; g) corrective action; h) preventive action; i) follow-up actions from previous management reviews; j) changes that could affect the quality management system; k) recommendations for improvement; l) applicable new or revised regulatory requirements. 管理评审的输入应包括，但不限于由以下方面产生的信息： a) 反馈；b) 投诉处置；c) 向监管机构的报告；d) 审核；e) 过程的监视和测量；f) 产品的监视和测量；g) 纠正措施，h) 预防措施；i) 以往管理评审的跟踪措施，j) 影响质量管理体系的变更，k) 改进的建议；l) 适用的新的或修订的法规要求。 5.6.3 Review output 评审输出 The output from management review shall be recorded (see 4.2.5) and include the input reviewed and any decisions and actions related to: a) improvement needed to maintain the suitability, adequacy, and effectiveness of the quality management system and its processes; b) improvement of product related to customer requirements; c) changes needed to respond to applicable new or revised regulatory requirements; d) resource needs. 管理评审的输出应形成记录（见 4.2.5），包括以下方面有关的输入评审和任何决定和措施： a) 保持质量管理体系及其过程适宜性、充分性和有效性所需的改进， b) 与顾客要求有关的产品的改进； c) 为响应适用的新的或修订的法规要求所需的变更； d) 资源需求。	d) 纠正和预防措施的实施情况； e) 上次管理评审跟踪措施的完成情况； f) 投诉处置； g) 向监管机构的报告； 4、管理评审输出信息：查见《管理评审报告》，列出了本次管理评审输出信息，并提出了改进的问题和改进措施，如： a) 各级领导和员工的质量意识有待进一步提高，尤其是全员质量意识； b) 因人员短缺的原因，部分档案记录、管理不完善； c) 对体系运行过程相关信息收集不够； 管理评审结论：查见《管理评审报告》，批准：尚月永，对公司的管理体系的运行进行了综合的评价，并得出评价结论：质量管理体系是适宜的、其运行是有效的；本组织的质量方针、目标是能满足市场形势发展需要的；管理评审进一步完善了我公司的质量管理体系，在实际工作中有力地推动了我公司质量管理体系持续、有效运行。管理评审计划、会议签到表、管理评审报告是否由相应人员亲笔签名？ ☒是 ☐否，已要求相应人员亲笔签名	
6.1	6.1 Provision of resources 资源提供 The organization shall determine and provide the resources needed to: a) implement the quality management system and to maintain and continually improve its effectiveness; b) meet applicable regulatory and customer requirements and enhance customer satisfaction. 组织应确定并提供以下方面所需的资源：	企业成立于 2007 年成立，注册地位于河北省雄安新区雄县张青口村一村 1 组 141-2 号，企业自有厂房（见村委证明），总面积约 6000 平方米，公司为产品实现及质量管理体系运行提供了必要的资源，具体包括： 基础设施：办公环境、生产车间、搬运设施、消防设施等。 生产设备：风批、电钻、调试台、空压机、自动捆包机等。 检测设备：百分表、数显卡尺、钢卷尺、数显角度尺等，满足来料、过程及成品检验需求。	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	a) 实施质量管理体系并保持其有效; b) 满足适用的法规和顾客要求	生产车间道路畅通、物品堆放整齐、区域明确、照明通风良好。现场配备了灭火器等消防设施, 公司通过内部审核和管理评审,定期对资源的充分性和适宜性进行评价,确保资源持续满足体系运行及产品实现的要求。	
8.1	8 MEASUREMENT, ANALYSIS & IMPROVEMENT 测量、分析和改进 8.1 General 总则 The organization shall plan and implement the monitoring, measurement, analysis and improvement processes needed to: a) demonstrate conformity of production their requirements; b) ensure conformity of the quality management system; c) maintain and continually improve the effectiveness of the quality management system. This shall include determination of appropriate methods, including statistical techniques, and the extent of their use. 组织应策划并实施以下方面所需的监视、测量、分析和改进过程: a) 证实产品的符合性 ; b) 确保质量管理体系的符合性; c) 保持质量管理体系的有效性。 这应包括对统计技术在内的适当方法及其应用程度的确定 。	企业编制了下列程序文件对监视、测量、分析和改进过程进行策划: 1. 制定《生产和服务提供控制程序》及《过程和产品监视测量程序》,对生产过程及最终产品进行检验,确保产品符合要求,并保留检验记录。 2. 制定《内部审核程序》和《管理评审控制程序》,定期开展内审和管理评审,评价质量管理体系的符合性和有效性,识别改进机会,确保体系持续有效运行,并保留相关记录。 3. 制定《数据分析控制程序》,运用适宜的统计技术对质量数据进行分析,形成分析报告,为改进提供依据。 4. 制定《纠正预防措施控制程序》,对不合格及潜在不合格采取纠正预防措施,实现持续改进。 上述监视测量活动均保留了相应记录。	
8.2.1	8.2 Monitoring and measurement 监视和测量 8.2.1 Feedback – Customer satisfaction 反馈 As one of the measurements of the effectiveness of the quality management system, the organization shall gather and monitor information relating to whether the organization has met customer requirements. The methods for obtaining and using this information shall be documented. The organization shall document procedures for the feedback process. This feedback process shall include provisions to gather data from production as well as post-production activities. The information gathered in the feedback process shall serve as potential input into risk management for monitoring and maintaining the product requirements as well as the product realization or improvement processes. If applicable regulatory requirements require the organization to gain specific experience from post-production activities, the review of this experience shall form part of the feedback process. 作为对质量管理体系有效性的一种测量,组织应收集和监视组织是否已满足顾客要求的信息,并应将获取和利用这种方法形成文件。 组织应将反馈过程程序形成文件。该反馈过程应包括从生产和生产后活动中收集数据的规定 。 从反馈过程中收集的信息应用作监视和保持产品要求的风险管理的潜在输入以及产品实现或改进过程的潜在输入。 如果适用法规要求组织从生产后活动获取特定经验,则对该经验的评审构成反馈过程的一部分。	组织编制了《顾客满意度测量控制程序》、《顾客抱怨处理程序》等文件,对生产后信息进行监视和管理。 组织识别的顾客相关要求的信息来源包括市场需求信息、客户投诉、售后服务记录、生产过程数据以及监管机构发布的同类产品质量公告等。 供销部定期对收集的反馈信息进行汇总和统计分析,识别趋势和潜在风险,形成分析报告,作为产品实现、改进及生产后阶段风险管理活动的输入。当分析发现新的危险情况时,生产部负责对风险管理文档进行评审和更新。 组织对用户反馈信息的进行监视,用于规范处理客户对于顾客要求的监视,对顾客不满意信息进行处置。 组织明确了客户满意度调查的要求,定期对客户进行调查,评分项目:产品质量,价格,售后服务,包装及防护的符合程度,交货期及时率交付及时性,售后服务等,见“顾客满意度调查表”如: 1) 宁波雅舸国际贸易有限公司,总分 95 分,调查日期:2026 年 5 月 15 日 2) 深圳猫胡子信息科技有限公司,总分 97 分,调查日期:2026 年 5 月 15 日 达成顾客满意度目标。 企业目前针对生产后信息的监控暂无发现新的风险。	
8.2.2	8.2.2 Complaint handling 投诉处理	组织编制了《顾客抱怨处理程序》《顾客满意度测量控制程序》,对用户反馈信息的	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	The organization shall document procedures for timely complaint handling in accordance with applicable regulatory requirements. These procedures shall include at a minimum requirements and responsibilities for: a) receiving and recording information; b) evaluating information to determine if the feedback constitutes a complaint; c) investigating complaints; d) determining the need to report the information to the appropriate regulatory authorities; e) handling of complaint-related products; f) determining the need to initiate corrections or corrective actions. If any complaint is not investigated, justification shall be documented. Any correction or corrective action resulting from the complaint handling process shall be documented. If an investigation determines activities outside the organization contributed to the complaint, relevant information shall be exchanged between the organization and the external party involved. Complaint handling records shall be maintained (see 4.2.5). 组织应按照适用的法规要求将及时处理抱怨的程序形成文件。这些程序应至少包括以下方面的最低要求和职责： a) 接收和记录信息； b) 评价信息以确定反馈是否构成投诉， c) 调查投诉； d) 确定是否需要向适当的监管机构报告信息； e) 处置与投诉相关产品； f) 确定是否需要启动纠正或纠正措施。 如果有任何投诉未经调查，应记录理由。应记录由投诉处置过程形成的纠正或纠正措施。 如果一项调查确定是组织外的活动导致了投诉，则应在组织和所涉及的外部方之间交换相关信息。应保留投诉处置记录（见4.2.5）。	进行监视，用于规范处理客户对于顾客要求的监视，对顾客不满意信息进行处置。 自体系运行以来，组织暂时未收到产品的投诉信息。	
8.2.3	8.2.3 Reporting to regulatory authorities向监管机构报告 If applicable regulatory requirements require notification of complaints that meet specified reporting criteria of adverse events or issuance of advisory notices, the organization shall document procedures for providing notification to the appropriate regulatory authorities. Records of reporting to regulatory authorities shall be maintained (see 4.2.5). 如果适用法规有要求，将符合不良事件报告准则的投诉、发布的忠告性通知报告，则组织应将向有关的监管机构报告的程序形成文件。 应保留向监管机构报告的记录（见4.2.5）。	组织编制了《医疗器械不良事件报告和再评价程序》、《忠告性通知和产品召回控制程序》，适用于公司产品的忠告性事件的信息发布、反馈和处理。 产品 26 年初完成备案，自体系建立以来，产品未曾发布忠告性通知或召回通知。	
8.2.4	8.2.4 Internal audit内部审核 The organization shall conduct internal audits at planned intervals to determine whether the quality management system: a) conforms to planned and documented arrangements, requirements of	公司策划编制了《内部审核程序》，明确了内审实施的具体要求，包括内审的频次（每年一次）、内审计划、审核组的选择、内审报告编制、不符合报告等要求，与最高管理者面谈沟通和查阅相关证据获取管理评审信息： 1、内审时间/审核组：2026.05.15-16，组长：尚月永，组员：王倩（生产部经理身	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	this International Standard, quality management system requirements established by the organization, and applicable regulatory requirements; b) is effectively implemented and maintained. The organization shall document a procedure to describe the responsibilities and requirements for planning and conducting audits and recording and reporting audit results. An audit program shall be planned, taking into consideration the status and importance of the processes and area to be audited, as well as the results of previous audits. The audit criteria, scope, interval and methods shall be defined and recorded (see 4.2.5). The selection of auditors and conduct of audits shall ensure objectivity and impartiality of the audit process. Auditors shall not audit their own work. Records of the audits and their results, including identification of the processes and areas audited and the conclusions, shall be maintained (see 4.2.5). The management responsible for the area being audited shall ensure that any necessary corrections and corrective actions are taken without undue delay to eliminate detected non-conformities and their causes. Follow-up activities shall include the verification of the actions taken and the reporting of verification results. NOTE Further information can be found in ISO 19011. 组织应按策划的时间间隔进行内部审核以确定质量管理体系是否： a) 符合策划并形成文件化的安排、本标准的要求以及组织所确定的质量管理体系的要求和适用的法规要求； b) 得到有效实施与保持。 组织应策划审核方案，策划时应考虑拟审核的过程和区域的状况和重要性以及以往的审核结果，应规定和记录审核的准则、范围、间隔和方法（见4.2.5），审核员的选择和审核的实施应确保审核过程的客观和公正。审核员不应审核自己的工作。 应保留审核和审核结果的记录。（见 4.2.5）。 负责受审区域的管理者应确保采取任何必要的纠正和纠正措施，应无不当拖延，以消除所发现的不合格及其原因。后续活动应包括验证所采取的措施并报告验证结果。 注： 进一步信息见 ISO 19011 。	份）； 2、覆盖范围/内审方式：ISO 13485:2016 标准及相关法律法规所覆盖的所有部门，本次内审覆盖了组织管理体系适用的全部条款和活动，采用集中式按部门审核。 3、内审组能力与公正性：内审组成员在实施内审前接受了内审实施要求的培训，并接受了审核员任命，内审组成员为公司内部正式员工，且审核组成员未审核本部门的工作。 4、本次审核期间核查的证据信息：查见《内部审核计划》、《内部审核检查表》、《内审不合格报告》和《内部质量审核报告》。 5.见内审计划明确了组内的分工并经管理代批准，查见内审签到表，总经理/管代和各部门负责人均参与了本次内审。 6.见生产部检查表，涵盖条款6.3/6.4/7.1/7.3/7.5.1/7.5.2...，内审检查表覆盖了部门所有条款 7.本次内审发现一项一般不符合，涉及条款 7.5.8 标识，不符合报告中明确了原因分析、纠正及纠正措施，并于 2026 年 5 月 17 日经验证关闭。 查内审报告，对本次的内审给出了整体评价，并经管理者代表审核、总经理批准下发至各部门及最高管理者作为管理评审输入的信息。 组织的内审实施符合标准要求。 首末次会议签到表、内审计划、不符合项报告、内审报告是否由相应人员亲笔签名？ ■是 □否，已要求相应人员亲笔签名	
8.4	8.4 Analysis of data数据分析 The organization shall document procedures to determine, collect and analyse appropriate data to demonstrate the suitability, adequacy and effectiveness of the quality management system and to evaluate where continual improvement of the effectiveness of the quality management system can be made. The procedures shall include determination of appropriate methods, including statistical techniques and the extent of their use. The analysis of data shall include data generated as a result of monitoring and measurement and from other relevant sources and include, at a minimum, input from: a) feedback and customer satisfaction;	企业编制了《数据分析控制程序》，通过数据的采集、分析和统计技术的应用，揭示生产过程中显示和隐藏的规律，提供可靠的质量改进信息，以保证质量管理体系的有效运行及持续改进。 提供《目标分解及考核报表》30/06/2026，周期：2026 年 1-6 月，质量目标及分解目标均达成，批准：尚月永。	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	b) conformity to product requirements; c) characteristics and trends of processes and products, including opportunities for improvement and preventive actions; d) suppliers; e) audits; f) service reports, as appropriate. If the analysis of data shows that the quality management system is not suitable, adequate or effective, the organization shall use this analysis as input for improvement as required in 8.5. Records of the results of analyses shall be maintained (see 4.2.5). 组织应将确定、收集和分析适当的数据的程序形成文件以证实质量管理体系的适宜性、充分性和有效性。这些程序应包含统计技术在内的适当方法及其使用程度的确定。 数据分析应包括由监视测量所产生的数据和其他相关来源的数据，至少包括以下方面的输入： a) 反馈； b) 与产品要求的符合性； c) 过程和产品的特性及趋势，包括改进的机会， d) 供方； e) 审核； f) 适当时，服务报告 如果数据分析表明质量管理体系不是适宜的，充分的或有效的，组织应按8.5 的要求将此分析结果作为改进的输入。 应保留数据分析结果的记录（见 4.2.5）。		
8.5.1	8.5 Improvement改进 8.5.1 General – Continual improvement总则—持续改进 The organization shall continually improve and identify and implement any changes necessary to ensure and maintain the continued suitability, adequacy and effectiveness of the quality management system as well as medical device safety and performance through the use of the quality policy, quality objectives, audit results, post-market surveillance, analysis of data, corrective actions, preventive actions and management review. 组织应利用质量方针、质量目标、审核结果、上市后监督、数据分析、纠正和预防措施以及管理评审来识别和实施任何必要的更改，以确保和保持质量管理体系的持续适宜性、充分性和有效性以及医疗器械的安全和性能。	总经理和管理者代表利用工作会议、监视质量方针和质量目标达成情况、内部审核、数据分析、纠正和预防措施以及管理评审等方法，识别和实施任何必要的改进，使质量管理体系不断完善，以确保质量管理体系的持续适宜性和有效性。	
8.5.2	8.5.2 Corrective action纠正措施 The organization shall take action to eliminate the cause of non-conformity in order to prevent recurrence. Any necessary corrective actions shall be taken without undue delay. Corrective actions shall be proportionate to the effects of the non-conformity encountered. The organization shall document a procedure to define requirements for: a) reviewing non-conformism (including customer complaints); b) determining the causes of non-conformism; c) evaluating the need for action to ensure that non-conformism do not recur;	企业建立了《纠正预防措施控制程序》，总经理和管理者代表利用工作会议、监视质量方针和质量目标达成情况、内部审核、数据分析、纠正和预防措施以及管理评审等方法，识别和实施任何必要的改进，使质量管理体系不断完善，以确保质量管理体系的持续适宜性和有效性。 如针对 16/05/2026 内审开具的不符合项（7.5.8），组织采取了纠正措施和预防措施，不符合项经验证可以关闭，验证日期：17/05/2026，验证人：尚月永。	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	d) planning and documenting action needed and implementing such action, including, as appropriate, updating documentation; e) verifying that the corrective action does not adversely affect the ability to meet applicable regulatory requirements or the safety and performance of the medical device; f) reviewing the effectiveness of corrective action taken. Records of the results of any investigation and of action taken shall be maintained (see 4.2.5). 组织应采取措施，以消除不合格的原因以防止不合格的再发生。组织应采取任何必要的纠正措施。纠正措施应与不合格的影响程度相适应； 组织应编制形成文件的程序，以规定以下方面的要求： a) 评审不合格（包括投诉）； b) 确定不合格的原因； c) 评价确保不合格不再发生的措施的需求； d) 对所需的措施进行策划、形成文件并实施，适当时，包括更新文件； e) 验证纠正措施对满足适用的法规要求的能力和对医疗器械的安全和性能无不良影响； f) 评审所采取的纠正措施的有效性。 应保留任何调查的结果和所采取措施的记录（见4.2.5）。		
8.5.3	8.5.3 Preventive action 预防措施 The organization shall determine action to eliminate the causes of potential non-conformity in order to prevent their occurrence. Preventive actions shall be proportionate to the effects of the potential problems. The organization shall document a procedure to describe requirements for: a) determining potential non-conformism and their causes; b) evaluating the need for action to prevent occurrence of non-conformism; c) planning and documenting action needed and implementing such action, including, as appropriate, updating documentation; d) verifying that the action does not adversely affect the ability to meet applicable regulatory requirements or the safety and performance of the medical device; e) reviewing the effectiveness of the preventive action taken, as appropriate. Records of the results of any investigations and of action taken shall be maintained (see 4.2.5). 组织应确定措施消除潜在不合格的原因以防止不合格的发生。预防措施应与潜在问题的影响程度相适应。 组织应将说明以下方面要求的程序形成文件： a) 确定潜在不合格及其原因， b) 评价防止不合格发生的措施的需求； c) 对所需的措施进行策划、形成文件并实施，适当时，包括更新文件， d) 验证预防措施对满足适用的法规要求的能力或医疗器械的安全和性能无不良影响；	企业建立了《纠正预防措施控制程序》，对于不合格或潜在不合格采取纠正和预防措施，以消除不合格或潜在不合格的原因，防止不合格和潜在不合格的发生。预防措施的来源包括数据分析、风险分析、过程监控、客户反馈趋势等。 如针对 16/05/2026 内审开具的不符合项（7.5.8），组织采取了纠正措施和预防措施，不符合项经验证可以关闭，验证日期：17/05/2026，验证人：尚月永。 经查阅相关记录并与管理者代表沟通，目前暂未发现需要采取预防措施的潜在不合格情况。因此，本年度暂无预防措施实施记录。	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	e) 适当时，评审所采取的预防措施的有效性。 应保留任何调查的结果和所采取措施的记录（见 4.2.5）。		
	Management system changes/logo usage/last non conformance verification /complaints from relevant parties 管理体系变更/标志使用/上次不符合项验证/相关方投诉（监审、再认证）（公信网核查组织是否被列入严重违法失信名单）	企业管理体系没有变更； 初次认证一阶段开具的问题项，经验证已经关闭； 暂不涉及标志使用； 暂无收到有关相关方的投诉； 经过信用中国网站查询，组织未被列入严重违法失信名单；	

Process Area 过程/区域	行政人事文件管理	Responsible Person 负责人	Dong Jingjing 董晶晶
-----------------------	----------	---------------------------	-------------------

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
6.2	6.2 Human Resources 人力资源 Personnel performing work affecting product quality shall be competent on the basis of appropriate education, training, skills and experience. The organization shall document the process(es) for establishing competence, providing needed training, and ensuring awareness of personnel. The organization shall: a) determine the necessary competence for personnel performing work affecting product quality; b) provide training or take other actions to achieve or maintain the necessary competence; c) evaluate the effectiveness of the actions taken; d) ensure that its personnel are aware of the relevance and importance of their activities and how they contribute to the achievement of the quality objectives; e) maintain appropriate records of education, training, skills and experience (see 4.2.5). NOTE The methodology used to check effectiveness is proportionate to the risk associated with the work for which the training or other action is being provided. 基于适当的教育、培训、技能和经验，从事影响产品质量工作的人员应是能够胜任的。组织应对建立人员能力、提供所需的培训和保证人员意识的过程形成文件。 组织应： a) 确定从事影响产品质量工作的人员所必要的能力； b) 提供培训或采取其他措施以达到或保持必要的能力； c) 评价所采取措施的有效性； d) 确保员工认识到所从事活动的相关性和重要性，以及如何为实现质量目标作出贡献； e) 保持教育、培训、技能和经验的适当记录（见4.2.5）。 注：对于培训或提供其他措施的有效性的检查方法应与相关的风险相一致。	编制了《培训控制程序》，对承担质量管理体系职责的人员规定相应岗位的能力要求，建立培训工作指导、提高公司员工素质，保证产品质量。 见 2026 年度培训计划表，抽培训执行情况： 1) 15/02/2026，ISO13485:2016 标准培训； 2) 15/06/2026，生产标准培训； 3) 10/05/2026，公司规章制度培训； 以上培训均经过了评价，培训效果良好。 提供宣贯公司方针和目标培训记录，15/06/2026，培训内容：1.宣贯公司方针和目标 2.满足顾客需求是我们长期的质量方针 3.持续质量改进是公司发展的根本动力，培训通过笔试及提问的方式进行考核，培训效果良好。 员工健康管理，提供《从业人员健康证明》，如： 1) 尚悦茂，编号：13310170369202601505，有效期至：14/06/2027，发证机构：雄县中医医院； 2) 董晶晶，编号：13310170369202601422，有效期至：14/06/2027，发证机构：雄县中医医院； 3) 尚骏浩，编号：BDXX0126008667，有效期至：20/05/2027，发证机构：保定市雄县中医医院；	
4.2.4	4.2.4 Control of documents 文件控制 Documents required by the quality management system shall be controlled. Records are a special type of document and shall be controlled according to the requirements given in 4.2.5. A documented procedure shall define the controls needed to: a) review and approve documents for adequacy prior to issue; b) review, update as necessary and re-approve documents; c) ensure that the current revision status of and changes to documents are identified; d) ensure that relevant versions of applicable documents are available at points of use; e) ensure that documents remain legible and readily identifiable; f) ensure that documents of external origin, determined by the organization to be necessary for the planning and operation of the quality management system, are identified and their distribution controlled;	公司已建立《文件控制程序》，通过文件编号、版本号和修订状态对质量管理体系文件进行编制、更新、发放回收的管理和控制。 查《文件发放/回收记录表》，文件名称、编号、版次、生效日期等信息完整，现场使用的均为有效版本。 见《质量手册》，文件编号：XASL-QM，版次：A/0，批准：05/02/2026，批准：尚月永； 见《不合格品控制程序》：文件编号 XASL-QP-013，版次 A/0，编制：编写小组，审核：王倩 批准：尚月永，分发日期：05/02/2026； 查作废文件管控，如有作废文件，由文件管理员收回后，其中一份加盖“作废”标识留存，其余销毁处理，与有效文件隔离存放，防止误用。组织体系运行时间不长，经确认暂无作废文件，暂无文件更改情况。 外来文件控制：查《外来文件清单》，收集了适用的法律法规及标准，如：《中华人民共和国民法典》、《医疗器械注册管理办法》、《医疗器械生产监督管理办法》、《医疗器械经营质量管理规范-GMP》等；	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	g) prevent deterioration or loss of documents; h) prevent the unintended use of obsolete documents and apply suitable identification to them. The organization shall ensure that changes to documents are reviewed and approved either by the original approving function or another designated function that has access to pertinent background information upon which to base its decisions. The organization shall define the period for which at least one copy of obsolete documents shall be retained. This period shall ensure that documents to which medical devices have been manufactured and tested are available for at least the lifetime of the medical device as defined by the organization, but not less than the retention period of any resulting record (see 4.2.5), or as specified by applicable regulatory requirements. 质量管理体系所要求的文件应予以控制。记录是一种特殊类型的文件，应依据 4.2.5 的要求进行控制。 应编制形成文件的程序，以规定以下方面所需的控制： a) 文件发布前得到评审和批准，以确保文件是充分的； b) 必要时对文件进行评审与更新，并再次批准； c) 确保文件的更改和现行修订状态得到识别； d) 确保在使用处可获得适用文件的有关版本； e) 确保文件保持清晰、易于识别； f) 确保组织所确定的策划和运行质量管理体系所需的外来文件得到识别，并控制其分发； g) 防止文件退化或遗失； h) 防止作废文件的非预期使用，并对这些文件进行适当的标识。 组织应确保文件的更改得到原审批部门或指定的其他审批部门的评审和批准，该被指定的审批部门应能获取用于作出决定的相关背景资料。 组织应至少保存一份作废的文件，并确定其保存期限。这个期限应确保至少在组织所规定的医疗器械寿命期内，可以得到这些医疗器械的制造和试验的文件，还应不少于记录（见 4.2.5）或相关法规要求所规定的保存期限。	文件控制符合要求。	
4.2.5	4.2.5 Control of records记录控制 Records shall be maintained to provide evidence of conformity to requirements and of the effective operation of the quality management system. The organization shall document procedures to define the controls needed for the identification, storage, security and integrity, retrieval, retention time and disposition of records. The organization shall define and implement methods for protecting confidential health information contained in records in accordance with the applicable regulatory requirements. Records shall remain legible, readily identifiable and retrievable. Changes to a record shall remain identifiable. The organization shall retain the records for at least the lifetime of the medical device as defined by the organization, or as specified by applicable regulatory requirements, but not less than two years from the medical device release by the organization.	企业建立了《记录控制程序》，规定了记录的标识、贮存、安全与完整性、检索、保存期限及处置要求，确保记录保持清晰、易于识别和检索。 其中规定：生产和质量记录应确保在保存期限内可识别，与产品质量有关的记录保存期限：产品交付后 10 年；与产品设计开发有关的质量记录保存期限不少于最后一批产品生产后 10 年；其他记录为 5 年。 见记录清单如： XASL-OR-007 《采购订单》，XASL-OR-016 《发货通知单》，XASL-OR-012 《合同、订单评审表》等，均有按要求保存留档；保留的记录清晰，易于识别。	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	应保持记录以提供符合要求和质量管理体系有效运行的证据。 组织应建立形成文件的程序，以规定记录的标识、贮存、安全和完整性、检索、保存期限和处置所需的控制，按适用的法规要求，组织应对记录中健康信息规定并实施保护的方法。 记录应保持清晰、易于识别和检索，记录的变更应保持可识别。 组织保存记录的期限应至少为组织所规定的医疗器械的寿命期，而且还应从组织放行医疗器械起不少于 2 年。		

Process Area 过程/区域	生产、检验过程 Production, and inspection processes	负责人 Responsible Person	Wang Qian 王倩/ Dong Yuemao 尚悦茂
-----------------------	---	---------------------------	-------------------------------

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
6.3	6.3 Infrastructure基础设施 The organization shall document the requirements for the infrastructure needed to achieve conformity to product requirements, prevent product mix-up and ensure orderly handling of products. Infrastructure includes, as appropriate: a) buildings, workspace and associated utilities; b) process equipment (both hardware and software); c) supporting services (such as transport, communication, or information systems). The organization shall document requirements for the maintenance activities, including the interval of performing the maintenance activities, when such maintenance activities, or lack thereof, can affect product quality. As appropriate, the requirements shall apply to equipment used in production, the control of the work environment and monitoring and measurement. Records of such maintenance shall be maintained (see 4.2.5). 为达到产品符合要求、防止产品混淆和保证产品的有序处置, 组织应将所需的基础设施的要求形成文件。适当时, 基础设施包括: a) 建筑物、工作场所和相关的设施; b) 过程设备(硬件和软件), c) 支持性服务(如运输、通讯或信息系统)。 当维护活动或缺少维护活动可能影响产品质量, 组织应将此类维护活动的要求包括执行维护活动的时间间隔形成文件。适当时, 要求应适用于生产设备、工作环境控制设备和监视测量设备。 应保持此类维护记录(见 4.2.5)。	企业成立于 2007 年成立, 注册地位于河北省雄安新区雄县张青口村一村 1 组 141-2 号, 企业自有厂房(见村委证明), 总面积约 6000 平方米, 公司为产品实现及质量管理体系运行提供了必要的资源, 具体包括: 基础设施: 办公环境、生产车间、搬运设施、消防设施等。 生产设备: 风批、电钻、调试台、空压机、自动捆包机等。 检测设备: 百分表、数显卡尺、钢卷尺、数显角度尺等, 满足来料、过程及成品检验需求。 生产车间道路畅通、物品堆放整齐、区域明确、照明通风良好。现场配备了灭火器等消防设施, 厂区大院严禁吸烟, 工作环境基本符合规定要求。 提供《设备台账》, 组织主要生产设备有: 流水线、空压机、手电钻、风批等, 满足产品工艺要求。 编制了《电动扳手操作规程》、《手电钻操作和保养规程》等, 明确了设备定期保养和维护的管理要求。 提供《设备维护保养记录》30/06/2026, 设备名称: 手电钻、风批等, 日期: 30/06/2026, 负责人: 王倩; 组织未能提供定期对自动捆包机进行保养维护的有效证据, 违反了《基础设施及设备控制程序》、《自动捆包机操作和保养规程》的有关规定。 储气罐安全附件: 1) 压力表, 编号: 26000230862 号, 型号: 0-1.6Mpa, 检定结论: 合格, 检定日期: 13/07/2026; 2) 安全阀, 编号: YS/BG-26070193429-23, 型号: A27W-16T, 校验结论: 符合要求, 校验日期: 13/07/2026;	m1
6.4.1	6.4 Work environment工作环境 6.4.1 Work environment工作环境 The organization shall document the requirements for the work environment needed to achieve conformity to product requirements. If the conditions for the work environment can have an adverse effect on product quality, the organization shall document the requirements for the work environment and the procedures to monitor and control the work environment. The organization shall: a) document requirements for health, cleanliness and clothing of personnel if contact between such personnel and	企业建立《工作环境管理制度》, 保持生产设施处于清洁、整齐、有序的状态,并持续不断地改善工作环境的条件,以提高员工的工作积极性和工作效率,确保为产品质量创造条件等; 企业生产产品-手动轮椅车不属于无菌医疗器械产品, 对于产品、生产环境及人员等无灭菌及无菌屏障要求, 现场生产环境只需要区域清晰, 无消防及安全隐患即可, 无特殊要求。 基本符合管控要求。	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	the product or work environment could affect medical device safety or performance; b) ensure that all personnel who are required to work temporarily under special environmental conditions within the work environment are competent or supervised by a competent person. NOTE Further information can be found in ISO 14644 and ISO 14698. 组织应将达到符合产品要求所需工作环境的要求形成文件。如果工作环境的条件可能对产品质量有不良影响，组织应将工作环境要求以及监视和控制工作环境的要求形成文件。 组织应： a) 将对特定人员的健康、清洁和着装的要求文件；此类人员与产品或工作环境的接触可能影响医疗器械的安全或性能 b) 确保所有在特殊环境条件下临时工作的人员是胜任的或在胜任的人员监督下工作。 注：更多信息见 ISO14644 和 ISO14698 。		
6.4.2	6.4.2 Contamination control污染控制 As appropriate, the organization shall plan and document arrangements for the control of contaminated or potentially contaminated product in order to prevent contamination of the work environment, personnel, or product. For sterile medical devices, the organization shall document requirements for control of contamination with microorganisms or particulate matter and maintain the required cleanliness during assembly or packaging processes. 适当时，为了防止对工作环境、人员或产品的污染，组织对受污染产品的控制应进行策划并将安排形成文件。 对于无菌医疗器械，组织应将微生物或颗粒物污染的要求形成文件，并在组装或包装过程中保持所要求的洁净度。	企业生产产品-手动轮椅车不属于无菌医疗器械产品，对于产品、生产环境及人员等无灭菌及无菌屏障要求，现场生产环境只需要区域清晰，无消防及安全隐患即可，无特殊关于医疗器械的污染控制要求。 提供《从业人员健康证明》，如： 1) 尚悦茂，编号：13310170369202601505，有效期至：14/06/2027，发证机构：雄县中医院； 2) 董晶晶，编号：13310170369202601422，有效期至：14/06/2027，发证机构：雄县中医院； 3) 尚骏浩，编号：BDXX0126008667，有效期至：20/05/2027，发证机构：保定市雄县中医院；	
7.1	7 PRODUCT REALIZATION产品实现 7.1 Planning of product realization产品实现的策划 The organization shall plan and develop the processes needed for product realization. Planning of product realization shall be consistent with the requirements of the other processes of the quality management system. The organization shall document one or more processes for risk management in product realization. Records of risk management activities shall be maintained (see 4.2.5). In planning product realization, the organization shall determine the following, as appropriate: a) quality objectives and requirements for the product; b) the need to establish processes and documents (see 4.2.4) and to provide resources specific to the product,	产品的生产在受控的条件下进行，企业建立了《生产和服务提供控制程序》等对产品的实现进行策划。 生产部依据产品的实现要求策划明确了产品实现流程。 1) 产品实现过程：原材料采购→进货检验→部件装配→过程检验→成品组装→成品检验→包装→入库 2) 生产过程：制定了工艺文件包括：工艺规程、操作规程、管理规程等。如：手动轮椅车生产工艺流程图；手动轮椅车装配作业指导书、钢卷尺操作规程、手电钻操作和保养规程等； 3) 检验过程：组织按照产品技术要求、内部控制文件等进行检验。如：手动轮椅车原材料进货检验规程；手动轮椅车过程检验规程；手动轮椅车成品检验作业指导书； 4) 策划了相应的生产和检验记录，如：进货检验记录、手动轮椅车生产记录等； 5) 外包过程：无；关键过程：成品组装；特殊过程：无； 以上策划适用于公司产品的生产，通过记录，以及确认产品合格情况，来验证符合性，暂无策划变更。	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	including infrastructure and work environment; c) required verification, validation, monitoring, measurement, inspection and test, handling, storage, distribution and traceability activities specific to the product together with the criteria for product acceptance; d) records needed to provide evidence that the realization processes and resulting product meet requirements (see 4.2.5). The output of this planning shall be documented in a form suitable for the organization's method of operations. NOTE Further information can be found in ISO 14971. 组织应策划和开发产品所需的过程，产品实现的策划应与质量管理体系的其他过程的要求一致。 在产品的实现过程中，将风险管理的一个或多个过程形成文件。应保持风险管理活动的记录（见 4.2.5）。 在策划产品实现的过程中，适当时，组织应确定以下方面的内容： a) 产品的质量目标和要求； b) 针对产品建立过程、文件（见 4.2.4）和提供资源的需求，包括基础设施和工作环境； c) 特定的产品所要求的验证、确认、监视、测量、检验和试验、处理、贮存、销售和追溯活动，以及产品接收准则； d) 为实现过程及其产品满足要求提供证据所需的记录（见 4.2.5）。 策划的输出应以适合于组织的运作方式形成文件。注：更多信息见 ISO 14971。	组织编制了《风险管理控制程序》，判断本公司的产品有关的危害，估计和评估风险，控制这些风险，并监控控制的有效性；提供《手动轮椅车风险管理报告》，针对手动轮椅车开展风险管理，识别各类潜在危害，评估风险等级并落实管控措施，对风险进行了分析和分级，对安全性特征进行了判断等，结论：风险管理计划已被适当的实施，全部剩余风险均处于可接受范围内，且收益超过风险等。 经与管理者代表沟通确认，组织通过上市后信息收集（客户反馈、生产过程数据、售后记录等）持续监控风险，目前暂未发现需要更新风险管理文档的新风险。	
7.3	7.3 Design and development设计和开发 7.3.1 General总则 The organization shall document the design and development procedures. 组织应对设计和开发的程序形成文件。 7.3.2 Design and development planning设计和开发策划 The organization shall plan and control the design and development of products. When appropriate, the design and development plan documents shall be maintained and updated as the design and development progresses. In the process of design and development planning, the organization shall form the following documents: a) Design and development stage; b) Review required in each design and development stage; c) It is applicable to the verification, validation and design conversion activities in each design and development stage; d) Responsibility and authority of design and development; e) Methods to ensure traceability from design and	7.3.1 组织编制了《设计开发控制程序》，对设计和开发的全过程（策划、输入、输出、评审、验证、确认、更改等）进行控制，确保产品满足顾客的需求和期望及有关法律、法规的要求。 7.3.2 项目名称：手动轮椅车开发 开发规格：SYIV 100-SL02A；SYIV 100-SL02B；SYIV 100-SL03A..... 提供《设计和开发策划书》03/01/2025 及《设计开发输入清单》04/01/2025，设计内容： 预期用途：供行动困难的残疾人、病人及年老体弱者作代步工具使用； 主要性能：滑行偏移量≤350mm；驻坡性能≥8°；静态稳定性（侧倾）≥10°等； 对设计开发各阶段进行了划分，如：设计开发策划阶段、设计开发输入阶段、设计开发转换阶段等，各阶段设计开发人员：设计开发策划阶段-董晶晶、王倩；设计开发转换阶段-王倩、尚月茂 7.3.3/7.3.5 见设计开发输入清单和设计开发策划书： 性能指标如：滑行偏移量≤350mm；驻坡性能≥8°；静态稳定性（侧倾）≥10°等； 产品参考的法律法规：《GB/Z 13800:2021 手动轮椅车》、《GB/Z 18029.1:2008 轮椅车 第1部分：静态稳定性的测试》、《GB/Z 18029.3:2008 轮椅车 第3部分：制动性能的测定》等； 预期用途：供行动困难的残疾人、病人及年老体弱者作代步工具使用； 风险管理控制要求如：风险分析；风险评价；风险控制；综合剩余风险评价等；	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	development output to design and development input; f) Resource requirements, including the necessary personnel capabilities. 组织应策划和控制产品的设计和开发。适当时，随着设计和开发的进展，应保持和更新设计和开发计划文件。 设计和开发策划过程中，组织应对以下形成文件： a) 设计和开发阶段； b) 每个设计和开发阶段所需要的评审； c) 适用于每个设计和开发阶段的验证、确认和设计转换活动； d) 设计和开发的职责和权限； e) 为确保设计和开发输出到设计和开发输入可追溯性的方法； f) 包括必要的人员能力在内的所需资源。 7.3.3 Design and development input设计和开发输入 Inputs related to product requirements shall be identified and recorded (see 4.2.5). These inputs shall include: a) According to the intended use, function, performance, availability and safety requirements; b) Applicable regulatory requirements and standards; c) Applicable risk management output; d) Where appropriate, information provided by previous similar designs; e) Other requirements necessary for the design and development of products and processes; The adequacy and suitability of these inputs shall be reviewed and approved. 应确定与产品要求有关的输入并保持记录（见 4.2.5），这些输入应包括： a) 依据预期用途，功能、性能、可用性和安全要求； b) 适用的法规要求和标准； c) 适用的风险管理输出； d) 适当时，以前类似设计提供的信息； e) 产品和过程的设计和开发所必需的其他要求； 应对这些输入的充分性和适宜性进行评审并批准。 7.3.4 Design and development output设计和开发输出 The design and development output shall: a) Meet the requirements of design and development input; b) Provide appropriate information for purchasing, production and service; c) Include or refer to product acceptance criteria; d) Specify the product characteristics necessary for the safety and normal use of the product; The form of design and development output shall be suitable for verification of design and development input and shall be	监视和测量器具：钢卷尺、数显卡尺、百分表等； 输入评审记录：产品性能指标符合国家推荐标准要求；输入的法律法规齐全、适用、有效；风险管理要求符合有关标准要求等，结论：输入文件完整、清楚，无互相矛盾之处，是充分和适宜的，评审人：王倩、尚月茂、董晶晶等，日期：04/01/2025。 7.3.4/7.3.5 见设计开发输出清单： 1) 试制样机； 2) 手动轮椅车产品技术要求； 3) 风险管理报告； 4) 手动轮椅车产品使用说明书； 5) 作业规程如：手动轮椅车原材料进货检验规程；手动轮椅车过程检验规程；手动轮椅车成品检验作业指导书；手动轮椅车生产工艺流程图；手动轮椅车装配作业指导书等 批准：尚月永，日期：01/02/2025； 输出评审：输出文件正确齐全，满足设计开发输入要求；样品性能和功能满足用户需求和产品设计要求等，结论：输出文件完整、清楚，无互相矛盾之处，是充分和适宜的，评审人：王倩、尚月茂、尚月兴等，日期：01/02/2025。 7.3.6 提供《设计开发验证记录》05/02/2025，对产品的可生产性和部件的可获得性进行验证：生产检验人员进行了培训；提供了产品生产所需的生产和检验设备；对供应商进行了评价，选择和确定了合格供方等；抽查《成品组装工序工艺验证报告》，验证日期：12/05/2025，验证结果：按装配工艺流程图装配的样品经检验全部合格，平均用时 30min，结论：装配工艺流程图设计合理，确定在生产实际中采用。验证人员：王倩、尚月茂；另见《产品使用期限试验报告》、《产品运输稳定性试验报告》等文档，均符合规定的要求。 7.3.7 提供《设计和开发确认记录》09/01/2026，临床评价：按照《免于进行临床试验医疗器械目录汇总》的规定，产品免于进行临床试验；确认结论：符合手动轮椅车产品技术要求，送交国家康复辅具研究中心康复辅具质量监督检验中心进行注册检验，编制注册资料上报河北省药品监督管理局，已获准上市许可，确认评审：王倩、尚月茂、董晶晶、尚月兴； 7.3.8 提供《设计和开发转换记录》05/02/2025，组织为组织生产策划了《岗位人员考核评价记录》、《合格供方名录》、《生产记录表》、《成品检验记录》等文件，规定了生产成品的验证要求和生产能力要求。 设计开发输出已成功转换，产品经过了备案登记（编号：冀雄安药监生产备 20260001 号）并形成了订单，产品批量生产后，保留了成品检验放行的证据。 7.3.9 暂无设计开发更改 7.3.10 提供《技术文件》汇编，如：《手动轮椅车生产工艺流程图》、《钢卷尺操作规程》、《自动捆包机操作和保养规程》、《手动轮椅车成品检验规程》等； 提供电动轮椅车设计开发过程记录（设计和开发策划书、设计和开发验证报告、设计和开发确认报告等） 管理基本有效。 当前产品设计开发情况：暂未有新产品的设计开发活动。企业设计开发人员王倩、尚月茂等有多	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	approved before release Records of design and development outputs shall be maintained (see 4.2.5). 设计和开发输出应: a) 满足设计和开发输入的要求; b) 给出采购、生产和服务提供适当的信息; c) 包含或引用产品接收准则; d) 规定对产品的安全和正常使用所必需的产品特性; 设计和开发输出的形式应适合于设计和开发输入的验证, 并应在发布前批准 应保持设计和开发输出的记录 (见 4.2.5)。 7.3.5 Design and development review设计和开发评审 At the appropriate stage, the design and development shall be systematically reviewed according to the planning and documentation arrangement, so as to: a) Evaluate the ability of the design and development results to meet the requirements; b) Identify and propose necessary measures. Participants in the review shall include representatives of functions and other experts related to the design and development phase being reviewed. Records of the review results and any necessary measures shall be maintained (see 4.2.5). 在适宜的阶段, 应依据策划和文件化的安排, 对设计和开发进行系统的评审, 以便: a) 评价设计和开发的结果满足要求的能力; b) 识别和提出必要的措施。 评审的参加者应包括与所评审的设计和开发阶段有关的职能的代表和其他专家。 评审结果及任何必要措施的记录应予保持 (见 4.2.5)。 7.3.6 Design and development verification设计和开发验证 In order to ensure that the design and development output meets the requirements of design and development input, the design and development shall be verified according to the planned and documented arrangement. The organization shall document the validation plan, including methods, acceptance criteria, and, where appropriate, the statistical techniques and principles used to determine the sample size. If the intended use requires the medical device to be connected or joined with other medical devices, the verification shall include the content that the design output meets the design input when connecting or joining according to this. Records of verification results, conclusions and necessary measures shall be kept(See 4.2.4 and 4.2.5) 为确保设计和开发输出满足设计开发输入的要求, 应依据所策	年行业经验, 了解行业有关技术要求。企业配备了百分表、游标卡尺、钢卷尺、AUTOCAD 等设计开发资源, 配备了螺丝刀、手电钻、自动捆包机等生产资源, 具备了设计开发和试生产的必要资源; 针对设计开发活动, 企业编制了《设计和开发控制程序》, 对设计和开发的全过程 (策划、输入、输出、评审、验证、确认、更改等) 进行控制, 确保产品满足顾客的需求和期望及有关法律、法规的要求; 企业编制了《手动轮椅车产品使用说明书》、《手动轮椅车装配作业指导书》等设计开发及工艺文件, 可为后续设计开发活动提供技术参考。	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	划和文件化的安排对设计和开发进行验证。组织应将验证计划形成文件，包括方法、接收准则，适当时，为确定抽样量所采用的统计技术与原理。如果预期用途需要医疗器械与其他医疗器械连接或接合，验证应包含依此连接或接合时，证实设计输出满足设计输入的内容。验证结果和结论以及必要措施的记录应予保持。（见 4.2.4 和 4.2.5）。 7.3.7 Design and development validation设计和开发确认 In order to ensure that the product can meet the specified applicable requirements or the requirements of the intended use, the design and development shall be confirmed according to the planned and documented arrangement. The organization shall document the validation plan, including methods, acceptance criteria, and, where appropriate, the statistical techniques and principles used to determine the sample size. Design confirmation shall be conducted for representative products, including the original production unit, batch or other equivalent. The reasonableness of the product used for validation shall be documented (see 4.2.5). As part of design and development validation, the organization shall conduct clinical or performance evaluations in accordance with applicable regulatory requirements. Medical devices used for clinical evaluation or performance evaluation shall not be considered as released to customers. If the intended use requires the medical device to be connected or joined with other medical devices, the confirmation shall include the content that proves that the specified applicable requirements or the intended use have been met when the medical device is connected or joined. The confirmation shall be completed before the product is delivered to the customer for use. Records of confirmation results and necessary measures shall be kept (see 4.2.4 and 4.2.5) 为确保产品能够满足规定的适用要求或预期用途的要求，应依据所策划并文件化的安排对设计和开发进行确认。 组织应将确认计划形成文件，包括方法、接收准则，适当时，为确定抽样量所采用的统计技术与原理。 应对代表性产品进行设计确认，代表性产品包括最初的生产单位、批或其他等同物。应记录用于进行确认的产品的合理性（见 4.2.5）。 作为设计和开发确认的一部分，组织应按照适用的法规要求进行临床评价或性能评价。 用于临床评价或性能评价的医疗器械不应视作放行给顾客使用。		

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	如果预期用途需要医疗器械与其他医疗器械连接或接合，确认应包含依此连接或接合时，证实规定的适用要求或预期用途已得到满足的内容。 确认应在产品交付给客户使用之前完成。 确认结果及必要措施的记录应予保持（见 4.2.4 和 4.2.5）。 7.3.8 Design and development transformation设计和开发转换 The organization shall document the conversion procedures from design and development output to manufacturing. These procedures shall ensure that the output of design and development is verified in a way suitable for production before becoming the final production specification, and the production capacity can meet the product requirements. The results and conclusions of the conversion should be recorded (see 4.2.5). 组织应将设计和开发输出到制造的转换程序形成文件。这些程序应确保设计和开发的输出在成为最终生产规范之前以适用于生产的方式经过验证，并且生产能力能满足产品要求。 转换的结果和结论应予以记录（见 4.2.5）。 7.3.9 Control of design and development changes设计和开发更改的控制 The organization shall document the procedures for controlling design and development changes. The organization shall identify significant changes related to the function, performance, availability, safety and applicable regulatory requirements of the medical device and its intended use. Design and development changes shall be identified. Before implementation, these changes shall: a) After evaluation; b) It has been verified; c) If appropriate, confirmed; d) Approved. The review of design and development changes shall include the changes of parts and products in the process or that have been distributed and the evaluation of the impact of risk management and changes of input and output in the product realization process. The review results of changes and records of any necessary measures shall be kept (see 4.2.5). The review of design and development changes shall include the evaluation of the impact of changes on product components and work in process or delivered products, and the evaluation of the impact of changes on the input / output of risk management and the process of product realization. 组织应将控制设计和开发变更的程序形成文件。组织应确定与医疗器械的功能、性能、可用性、安全和适用的医疗器械法规		

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	要求和其预期使用有关的重要变更。 设计和开发变更应被识别，实施前， 这些变更应： a) 经过评审； b) 经过验证； c) 适当时，经确认； d) 经过批准。 设计和开发变更的评审应包括过程中或已经配送的部件和产品的变化和风险管理对产品实现过程的输入和输出的变化的影响的评价，更改的评审结果及任何必要措施的记录应予以保持（见 4.2.5）。 设计和开发更改的评审应包括评价更改对产品组成部分和在制品或已交付产品的影响，评价更改对风险管理的输入/输出和产品实现的过程的影响。 7.3.10 Design and development documents 设计和开发文件 The organization shall maintain the design and development documents of each medical device type or medical device family, which shall include or quote the records generated to verify the compliance with the design and development requirements, as well as the records of design and development changes. 组织应保持每一医疗器械类型或医疗器械族的设计和开发文件，此文件应包括或引用为证实符合设计和开发要求所产生的记录，以及设计和开发变更的记录。		
7.4.3	7.4.3 Verification of purchased product 采购产品的验证 The organization shall establish and implement the inspection or other activities necessary for ensuring that purchased product meets specified purchasing requirements. The extent of verification activities shall be based on the supplier evaluation results and proportionate to the risks associated with the purchased product. When the organization becomes aware of any changes to the purchased product, the organization shall determine whether these changes affect the product realization process or the medical device. When the organization or its customer intends to perform verification at the supplier's premises, the organization shall state the intended verification activities and method of product	组织制定了《手动轮椅车原材料进货检验规程》，采购产品的验证见进货检验记录，如： 1) 日期：21/06/2026，品名：后轮，数量：1000 个，检验项目：外观（表面平滑、无开裂；不锈钢辐条表面不应有锈点）；装配质量（轮子应转动均匀，灵活，间隙适当，不得有卡滞或松弛现象等）；尺寸（符合要求）等，检验结论：合格，检验人：尚月茂； 2) 日期：25/06/2026，品名：越野轮胎，数量：1500 个，检验项目：外观（颜色色泽一致）；装配质量（配合良好）；尺寸（符合要求）；性能（有足够的强度和富于弹性）等，检验结论：合格，检验人：尚月茂； 3) 日期：30/06/2026，品名：黑色车架，数量：800 个，检验项目：外观（镀锌表面应平滑、无起皮；焊接表面应光滑平整，焊缝均匀，不得有气孔、裂纹、漏焊等现象）；装配质量（配合良好）；尺寸（符合要求）等，检验结论：合格，检验人：尚月茂； 4) 日期：20/05/2026，品名：座垫，数量：500 个，检验项目：外观（颜色色泽一致）；缝制（缝边整齐）；性能（有足够的强度和富于弹性）等，检验结论：合格，检验人：尚月茂； 查供应商提供的出厂检验报告： 1) 20/05/2026，座垫，检验项目：基本尺寸、缝纫针尺寸、材质、性能等，检验结论：符合，检验单位：雄县雄州镇久奥座垫加工厂； 2) 17/05/2026，车架，检验项目：钢管尺寸、厚度、接焊口外观、材质、性能等，检验结论：符合，检验单位：河北雄安远智医疗科技有限公司；	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	release in the purchasing information. Records of the verification shall be maintained (see 4.2.5). 组织应建立实施检验或其他必要的活动，以确保采购的产品满足规定的采购要求。验证活动的范围应基于供方的评价结果，并与采购产品的风险相适应。当组织觉察到采购产品的任何更改，组织应确定这些更改是否影响产品实现过程或医疗器械最终产品。 当组织或顾客拟在供方的现场实施验证时，组织应在采购信息中对拟验证的活动和产品放行的方法作出规定。 应保留验证记录（见 4.2.5）。		
7.5	7.5 Production and service provision生产和服务实现 7.5.1 Control of production and service provision生产和服务提供的控制 Production and service provision shall be planned, carried out, monitored and controlled to ensure that products conform to specification. As appropriate, production controls shall include but are not limited to: a) documentation of procedures and methods for the control of production (see 4.2.4); b) qualification of infrastructure; c) implementation of monitoring and measurement of process parameters and product characteristics; d) availability and use of monitoring and measuring equipment; e) implementation of defined operations for labelling and packaging; f) implementation of product release, delivery and post-delivery activities. The organization shall plan and carry out production and service provision under controlled conditions, that shall include as applicable also: Availability of information that describes the characteristics of the product. Availability of work instructions. Use of suitable equipment. The organization shall establish and maintain a record (see 4.2.5) for each medical device or batch of medical devices that provides traceability to the extent specified in 7.5.9 and identifies the amount manufactured and the amount approved for distribution. The record shall be verified and approved. If the product is cleaned in accordance with a) or b) above,	公司编制了《生产和服务提供控制程序》，编制了工艺指导书等，保证对过程中影响产品和服务质量的因素处于受控状态，确保产品和服务质量符合规定要求，通过职责、权限和相对应的工作文件和记录表明确了过程的顺序和相互关系，确保产品在生产各个工序处于受控状态，组织结构合理，岗位职责明确： 依据销售订单以及库存情况，合理安排生产，提供《手动轮椅车生产记录》 1) 生产日期：23/06/2026，批号：20260623，型号：SYIV100-SL-01A（手动轮椅车），生产数量：100台； 2) 生产日期：15/05/2026，批号：20260515，型号：SYIV100-SL-01B（手动轮椅车），生产数量：50台； 3) 生产日期：02/07/2026，批号：20260702，型号：SYIV100-SL-01A（手动轮椅车），生产数量：60台； 人员能力：作业人员上岗前均经培训并考核合格。现场与作业人员左岩、徐翠娇等沟通，均了解相应工序的作业要求； 设备设施：组装工具/包装设备状态良好，定期进行盘点/点检，保证生产的顺利开展； 作业文件：作业指导书受控，如《手动轮椅车装配作业指导书》，现场操作有据可依。 工作环境：无特殊环境要求； 过程监视：生产过程所用的计量器具均处于校准有效期内，如百分表、数显卡尺等； 审核生产现场：型号：SYIV100-SL-03B（手动轮椅车），批号：20260716，生产数量：65台； 生产过程中提供了适宜的基础设施，包括厂房和设备/工具，如风批、手电钻、空压机等；监视和测量设备主要为百分表、数显卡尺等； 现场工序控制，作业人员：左岩、徐翠娇、刘兴杰等 1 生产人员组装车架/脚托/前后轮/制动器：工作人员按要求在工作台操作（如安装脚托：将脚踏下节穿入脚踏上节，将脚踏板套在脚踏下节上，将脚踏上节与椅架上的脚踏连接柱连接在一起），对配合较紧密的零部件使用橡胶锤或木槌击打，不得野蛮装配；装配后检查各转动部位转动均匀、灵活，间隙适当；紧固件安装牢固无松动；制动器装配可靠，即时调试，操作简便易行等； 2 轮椅车调试：车轮着地性测试，将轮椅放在测试平台上，平台处于水平位置，按轮椅车最大	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	the requirements contained in 6.4.1 do not apply prior to the cleaning process. 生产和服务提供应进行策划、实施、监视和控制以确保产品符合规范。适当时，生产控制应包括但不限于： a) 编制生产控制程序和控制方法的文件（见 4.2.4）； b) 基础设施鉴定； c) 实施过程参数和产品特性监视和测量； d) 获得和使用监视和测量装置； e) 对标记和包装实施规定的操作； f) 实施产品放行、交付和交付后活动。 组织应为每一台或每一批医疗器械建立并保留记录（见 4.2.5），该记录提供了 7.5.9 中规定的可追溯的范围和程度的信息并标明生产数量和批准销售的数量。应核验和批准该记录。	使用者质量配重后，在轮椅车任何三个轮子下垫一张纸，另一个轮子塞入一个20±1mm的垫块，三个轮子下面的纸抽不出等； 作业人员动作熟练，按作业指导书进行操作，基本符合工艺要求，生产现场管控基本有效。	
7.5.2	7.5.2 Cleanliness of product产品的清洁 The organization shall document requirements for cleanliness of product or contamination control of product if: a) product is cleaned by the organization prior to sterilization or its use; b) The product is supplied non-sterile and is to be subjected to a cleaning process prior to sterilization or its use; c) The product cannot be cleaned prior to sterilization or its use, and its cleanliness is of significance in use; d) The product is supplied to be used non-sterile, and its cleanliness is of significance in use; e) Process agents are to be removed from the product during manufacture. If the product is cleaned in accordance with a) or b) above, the requirements contained in 6.4.1 do not apply prior to the cleaning process. 在下列情况下，组织应将产品清洁或污染控制的要求形成文件： a) 产品在灭菌和或使用由组织进行清洁； b) 产品是以非无菌形式提供且在灭菌或使用先进行清洁处理； c) 产品在灭菌或使用前不能进行清洁处理，使用时其清洁是至关重要的； d) 提供的产品以非无菌使用，使用时其清洁是至关重要的； e) 在制造过程中从产品中除去过程添加物。 如产品是按照上述 a) 或 b) 要求进行清洁，则6.4.1中包含的要求不适用于清洁处理前的过程。	组织产品非灭菌提供，无灭菌及无菌屏障要求。 产品在生产、包装、贮存过程中，主要通过对环境、设备、人员的控制，保持生产车间的整洁； 车间环境：保持基本整洁，无安全隐患； 包装：操作人员经培训上岗，按《产品包装作业指导书》进行操作； 贮存：产品无灭菌要求，一般性清洁环境储存。 制程中产品表面无明显脏污、色泽均匀、表面光滑平整等，通过成品外观检验予以控制，无特殊要求。 提供《手动轮椅车使用说明书》，明确了交付后成品清洁维护要求，如： 1) 使用安全的清洁剂或消毒剂对轮椅车进行常规清洗或消毒，清洗后要用清水将轮椅车表面擦拭干净，以防清洁剂或消毒剂残留对人体皮肤的刺激； 2) 常规清洁用擦拭布巾擦拭，布巾用后清洗干净，在250mg/二氧化氯消毒液中浸泡30min后冲洗消毒液，干燥备用。 3) 当轮椅车不需要使用时，要将表面擦拭干净，折叠后装入原包装箱存放至通风干燥处，要防止雨淋、挤压和摔掷。 	
7.5.3	7.5.3 Installation activities. 安装活动 Where appropriate, the organization shall document the medical device installation requirements and installation verification acceptance guidelines.	手动轮椅车用户简单折叠展开即能使用，不涉及安装活动，删减 7.5.3 安装活动。	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	If an agreed customer requests permission to install a medical device by an external party other than the Organization or its supplier, the Organization shall provide documented requirements for the installation and installation verification of the medical device. Records of medical device installation and installation verification performed by the organization or its suppliers shall be maintained (see 4.2.5) 适当时，组织应将医疗器械安装要求和安装验证接收准则形成文件。 如果经同意的客户要求允许除组织或其供方以外的外部方安装医疗器械，则组织应提供医疗器械安装和安装验证的形成文件的要求。 应保留由组织或其供方完成的医疗器械安装和安装验证的记录（见4.2.5）		
7.5.4	7.5.4 Servicing activities服务活动 If servicing of the medical device is a specified requirement, the organization shall document servicing procedures, reference materials, and reference measurements, as necessary, for performing servicing activities and verifying that product requirements are met. The organization shall analyse records of servicing activities carried out by the organization or its supplier: a) to determine if the information is to be handled as a complaint; b) as appropriate, for input to the improvement process. Records of servicing activities carried out by the organization or its supplier shall be maintained (see 4.2.5). 如果对医疗器械服务有规定的要求，必要时，组织应将服务程序、所涉及的材料和所涉及的测量形成文件。用于实施服务活动并验证产品满足要求。组织应分析由组织或其供方实施的服务活动记录： a) 以确定该信息是否作为投诉进行处置； b) 适当时，作为改进过程形成输入。 应保持组织或其供方所实施服务活动的记录（见4.2.5）。	组织编制了《与顾客有关的过程控制程序》，供销部负责与客户联络，组织处理顾客投诉，负责保存相关的服务记录，负责分析顾客反馈信息，分析结果报管理者代表，管理者代表对供销部的分析结果进行评审，确定责任部门，制定整改计划、纠正预防措施（必要时）并监督实施。 组织定期对客户满意度进行调查，见客户满意度调查表； 组织暂时未收到相关的投诉信息。若有，依照《质量信息报警反馈控制程序》、《顾客抱怨处理程序》处理。	
7.5.5	7.5.5 Particular requirements for sterile medical devices. (1) 无菌医疗器械的专用要求 The organization shall maintain records of the sterilization process parameters used for each sterilization batch (see 4.2.5). Sterilization records shall be traceable to each production batch of medical devices. 组织应保持每个灭菌批所用灭菌工艺参数的记录（见4.2.5）。灭菌记录应当溯源到每一批生产的医疗器械。	组织的产品不是无菌医疗器械产品，均不执行无菌医疗器械的专用要求，删减 7.5.5	
7.5.6	7.5.6 Validation of processes for production and service	企业制定了各工序的装配作业指导书和检验操作规程，岗前对员工进行培训。通过员工感官+测	m2

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	provision生产和服务过程的确认 The organization shall validate any processes for production and service provision where the resulting output cannot be or is not verified by subsequent monitoring or measurement and, as a consequence, deficiencies become apparent only after the product is in use or the service has been delivered. Validation shall demonstrate the ability of these processes to achieve planned results consistently. The organization shall document procedures for validation of processes, including: a) defined criteria for review and approval of the processes; b) equipment qualification and qualification of personnel; c) use of specific methods, procedures and acceptance criteria; d) as appropriate, statistical techniques with rationale for sample sizes; e) requirements for records (see 4.2.5); f) re-validation, including criteria for re-validation; g) approval of changes to the processes. The organization shall document procedures for the validation of the application of computer software used in production and service provision. Such software applications shall be validated prior to initial use and, as appropriate, after changes to such software or its application. The specific approach and activities associated with software validation and re-validation shall be proportionate to the risk associated with the use of the software, including the effect on the ability of the product to conform to specifications. Records of the results and conclusion of validation and necessary actions from the validation shall be maintained (see 4.2.4 and 4.2.5). 当生产和服务提供过程的输出不能或不是由后续的监视或测量加以验证，并因此使问题产品在产品使用后或服务已交付后才会显现时，组织应对这样的过程进行确认。确认应证实这些过程具有持续实现所策划的结果的能力。组织应将过程的确认程序形成文件，包括： a) 为过程的评审和批准所规定的准则； b) 设备的鉴定和人员资格鉴定； c) 使用特定的方法、程序和接收准则； d) 适当时，包含样本量原理与统计技术 e) 记录的要求（见 4.2.5）； f) 再确认，包括再确认的准则； g) 对过程变更的批准。 组织应将用于生产和服务提供的计算软件确认程序形成文件。此类软件的确认应在初次使用前确认。适当时，在此类软件应用或其应用更改后也应予以确认。有关软件确认和再确认的特	量自检、上下工序互检的方式确保工序质量合格，防止不良产品流出。工序不良品可返工处理后重新进入产线，企业当前无特殊工序。 企业用于生产和服务的计算机软件：如畅捷通 T+ V13.0 版；微打标签打印软件 2.0.2019.920 版等； 组织未提供对用于生产和服务的计算机软件：如畅捷通 T+ V13.0 版；微打标签打印软件 2.0.2019.920 版定期进行软件确认的有效证据，违反了《计算机软件应用确认控制程序》的有关规定。	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	定方法和活动应与软件使用有关的风险相适应，包括对产品符合规范能力的影响。 应保留确认的结果和结论以及确认所采取的必要措施的记录（见 4.2.4 和 4.2.5）。		
7.5.7	7.5.7 Particular requirements for validation of processes for sterilization and sterile barrier systems.灭菌过程和无菌屏障系统确认的专用要求 The organization shall document the validation procedures for sterilization and aseptic barrier system processes (see 4.2.4). Sterilization and aseptic barrier systems should be confirmed prior to implementation and, where appropriate, prior to implementation of subsequent product or process changes. Records of the findings and conclusions of the confirmation and the necessary measures taken should be maintained (see 4.2.4 and 4.2.5). 组织应对灭菌和无菌屏障系统过程的确认程序形成文件（见 4.2.4）。 灭菌和无菌屏障系统应在实施前得到确认，适当时，还应在后续的产品或过程更改实施前得到确认。 应保留确认的结果和结论以及确认所采取的必要措施的记录（见4.2.4和4.2.5）	组织的产品不是无菌医疗器械产品，均不执行无菌医疗器械的专用要求，删减 7.5.7	
7.5.8	7.5.8 Identification标识 The organization shall document procedures for product identification and identify products by suitable means throughout product realization. The organization shall identify product status with respect to monitoring and measurement requirements throughout product realization. Identification of product status shall be maintained throughout production, storage, installation and servicing of product to ensure that only product that has passed the required inspections and tests or released under an authorized concession is dispatched, used or installed. If required by applicable regulatory requirements, the organization shall document a system to assign unique device identification to the medical device. 组织应将产品标识程序形成文件，并在产品实现的整个过程中使用适当的方法识别产品。 组织应在产品实现的整个过程中按照监视测量的要求识别产品状态。在产品的生产、贮存、安装和服务的全过程中应保持产品的状态标识，以确保只有通过所要求的检验和试验或经授权让步放行的产品才能被发送、使用或安装。 如果有适用的法规要求，组织应将医疗器械指定唯一器械标识的系统形成文件。	编制了《产品标识和可追溯性控制程序》，对产品标识、状态标识等管理控制进行了规定。 查看现场：区域划分明显，包括手动轮椅车生产线、原材料仓库、成品仓库、办公区域等； 检查最小销售单元的标签，明确了规格型号、生产备案凭证编号、产品备案凭证编号、产品技术要求编号、最大使用者质量、行驶限制、使用期限、生产地址、住所、生产企业、联系电话等信息； 检查轮椅成品外包装箱，明确了生产备案凭证编号、产品备案凭证编号、产品技术要求编号、生产单位、生产地址、服务电话等信息，以上基本符合医疗器械有关法规要求。	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	组织应建立程序并形成文件以确保返回组织的医疗器械能被识别且能与合格的产品区分开。		
7.5.9.1	7.5.9 Traceability可追溯性 7.5.9.1 General总则 The organization shall document procedures for traceability. These procedures shall define the extent of traceability in accordance with applicable regulatory requirements and the records to be maintained (see 4.2.5). 组织应将可追溯性程序形成文件，这些程序应依据适用的法规要求规定的可追溯性的范围和程度以及拟保留的记录（见4.2.5）。	企业编制了《产品标识和可追溯性控制程序》，通过装车单、出库单、订货合同、生产检验记录等进行生产交付过程的全流程追溯。 如抽查：追溯产品出库交付过程：手动轮椅车，规格：24 寸，型号：SYIV100-SL02A，数量：918 台，顾客：天津高米迪自行车有限公司，订单日期：07/06/2026，交付方式：派集装箱到乙方提货，以防负责装货；出库单：出库日期：01/07/2026，米总柜一：459 台；米总柜二：459 台；装车单：箱号：TRHU6407712，车号：津 C56230，司机电话：15222779803 等； 查《手动轮椅车手动轮椅车生产记录》15/05/2026，未填写可追溯各主要原材料的批次信息，违反了《医疗器械生产质量管理规范》、《产品标识和可追溯性控制程序》的有关规定。	m3
7.5.9.2	7.5.9.2 Particular requirements for implantable medical devices植入性医疗器械的专用要求 The records required for traceability shall include records of components, materials, and conditions for the work environment used, if these could cause the medical device not to satisfy its specified safety and performance requirements. The organization shall require that suppliers of distribution services or distributors maintain records of the distribution of medical devices to allow traceability and that these records are available for inspection. Records of the name and address of the shipping package consignee shall be maintained (see 4.2.5). 如果所使用的组件、材料和工作环境条件等因素可能导致医疗器械不满足其规定的安全和性能要求，可追溯性所要求的记录应包括这些相关因素的记录。 组织应要求提供流通服务的供方或经销商保留医疗器械流通记录以便追溯，当检查需要，可获得这些记录。 应保留货运包装收件人的名字和地址的记录（见4.2.5）	组织为非植入性医疗器械，故植入性医疗器械专用要求不适用。	
7.5.10	7.5.10 Customer property顾客财产 The organization shall identify, verify, protect, and safeguard customer property provided for use or incorporation into the product while it is under the organization's control or being used by the organization. If any customer property is lost, damaged or otherwise found to be unsuitable for use, the organization shall report this to the customer and maintain records (see 4.2.5). 若顾客财产在组织的控制下或由组织使用，对组织使用的或构成产品一部分的顾客财产，组织应予以识别、验证、保护和防护。若任何顾客财产发生丢失、损坏或发现不适用情况时，组织应向顾客报告并保留记录（见4.2.5）	建立了《顾客财产控制程序》，规定了顾客财产的识别、验证、保护和防护要求。 顾客财产包括：客户提供的标签、外文说明书、外箱贴等；如日期：30/06/2026，收到顾客：杭州瑞憬照明科技有限公司提供的终端客户的产品贴、外箱贴的设计稿件，顾客财产唯一接收和保管人：王倩，目前未有顾客财产的损失。	
7.5.11	7.5.11 Preservation of product产品防护 The organization shall document procedures for preserving the conformity of products to requirements during processing.	企业建立了《产品防护管理程序》，成品直接折叠后放入包装箱内。手动轮椅车应存放在干燥、通风良好，环境温度在-40℃~+40℃，相对湿度不超过 80%。室内应避免强烈日光。周围无腐蚀	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	storage, handling, and distribution in order to maintain conformity to requirements. Preservation shall apply to the constituent parts of a product or medical device. The organization shall protect the product from alteration, contamination or damage when exposed to expected conditions and hazards during processing, storage, handling, and distribution by: a) designing and constructing suitable packaging and shipping containers; b) documenting requirements for special conditions needed if packaging alone cannot provide preservation. If special conditions are required, they shall be controlled and recorded (see 4.2.5). 在产品的加工、贮存、处置和流通期间，组织应将产品符合要求提供防护的程序形成文件。防护应适用于医疗器械的组成部分。 在产品的加工、贮存、处置和流通期间，当其暴露在预期条件和危险（源）时，组织应通过以下方式来保护产品发生变化、污染或损坏： a) 设计和制作适当的包装和货运包装箱； b) 如果包装本身不能提供防护，将所需的特殊条件要求形成文件。 如要求特殊条件，其应受控并予以记录（见4.2.5）。	性的物质。 包装好的手动轮椅车可以用一般运输工具运输，运输中应注意包装箱上的“向上”、“怕雨”、“禁止翻滚标识”，防止由于不当搬运而造成的破坏；当轮椅车不需要使用时，要将表面擦拭干净，折叠后装入原包装箱存放至通风干燥处，要防止雨淋、挤压和摔掷等。 无其他针对物料、成品等的其他特殊防护要求。	
7.6	7.6 Control of monitoring and measuring equipment监视和测量设备的控制 The organization shall determine the monitoring and measurement to be undertaken and the monitoring and measuring equipment needed to provide evidence of conformity of product to determined requirements. The organization shall document procedures to ensure that monitoring and measurement can be carried out and are carried out in a manner that is consistent with the monitoring and measurement requirements. As necessary to ensure valid results, measuring equipment shall: a) be calibrated or verified, or both, at specified intervals, or prior to use, against measurement standards traceable to international or national measurement standards: when no such standards exist, the basis used for calibration or verification shall be recorded (see 4.2.5); b) be adjusted or re-adjusted as necessary: such adjustments or re-adjustments shall be recorded (see 4.2.5); c) have identification in order to determine its calibration status; d) be safeguarded from adjustments that would invalidate the measurement result;	企业建立了《监视和测量装置控制程序》，监测和测量活动有助于制定监测过程、产品和服务的指标。抽样计量器具校准执行情况，如： 1) 数显角度尺，型号：0-180°，编号：HC2512240022，校准结论：合格，校准日期：24/12/2025； 2) 百分表，型号：0-10mm，编号：HC2512240019，校准结论：合格，校准日期：24/12/2025； 3) 数显卡尺，型号：0-150mm，编号：HC2512240020，校准结论：合格，校准日期：24/12/2025； 4) 钢卷尺，型号：3m，编号：HC2512240021，校准结论：合格，校准日期：24/12/2025； 企业无用于监视和测量要求的计算机软件。 管理基本有效。	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	e) be protected from damage and deterioration during handling, maintenance and storage. The organization shall perform calibration or verification in accordance with documented procedures. In addition, the organization shall assess and record the validity of the previous measuring results when the equipment is found not to conform to requirements. The organization shall take appropriate action in regard to the equipment and any product affected. Records of the results of calibration and verification shall be maintained (see 4.2.5). The organization shall document procedures for the validation of the application of computer software used for the monitoring and measurement of requirements. Such software applications shall be validated prior to initial use and, as appropriate, after changes to such software or its application. The specific approach and activities associated with software validation and re-validation shall be proportionate to the risk associated with the use of the software, including the effect on the ability of the product to conform to specifications. Records of the results and conclusions of validation and necessary actions from the validation shall be maintained (see 4.2.4 and 4.2.5). NOTE Further information can be found in ISO 10012. 组织应确定需实施的监视和测量以及所需的监视和测量装置，为产品符合确定的要求提供证据。 组织应建立程序并形成文件，以确保监视和测量活动可行并符合与监视和测量的要求相一致的方式实施。 为确保结果有效，必要时，测量设备应： a) 对照能溯源到国际或国家标准的测量标准，按照规定的时间间隔或使用前进行校准和（或）检定。当不存在上述标准时，应记录校准或检定的依据（见4.2.5）； b) 必要时得到调整或再调整；应记录这种调整或再调整（见4.2.5）； c) 具有标识，以确定其校准状态； d) 予以防护，防止由于调整使测量结果失效； e) 予以保护，防止处置、搬运和贮存期间损坏和衰减。 组织应按照所形成文件的程序进行校准或检定。 此外，当发现设备不符合要求时，组织应对以往测量结果的有效性进行评价和记录。组织应对该设备和任何受影响的产品采取适当的措施。 应保留校准和检定（验证）结果的记录（见4.2.5）。 组织应对用于监视测量要求的计算机软件应用的确认程序形成文件。此类软件应在初次使用前进行确认，适当时，此类软件		

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	或其应用更改后也应予确认。 有关软件确认和再确认的特定方法和活动应与软件使用相关的风险包括对产品符合规范能力的影响）相适应。 应保留确认的结果和结论以及确认所采取的必要措施的记录（见4.2.4 和4.2.5）。 注：进一步信息见 ISO 100127.		
8.2.5	8.2.5 Monitoring and measurement of processes过程的监视和测量 The organization shall apply suitable methods for monitoring and, as appropriate, measurement of the quality management system processes. These methods shall demonstrate the ability of the processes to achieve planned results. When planned results are not achieved, correction and corrective action shall be taken, as appropriate. 组织应采用适宜的方法对质量管理体系过程进行监视，并在适当时进行测量。这些方法应证实过程实现所策划结果的能力。若未能实现策划结果，适当时，应采取纠正和纠正措施。	组织采用适宜的方法对质量管理体系中的各项过程进行动态跟踪和评估，以验证其是否具备稳定输出预期结果的能力。当监视结果表明过程未达成策划结果时，组织按《纠正预防措施控制程序》采取纠正措施。 经查阅，目前各过程及均达成策划结果。	
8.2.6	8.2.6 Monitoring and measurement of product产品的监视和测量 The organization shall monitor and measure the characteristics of the product to verify that product requirements have been met. This shall be carried out at applicable stages of the product realization process in accordance with the planned and documented arrangements and documented procedures. Evidence of conformity to the acceptance criteria shall be maintained. The identity of the person authorizing release of product shall be recorded (see 4.2.5). As appropriate, records shall identify the test equipment used to perform measurement activities. Product release and service delivery shall not proceed until the planned and documented arrangements have been satisfactorily completed. For implantable medical devices, the organization shall record the identity of personnel performing any inspection or testing. 组织应对产品的特性进行监视和测量，以验证产品要求已得到满足。这种监视和测量应依据策划并形成文件的安排和形成文件的程序，在产品实现过程的适当阶段进行。 应保持符合接收准则的证据。应记录有权放行产品的人员身份（见4.2.5）。 适当时，记录应识别用于执行测量活动的检测设备。 在策划并形成文件的安排已圆满完成前，不应放行产品和交付服务。 对于植入性医疗器械，组织应记录进行任何检验和试验人员的	企业建立了《过程和产品监视测量程序》、《手动轮椅车原材料进货检验规程》、《手动轮椅车过程检验规程》、《手动轮椅车成品检验作业指导书》等，对产品的特性进行监视和测量，以验证产品满足要求。 组织从事手动轮椅车的生产和销售，不涉及植入性医疗器械管控。 过程检验，《手动轮椅车过程检验规程》中明确了对装配工序的检验要求，如：车架检验要求（零部件的装配应准确、可靠，各紧固件的车架装配装配应牢固，不得有松动现象。可转动部位应转动灵活，不得有卡滞现象等）；脚托架、脚托检验要求（装配牢固可靠，不得有松动现象。脚托应易于翻动，翻起后应不会自行落下）；前轮、后轮检验要求（装配牢固可靠，转动灵活，后轮的辐条应调节均匀，辐条头不允许突出条母之外）等。提供《过程检验记录》，如： 1) 日期：08/05/2026，型号：SYIV100-SL01A，检查工序：车架装配质量；脚托装质量等，均合格，操作者：徐翠姣，检验员：尚月茂； 2) 日期：15/06/2026，型号：SYIV100-SL03AB，检查工序：装前后轮质量；装驻车制动器质量等，均合格，操作者：左岩、刘兴杰，检验员：尚月茂； 按策划要求，对产品进行了成品检验，提供《成品检验记录》，如： 1) 检验日期：23/06/2026，型号：SYIV100-SL01A，数量：100，检验项目：表面质量（手能触及的零部件平整、光滑，无毛刺尖角等；镀锌表面应平滑，无滴瘤，无起皮、无残留溶剂渣；焊接表面不得有焊瘤或凹坑，焊缝均匀，不得有气孔、裂纹、漏焊、烧穿等缺陷等）；外形尺寸（符合要求）；装配后所有转动、移动部件应运动均匀、灵活，间隙适当（符合要求）；装配后轮毂径跳 2mm；手圈端跳 1mm；脚托翻下后与脚托架的夹角为 89°；驻坡性能≥8°（符合要求）；静态稳定性（纵向前倾/纵向后倾/侧倾）≥10°（符合要求）等，检验结果：合格，检验员：尚月茂； 2) 检验日期：15/05/2026，型号：SYIV100-SL01B，数量：50，检验项目：表面质量（手能触及的零部件平整、光滑，无毛刺尖角等；镀锌表面应平滑，无滴瘤，无起皮、无残留溶剂渣；焊接表面不得有焊瘤或凹坑，焊缝均匀，不得有气孔、裂纹、漏焊、烧穿等缺陷等）；外形尺寸（符合要求）；装配后所有转动、移动部件应运动均匀、灵活，间隙适当（符合要求）；装配后轮毂径跳 2mm；手圈端跳 1mm；脚托翻下后与脚托架的	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	身份。	夹角为 89°；驻坡性能≥8°（符合要求）；静态稳定性（纵向前倾/纵向后倾/侧倾）≥10°（符合要求）等，检验结果：合格，检验员：尚月茂； 3) 检验日期：09/07/2026，型号：SYIV100-SL03B，数量：80，检验项目：表面质量（手能触及的零部件平整、光滑，无毛刺尖角等；镀锌表面应平滑，无滴瘤，无起皮、无残留溶剂渣；焊接表面不得有焊瘤或凹坑，焊缝均匀，不得有气孔、裂纹、漏焊、烧穿等缺陷等）；外形尺寸（符合要求）；装配后所有转动、移动部件应运动均匀、灵活，间隙适当（符合要求）；装配后轮毂径跳 3mm；手圈端跳 2mm；脚托翻下后与脚托架的夹角为 87°；驻坡性能≥8°（符合要求）；静态稳定性（纵向前倾/纵向后倾/侧倾）≥10°（符合要求）等，检验结果：合格，检验员：尚月茂； 所有产品经过成品检验合格后入库。 第三方检验，如： 1) 手动轮椅车，型号：SYIV100-SL01A，报告编号：155/2025，检验依据：《手动轮椅车》医疗器械产品技术要求，检验项目：表面质量、外形尺寸、质量、装配质量、性能、体位固定装置阻燃性、静态强度、冲击强度、疲劳强度等，检验结论：样品符合要求，签发日期：09/01/2026，检验单位：国家康复辅具研究中心康复辅具质量监督检验中心； 2) 手动轮椅车，型号：SYIV100-SL06B，报告编号：153/2025，检验依据：《手动轮椅车》医疗器械产品技术要求，检验项目：表面质量、外形尺寸、质量、装配质量、性能、体位固定装置阻燃性、静态强度、冲击强度、疲劳强度等，检验结论：样品符合要求，签发日期：09/01/2026，检验单位：国家康复辅具研究中心康复辅具质量监督检验中心； 产品无年度第三方检验要求。	
8.3.1	8.3 Control of nonconforming product不合格品控制 8.3.1 General总则 The organization shall ensure that products which do not conform to product requirements are identified and controlled to prevent its unintended use or delivery. The organization shall document a procedure to define the controls and related responsibilities and authorities for the identification, documentation, segregation, evaluation and disposition of nonconforming product. The evaluation of nonconformity shall include a determination of the need for an investigation and notification of any external party responsible for the nonconformity. Records of the nature of the non-conformities and any subsequent action taken, including the evaluation, any investigation and the rationale for decisions shall be maintained (see 4.2.5). 组织应确保对不符合产品要求的产品进行识别和控制，以防止其非预期的使用或交付。组织应建立程序并形成文件以规定不合格品控制以及规定不合格品识别、记录、隔离、评价和处置有关职责和权限，不合格的评价应包括确定是否需要调查和通知对不合格负责的所有外部方。应保留不合格的性质以及随后采取任何措施的记录，包括评价、任何调查和决策的理由说明（见4.2.5）。	企业编制了《不合格品控制程序》对不合格品的控制，程序文件规定了不合格品的识别、记录、隔离、评价和处置有关对职责和权限，不合格的评价包括确定是否需要调查和通知对不合格负责的所有外部方，如来料不合格时，应将不合格信息反馈给供应商采取有效措施；如发现已出货的成品中出现不合格时，根据严重程度和承担风险大小，经讨论后决定采取如发布忠告性通知或直接通知客户，将产品召回等措施。	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
8.3.2	8.3.2 Actions in response to nonconforming product detected before delivery交付前不符合产品的响应措施 The organization shall deal with nonconforming product by one or more of the following ways: a) taking action to eliminate the detected nonconformity; b) taking action to preclude its original intended use or application; c) authorizing its use, release or acceptance under concession by a relevant authority and, where applicable, by the customer. The organization shall ensure that nonconforming product is accepted by concession only if the justification is provided, approval is obtained and applicable regulatory requirements are met. Records of the acceptance by concession and the identity of the person authorizing the concession shall be maintained (see 4.2.5). 组织应通过下列一种或几种途径处置不合格品: a) 采取措施消除已发现的不合格; b) 采取措施防止其原预期使用或应用 , c) 授权让步使用、放行或接受。 组织应确保不合格品仅在提供理由、获得批准并且满足适用的法规要求情况下才能实施让步接收。 应保留让步接受和授权让步人员身份的记录 (见4.2.5) 。	企业编制了《不合格品控制程序》，对交付前产品，如：来料、生产过程品、成品检验过程等发现的不合格品的控制。 提供《进货不合格检验记录》30/06/2026，品名：黑色车架，数量：100个，外观表面2件不良，处理意见：1件退回厂家重新褪色氧化；1件返修重新补焊，结论：两件不良品退回厂家，其余98件入库，检验员：尚月茂，复核：王倩；	
8.3.3	8.3.3 Actions in response to nonconforming product detected after delivery交付后不符合产品的响应措施 When nonconforming product is detected after delivery or use has started, the organization shall take action appropriate to the effects, or potential effects, of the nonconformity. Records of actions taken shall be maintained (see 4.2.5). The organization shall document procedures for issuing advisory notices in accordance with applicable regulatory requirements. These procedures shall be capable of being put into effect at any time. Records of actions relating to the issuance of advisory notices shall be maintained (see 4.2.5). 当交付后或开始使用后发现不合格品时，组织应采取与不合格的影响或潜在影响程度相适应的措施。应保留采取措施的记录 (见 4.2.5) 。 组织应按照适用的法规要求将忠告性通知的发布程序形成文件。这些程序应能随时付诸实施。应保留与发布忠告性通知有关的措施记录。(见 4.2.5) 。	企业编制了《不合格品控制程序》，《忠告性通知和产品召回控制程序》，《顾客抱怨处理程序》，对交付后后发现的不合格品的控制。 自体系运行以来，体系覆盖内医疗器械产品未出现交付后的产品不合格情况。	
8.3.4	8.3.4 Rework返工 The organization shall perform rework in accordance with documented procedures that take into account the potential adverse effect of the rework on the product. These procedures shall undergo the same review and approval as	企业编制《不合格品控制程序》，对返工做出了规定，对需要返工的产品，有关部门制定返工作业指导书，并通过与原作业指导书相同的审批程序。对产品返工过程进行管理。 自体系运行以来，体系覆盖内医疗器械产品暂未进行过批次返工作业。	

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	the original procedure. After the completion of rework, product shall be verified to ensure that it meets applicable acceptance criteria and regulatory requirements. Records of rework shall be maintained (see 4.2.5). 组织应按照考虑了返工对产品潜在的不良影响所形成文件的程序进行返工，这些程序应经过与原程序相同的评审和批准。 返工完成后，产品应经验证以确保其满足适用的接收准则和法规要求。 应保留返工的记录（见 4.2.5）。		

Process Area 过程 / 区域	采购过程、销售过程 Procurement process, sales process	负责人 Responsible Person	Shang Yuexing 尚月兴
-------------------------	---	---------------------------	-------------------

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
7.2.1	7.2 Customer-related processes与顾客有关的过程 7.2.1 Determination of requirements related to the product与产品有关要求的确定 The organization shall determine: a) requirements specified by the customer, including the requirements for delivery and post-delivery activities; b) requirements not stated by the customer but necessary for specified or intended use, as known; c) applicable statutory and regulatory requirements related to the product; d) any user training needed to ensure specified performance and safe use of the medical device; e) any additional requirements determined by the organization. 组织应确定: a) 顾客规定的要求, 包括对交付及交付后活动的要求 ; b) 顾客虽然没有明示 , 但规定的或已知的预期用途所必需的要求; c) 与产品有关的适用的法规要求; d) 确保医疗器械的特定性能和安全使用所需的用户培训 ; e) 组织确定的任何附加要求。	企业建立了《与顾客有关的过程控制程序》。识别顾客的需求和期望, 在满足医疗器械要求的法规前提下, 确保顾客要求得到有效满足; 组织通过合同、订单及技术协议, 识别并确定与产品有关的要求, 涵盖以下方面: 1) 产品要求规定, 如明示的要求: 车架漆面光洁无划痕, 车轮转动不能偏摆, 前后轮转动顺畅等; 2) 法律法规要求: 如《中华人民共和国民法典》; 3) 公司满足合同要求如: 型号、交货期、价格; 4) 其他: 违约责任、付款方式、纠纷解决方式等; 公司与所有客户签订合同, 并在合同签订前进行评审。确保客户要求已明确并形成文件、组织具备满足要求的能力等, 提供《合同评审表》, 日期: 05/06/2026, 评审项目: 天津高米迪自行车有限公司的手动轮椅车交付项目, 评审内容包括: 规格要求、交货期、技术指标、价格与付款方式等, 评审通过, 评审人: 董晶晶、尚月兴等, 批准: 尚月永; 抽样签订的手动轮椅车销售合同: 1) 07/06/2026, 顾客: 天津高米迪自行车有限公司, 品名: 手动轮椅车, 型号: SY IV100-SL02A, 规格: 24寸, 数量: 918台; 2) 12/05/2026, 顾客: 艾梅儿国际贸易(广州)有限公司, 品名: 手动轮椅车, 型号: SYIV100-SL01B, 数量: 50台; 型号: SYIV100-SL01B, 数量: 50台; 3) 12/03/2026, 顾客: 衡水优尼德贸易有限公司, 品名: 手动轮椅车, 型号: SYIV 100-SL01A, 规格: 16寸, 数量: 3528台; 合同变更后应重新经过评审, 并确保相关的文件得到修改、文件的使用场所能够得到变更的文件。企业未发生合同变更的情况。	
7.2.2	7.2.2 Review of requirements related to the product产品要求的评审 The organization shall review the requirements related to product. This review shall be conducted prior to the organization's commitment to supply product to the customer (e.g. submission of tenders, acceptance of contracts or orders, acceptance of changes to contracts or orders) and shall ensure that: a) product requirements are defined and documented; b) contract or order requirements differing from those previously expressed are resolved; c) applicable regulatory requirements are met; d) any user training identified in accordance with 7.2.1 is available or planned to be available; e) the organization has the ability to meet the defined requirements. Records of the results of the review and actions arising from the review shall be maintained (see 4.2.5). When the customer provides no documented statement of requirement, the customer requirements shall be confirmed by the organization before acceptance. When product requirements are changed, the organization shall		

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	ensure that relevant documents are amended and that relevant personnel are made aware of the changed requirements. 组织应评审与产品有关的要求。评审应在组织向顾客做出提供产品的承诺（如：提交标书、接受合同或订单及接受合同或订单的更改）之前进行并应确保： a) 产品要求得到规定并形成文件； b) 与以前表述不一致的合同或订单的要求已得到解决； c) 满足适用的法规要求； d) 依据 7.2.1 识别的任何用户培训是可获得的或按计划可获得的； e) 组织有能力满足规定的要求。 应保留评审结果及评审所形成的措施的记录（见 4.2.5）。 若顾客没有提供形成文件的要求，组织在接受顾客要求前应对顾客要求进行确认。 若产品要求发生变更，组织应确保相关文件得到修改，并确保相关人员知道已变更的要求。		
7.2.3	7.2.3 Customer communication顾客沟通 The organization shall plan and document arrangements for communicating with customers in relation to: a) product information; b) enquiries, contracts or order handling, including amendments; c) customer feedback, including complaints; d) advisory notices. The organization shall communicate with regulatory authorities in accordance with applicable regulatory requirements. 组织应就以下方面与顾客的沟通进行策划并将安排形成文件： a) 产品信息； b) 处理问询、合同或订单，包括更改； c) 顾客反馈，包括投诉； d) 忠告性通知。 组织应依据适用的法规要求与监管机构进行沟通。	产品售前和售中，可通过电话、传真、电子邮件、广告宣传等渠道向顾客介绍产品，回答顾客咨询。 形成文件的信息如下： 1) 与客户签订的合同或订单，合同评审单； 2) 销售记录； 3) 客户对产品信息的反馈，含医疗器械成品在使用过程中涉及到本公司产品问题的信息反馈，含客户投诉； 4) 忠告性通知，产品交付后，客户依据产品使用说明使用。 组织制定了《医疗器械不良事件报告和再评价程序》、《上市后的监督程序》、《忠告性通知和产品召回控制程序》等，当需要发布忠告性通知时，通过正式信函、官网公告等方式告知客户。经了解，本年度未发布忠告性通知。	
7.4.1	7.4 Purchasing采购 7.4.1 Purchasing process采购过程 The organization shall document procedures (see 4.2.4) to ensure that purchased products conform to specified purchasing information. The organization shall establish criteria for the evaluation and selection of suppliers. The criteria shall be: a) based on the supplier's ability to provide products that meet the organization's requirements; b) based on the performance of the supplier; c) based on the effect of the purchased product on the quality of the medical device; d) proportionate to the risk associated with the medical device. The organization shall plan the monitoring and re-evaluation of	企业建立了《采购控制程序》、《供方评估和复评程序》，对采购过程实施控制，保证所采购的物资符合质量要求，确保公司生产产品符合产品要求与顾客的需求。 抽见合格供应商名录，明确了供方名称，评定表编号，采购产品名称、采购产品型号等信息，批准：尚月兴，日期：10/02/2026，如： 1) 河北雄安远智医疗科技有限公司，供应：车架； 2) 雄县雄州镇久奥座垫加工厂，供应：座垫、靠背、安全带； 3) 河北雄安耀鑫商贸有限公司，供应：轮胎、车轮； 4) 雄县靖凯商贸有限公司，供应：踏板； 对供应商收集了有关资质，填写了《合格供方评定记录》，包括评定材料品质，交货期、价格及付款方式、投诉处理服务等，考核合格列入合格供应商名录，评审小组：董晶晶、王倩、尚月兴、尚悦茂，审批：尚月永，日期：10/02/2026；	m4

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	suppliers. Supplier performance in meeting requirements for the purchased product shall be monitored. The results of the monitoring shall provide an input into the supplier re-evaluation process. Non-fulfilment of purchasing requirements shall be addressed with the supplier proportionate to the risk associated with the purchased product and compliance with applicable regulatory requirements. Records of the results of evaluation, selection, monitoring and re-evaluation of supplier capability or performance and any necessary actions arising from these activities shall be maintained (see 4.2.5). 组织将确保采购的产品符合规定的采购信息的程序形成文件（见 4.2.4）。组织应建立评价和选择供方的准则，准则应： a) 基于供方提供符合组织要求产品的能力； b) 基于供方的绩效； c) 基于采购产品对医疗器械质量的影响； d) 与医疗器械有关风险相一致。 组织应对供方的监视和再评价进行策划。应监视供方满足采购产品要求的绩效。监视的结果应作为供方再评价过程提供输入。 应表述不满足的采购 要求和 相应的有对 应风险的采购 产品的供方，并符合适用 的法规要求 。 对供方评价的结果、选择、监视和再评价的记录或因这些活动所采取的任何必要 措施的记录应予以保持（见 4.2.5）。	未提供对紧固件供应商进行识别并进行合格供方监视和评价的有效证据，违反了《供方评估和复评程序》的有关规定。 查采购信息，如： 1) 供应商：雄县雄州镇久奥座垫加工厂， 日期：15/05/2026，产品：安全带，数量：800条； 2) 供应商：河北雄安远智医疗科技有限公司，日期：25/06/2026，产品：黑色车架（碳钢），数量：800批； 3) 供应商：河北雄安耀鑫商贸有限公司，日期：21/06/2026，产品：越野轮胎，数量：1500条； 4) 供应商：雄县雄州镇久奥座垫加工厂， 日期：18/05/2026，产品：坐垫、背靠，数量：各500个； 5) 供应商：雄县靖凯商贸有效公司， 日期：15/06/2026，产品：塑料踏板，数量：850个； 采购信息规定了货品名称、数量、单价，金额等信息，符合要求。供方每批产品通过收集供方检验报告、进行来料验收等方式，避免产品不符合医疗器械质量影响的风险发生。	
7.4.2	7.4.2 Purchasing information采购信息 Purchasing information shall describe or reference the product to be purchased, including as appropriate: a) product specifications; b) requirements for product acceptance, procedures, processes and equipment; c) requirements for qualification of supplier personnel; d) quality management system requirements. The organization shall ensure the adequacy of specified purchasing requirements prior to their communication to the supplier. Purchasing information shall include, as applicable, a written agreement that the supplier notify the organization of changes in the purchased product prior to implementation of any changes that affect the ability of the purchased product to meet specified purchase requirements. To the extent required for traceability given in 7.5.9, the organization shall maintain relevant purchasing information in the form of documents (see 4.2.4) and records (see 4.2.5). 拟采购产品的采购信息应表述或引用，适当时包括： a) 产品规范； b) 产品接受准则、程序、过程和设备的要求；		

Clause 条款	Requirements 要求	Audit Evidence 审核证据	NC/OFI
	c) 供方人员资质要求； d) 质量管理体系的要求。 组织应确保在与供方沟通前所规定的采购要求是充分与适宜的。 适用时，采购信息应包含书面协议，该协议明确了在影响采购产品符合规定采购要求的能力的任何变更实施前，供方应将采购方面的变更通知组织。 按照 7.5.9 规定的可追溯性要求的范围和程度，组织应以文件（见 4.2.4）和记录（见 4.2.5）的形式保持相关的采购信息。		