



CC-2294

**Nishitronics
Instrumentation***We work for Customers Satisfaction*

Page 1 of 2

CALIBRATION CERTIFICATE

Customer : Yog Electro Process Pvt. Ltd. (Calibration Lab)	ULR No. : CC229426000000200F
	Discipline : Thermal
	Certificate No. : NI2026/01/0148
	Date of Issue : 10/01/2026
	Date of Receipt : 08/01/2026
	Date of Calibration : 09/01/2026
	Cal.Req.No. : 26/01/34/3
	Next Due Date : 08/01/2027
	Sub Grp. / Parameter : Temperature
	Calibration Method No. : NI / CP / T / 01

Plot No. G-63/8, Chakan Ind.Area, Phase III, Village
Kuruli, Tal-Khed, Dist.- Pune, Maharashtra Pin Code : 410501

Env.Conditions Temp.25±4°C, 60 ±15% RH

Location of Calibration : At Lab

Characteristic and Condition of Item : SATISFACTORY

Details of Items :

Name : RTD SENSOR with Indicator	Range : -50 to 400 °C
ID No. : YOG-27 , YOG-35	L.C. : 0.01 °C
Sr.No. : 22020573	Accuracy : Not Specified
Make : YOG / Yudian (AI5600)	Location : -
Model : -	
Type : PT-100	

Master Instruments Detail

Nomenclature	Serial No.	Id.No.	Valid Upto	Calibrated At./ CC NO.
SSPRT WITH INDICATOR	01294 / 33290042	NI/SSPRT/01	16/06/2027	GODREJ &BOYCE M-250606-30-2
SS PRT WITH INDICATOR	331134-29/332900 42	NI/SSPRT/02	16/06/2027	GODREJ &BOYCE M-250606-30-1

Observation**Calibration Results**

Calibration Point °C	Standard Reading °C	Standard Reading (Ω)	UUC Reading °C	Error °C	Expanded Uncertainty(±°C) (K=2)
-25	-25.029	89.912	-25.08	-0.051	0.040
-20	-20.021	91.921	-20.07	-0.049	0.040
-10	-10.034	95.918	-10.06	-0.026	0.040
0	0.019	99.931	0.03	0.011	0.030
25	24.991	109.736	25.02	0.029	0.030
50	49.990	119.397	50.04	0.050	0.030
100	99.993	138.503	100.06	0.067	0.030
200	199.985	175.842	200.08	0.095	0.060
300	299.978	212.038	300.10	0.122	0.060
400	399.982	247.104	400.14	0.158	0.060

Calibrated By

P.B.Desai

Technical Assistant



Reviewed & Authorized By

Digitally signed by Vijay Hingmire
Date: 2026.01.10 12:39:34 +05:30
Reason: Calibration Certificate
Location: Pune

Certificate No. : NI2026/01/0148

ULR No. : CC229426000000200F

The reported measurement uncertainty is estimated at a level of confidence of approximately 95 % with a coverage factor $k = 2$.

Remarks :


- 1) Result are related only to the item calibrated.
- 2) This certificate refers only to the particular items submitted for calibration.
- 3) This certificate shall not be reproduced except in full without our prior permission in writing.
- 4) The calibration results reported in this particular certificate are valid at the time of an under stated condition of measurement.
- 5) Standard used for calibration were traceable to National / International standard.
- 6) Readings given above are as on received condition of an instrument.
- 7) Above calibration done is meant for scientific & industrial purpose only.
- 8) Temperature Scale - ITS 90.
- 9) Immersion Depth of UUC was 150 mm.
- 10) Calibration Points as suggested by Customer.

Calibrated By



P.B.Desai

Technical Assistant



Reviewed & Authorized By

Digitally signed by Vijay Hingmire
Date: 2026.01.10 12:39:34 +05:30
Reason: Calibration Certificate
Location: Pune

--- END OF CERTIFICATE ---