



CC-2294

**Nishitronics
Instrumentation***We work for Customers Satisfaction***CALIBRATION CERTIFICATE**

Page 1 of 2

Customer : Yog Electro Process Pvt. Ltd. (Calibration Lab)	ULR No. : CC229426000000202F
	Discipline : Thermal
	Certificate No. : NI2026/01/0150
Plot No. G-63/8, Chakan Ind.Area,Phase III, Village	Date of Issue : 10/01/2026
Kuruli,Tal-Khed, Dist.- Pune, Maharashtra Pin Code : 410501	Date of Receipt : 08/01/2026
	Date of Calibration : 09/01/2026
Env.Conditions Temp.25±4°C, 60 ±15% RH	Cal.Req.No. : 26/01/34/2
Location of Calibration : At Lab	Next Due Date : 08/01/2027
Characteristic and : SATISFACTORY	Sub Grp. / Parameter : Temperature
Condition of Item	Calibration Method No. : NI / CP / T / 01

Details of Items :

Name : S Type Thermocouple with Indicator	Range : 0 to1200 °C
ID No. : YOG-04 ,YOG-35	L.C. : 0.1 °C
Sr.No. : 60657	Accuracy : Not Specified
Make : Tempsens / Yudian (AI5600)	Location : -
Model : -	
Type : Type S	


Master Instruments Detail

Nomenclature	Serial No.	Id.No.	Valid Upto	Calibrated At./ CC NO.
SS PRT WITH INDICATOR	331134-29/332900 42	NI/SSPRT/02	16/06/2027	GODREJ & BOYCE M-250606-30-1
Type-S Thermocouple with Cold Junction	01 / HP2823A12705	NI/S/TH/02	25/04/2027	NPL


Observation**Calibration Results**

Calibration Point °C	Standard Reading °C	Standard Reading (mV/Ω)	UUC Reading °C	Error °C	Expanded Uncertainty± °C
0	0.024	100.014	-0.1	-0.124	0.356
400	399.994	247.108	399.8	-0.194	0.356
600	599.990	313.774	599.7	-0.290	0.356
800	799.982	7.342	799.5	-0.482	1.000
1000	999.987	9.584	999.3	-0.687	1.021
1200	1199.983	11.947	1199.0	-0.983	1.021

Calibrated By


P.B.Desai

Technical Assistant



Reviewed & Authorized By

Digitally signed by Vijay Hingmire
Date: 2026.01.10 15:36:47 +05:30
Reason: Calibration Certificate
Location: Pune

Certificate No. : NI2026/01/0150

ULR No. : CC229426000000202F

The reported measurement uncertainty is estimated at a level of confidence of approximately 95 % with a coverage factor $k = 2$.

Remarks :


- 1) Result are related only to the item calibrated.
- 2) This certificate refers only to the particular items submitted for calibration.
- 3) This certificate shall not be reproduced except in full without our prior permission in writing.
- 4) The calibration results reported in this particular certificate are valid at the time of an under stated condition of measurement.
- 5) Standard used for calibration were traceable to National / International standard.
- 6) Readings given above are as on received condition of an instrument.
- 7) Above calibration done is meant for scientific & industrial purpose only.
- 8) Temperature Scale - ITS 90.
- 9) Coloumn no. 3 values given read unit in ohm upto 650 °C & above 650 °C read in mV.
- 10) Calibration Points as suggested by Customer.

Calibrated By



P.B.Desai

Technical Assistant



Reviewed & Authorized By

Digitally signed by Vijay Hingmire
Date: 2026.01.10 15:36:47 +05:30
Reason: Calibration Certificate
Location: Pune

--- END OF CERTIFICATE ---