



CC-2294

**Nishitronics
Instrumentation***We work for Customers Satisfaction*

Page 1 of 3

CALIBRATION CERTIFICATE

Customer : Yog Electro Process Pvt. Ltd. (Calibration Lab)

Plot No. G-63/8, Chakan Ind.Area, Phase III, Village
Kuruli, Tal-Khed, Dist.- Pune, Maharashtra Pin Code : 410501

Env. Conditions Temp. $23 \pm 1.5^{\circ}\text{C}$, $50 \pm 10\%$ RH

Location of Calibration : At Lab

Characteristic and Condition of Item : SATISFACTORY

ULR No. : CC229426000000208F

Discipline : Electro Technical

Certificate No. : NI2026/01/0160

Date of Issue : 09/01/2026

Date of Receipt : 08/01/2026

Date of Calibration : 08/01/2026

Cal.Req.No. : 26/01/34/1

Next Due Date : 07/01/2027

Parameter : Temperature

Calibration Method No. : NI/CP/ET/01

Details of Items :

Name : Digital Thermometer

ID No. : YOG-36

Sr.No. : 241089148

Make : Yudian

Model : AI5600

Type : Digital

Range : As per Specification

L.C. : As per Specification

Accuracy : As per Manual

Location :

Master Instruments Detail

Nomenclature	Serial No.	Id.No.	Valid Upto	Calibrated At./ CC NO.
Multifun. Calibrator with Current Coil	1204A12	NI/MFC/02	31/03/2026	NISHITRONICS INSTRUMENTATION


Observation**Temperature Simulation**

Mode	Type	Standard Readings	Unit of Measure	UUC Readings	Error	Expanded Uncertainty ($\pm$ $^{\circ}\text{C}$) (k = 2)
Measure	Type K	-100.00	$^{\circ}\text{C}$	-99.98	0.02	0.15
Measure	Type K	0.00	$^{\circ}\text{C}$	0.03	0.03	0.15
Measure	Type K	200.00	$^{\circ}\text{C}$	200.20	0.20	0.15
Measure	Type K	400.00	$^{\circ}\text{C}$	400.08	0.08	0.15
Measure	Type K	600.00	$^{\circ}\text{C}$	600.10	0.10	0.15
Measure	Type K	800.00	$^{\circ}\text{C}$	800.18	0.18	0.15
Measure	Type K	1000.00	$^{\circ}\text{C}$	1000.19	0.19	0.15
Measure	Type K	1200.00	$^{\circ}\text{C}$	1200.19	0.19	0.15
Measure	Type J	-100.00	$^{\circ}\text{C}$	-99.89	0.11	0.12
Measure	Type J	0.00	$^{\circ}\text{C}$	0.03	0.03	0.12
Measure	Type J	100.00	$^{\circ}\text{C}$	100.03	0.03	0.12
Measure	Type J	200.00	$^{\circ}\text{C}$	200.04	0.04	0.12
Measure	Type J	400.00	$^{\circ}\text{C}$	400.02	0.02	0.12

Calibrated By

T.T. Dhumal

Technical Assistant



Reviewed & Authorized By


Digitally signed by Jay Hingmire
Date: 2026.01.09 13:02:30 +05:30
Reason: Calibration Certificate
Location: Pune

Certificate No. : NI2026/01/0160


ULR No. : CC229426000000208F

Measure	Type J	600.00	°C	600.10	0.10	0.12
Measure	Type J	800.00	°C	800.12	0.12	0.12
Measure	Type N	-100.00	°C	-99.89	0.11	0.24
Measure	Type N	0.00	°C	0.02	0.02	0.24
Measure	Type N	200.00	°C	200.10	0.10	0.24
Measure	Type N	400.00	°C	400.05	0.05	0.24
Measure	Type N	600.00	°C	600.03	0.03	0.24
Measure	Type N	800.00	°C	800.10	0.10	0.24
Measure	Type N	1000.00	°C	999.99	-0.01	0.24
Measure	Type N	1200.00	°C	1200.14	0.14	0.24
Measure	Type T	-100.00	°C	-99.98	0.02	0.32
Measure	Type T	-50.00	°C	-49.97	0.03	0.32
Measure	Type T	0.00	°C	0.02	0.02	0.32
Measure	Type T	50.00	°C	49.97	-0.03	0.32
Measure	Type T	100.00	°C	99.97	-0.03	0.32
Measure	Type T	200.00	°C	200.13	0.13	0.32
Measure	Type T	300.00	°C	300.10	0.10	0.32
Measure	Type T	400.00	°C	399.96	-0.04	0.32
Measure	PT-100	-50.000	°C	-49.98	0.020	0.090
Measure	PT-100	-25.000	°C	-24.99	0.010	0.090
Measure	PT-100	-10.000	°C	-9.98	0.020	0.090
Measure	PT-100	0.000	°C	0.02	0.020	0.090
Measure	PT-100	25.000	°C	24.99	-0.010	0.090
Measure	PT-100	50.000	°C	49.98	-0.020	0.090
Measure	PT-100	100.000	°C	99.98	-0.020	0.090
Measure	PT-100	200.000	°C	199.98	-0.020	0.090
Measure	PT-100	300.000	°C	299.96	-0.040	0.090
Measure	PT-100	400.000	°C	399.96	-0.040	0.090
Measure	Type R	0.000	°C	0.2	0.200	0.480
Measure	Type R	200.000	°C	200.1	0.100	0.480
Measure	Type R	400.000	°C	400.0	0.000	0.480
Measure	Type R	600.000	°C	600.0	0.000	0.480
Measure	Type R	800.000	°C	800.1	0.100	0.480
Measure	Type R	1000.000	°C	1000.1	0.100	0.480
Measure	Type R	1200.000	°C	1200.1	0.100	0.480
Measure	Type S	0.000	°C	0.0	0.000	0.480
Measure	Type S	200.000	°C	200.1	0.100	0.480
Measure	Type S	400.000	°C	399.9	-0.100	0.480
Measure	Type S	600.000	°C	599.8	-0.200	0.480
Measure	Type S	800.000	°C	800.0	0.000	0.480

Calibrated By


T.T. Dhumal

Technical Assistant



Reviewed & Authorized By

Digitally signed by: Jay Hingmire
Date: 2026.01.09 13:02:30 +05:30
Reason: Calibration Certificate
Location: Pune

Certificate No. : NI2026/01/0160

ULR No. : CC229426000000208F


Measure	Type S	1000.000	°C	1000.0	0.000	0.480
Measure	Type S	1200.000	°C	1199.9	-0.100	0.480

The reported measurement uncertainty is estimated at a level of confidence of approximately 95 % with a coverage factor $k = 2$.


Remarks :

- 1) Result are related only to the item calibrated.
- 2) This certificate refers only to the particular items submitted for calibration.
- 3) This certificate shall not be reproduced except in full without our prior permission in writing.
- 4) The calibration results reported in this particular certificate are valid at the time of an under stated condition of measurement.
- 5) Standard used for calibration were traceable to National / International standard.
- 6) Readings given above are as on received condition of an instrument.
- 7) Above calibration done is meant for scientific & industrial purpose only.
- 8) Calibration points as suggested by Customer.

Calibrated By



T.T. Dhumal
Technical Assistant



Reviewed & Authorized By

Digitally signed by Vijay Hingmire
Date: 2026.01.08 13:02:30 +05:30
Reason: Calibration Certificate
Location: Pune

--- END OF CERTIFICATE ---