Title:

Certificate of Analysis (CoA)

Date: 2/7/2026
Date Tested: 2/6/2026
Customer: Forge Amino
Testing material: SS-31
Lot Number: FA-MTS3012601
BT Sample ID: 005000038768403
Labeled Peptide Content/Potency: 10 mg
Storage: R.T.
Visual Description: small clear vial: white sample, white label, silver crimp, dark blue plastic cap.
Labeled as: SS-31
Manufacturer: Forge Amino
Testing Purpose: FTIR and HPLC analysis for the identification, purity, potency and composition of a peptide product. It does not provide information on particulate matter, microbial contamination or presence of endotoxins.

Test	Method	Specification	Result
General Appearance	USP <630>	White powder	White powder
Mass	USP <41>	As recorded	131.7 mg
FTIR Identification and Composition Analysis	USP <197A>	Sample spectrum should confirm the content of peptide via characteristic bands	FTIR sample spectrum confirms the presence of SS-31 with addition of excipient(s)/fillers such as mannitol.
HPLC Purity of Peptide Assay	USP <621>	Specifications: ≥ 98 %	99.8 %
HPLC Potency Assay	USP <621>	Specifications: 95 – 105% of 10 mg	16 mg (160.3 %)

The results of the CoA relate only to the item(s) tested and applied to the sample as received.

Andrea Castro, AS
Scientist-I
BTLabs

Verna Zheng, AS
Scientist-I
BTLabs

5730 Corporate Way | Suite 220 | West Palm Beach, FL 33407
Phone: (561) 625-0133
E-mail: info@btlabtesting.com | Website: https://btlabtesting.com

Title:

Certificate of Analysis (CoA)

FTIR ID and Composition Analysis: SS-31 Lot FA-MTS3012601



HPLC Purity and Potency Assay @ 280 nm: SS-31 Lot FA-MTS3012601



SS-31 Lot FA-MTS3012601 @ 280 nm		
Peak #:	Retention Time (min)	Area (mAU*s)
1	4.017	769.3

Title:

Certificate of Analysis (CoA)

Bacterial Endotoxins Test (ABI- TEST 0012)

Date: 2/8/2026
Date Tested: 2/6/2026
Customer: Forge Amino
Testing material: SS-31
Lot Number: FA-MTS3012601
BT Sample ID: 005000038768403
Labeled Peptide Content/Potency: 131.7 mg
Storage: R.T.
Packing material: small clear vial: white sample, white label, silver crimp, dark blue plastic cap.
Labeled as: SS-31
Manufacturer: Forge Amino
Testing Purpose: (ABI-TEST 0012) test quantifies bacterial endotoxin levels to confirm that the product does not elicit pyrogenic reactions, maintaining patient safety and product integrity.

Test	Method	Specification	Result
Bacterial Endotoxins Test	USP <85>	40 EU/vial reconstituted	PASS

The results of the CoA relate only to the item(s) tested and applied to the sample as received.

Bacterial Endotoxins Test SS-31 Lot FA-MTS3012601

Sample ID	Description	Volume (mL)	Endotoxin Limit (EU/vial)	Lysate Sensitivity (λ, EU/mL)	Result
005000038768403	SS-31	3	40 EU/vial	0.25	9.5 EU/vial

The sample passed a valid gel clot endotoxin test based on the type, mode of administration and dosage.

Andrea Castro. AS
Scientist-I
BTLabs

Verna Zheng. AS
Scientist-I
BTLabs