




Inspection d'entrée

Information endoscope.

Modèle	0	**Ordre de rep.**	0	**CUS**	
Numéro de série:	0			**LOA**	
Nom du client:	0			**DEM**	
Date d'inspection:	27/10/2026	**Technicien Inspection:**			
Date contrôle qualité:		**Technicien reparation1:**			
Pistons	☐ Inclus ☐ Absents	**Technicien reparation2:**			
Accesoires:		**Technicien reparation3:**			
		Controleur Qualité	0		
Problème client:	0				

	Inspection Items			Inspection entrée	Contrôle apres REP
	**Inspection1 visuelle G5/G7 OSA**			OK ☐ NG ☐	OK ☐ NG ☐
	Inspection externe ISA/DTA			OK ☐ NG ☐	OK ☐ NG ☐
	DTA diametre exterieur		**#N/A**	OK ☐ NG ☐	OK ☐ NG ☐
	FSA diametre exterieur		**#N/A**	OK ☐ NG ☐	OK ☐ NG ☐
	Longueur de travail		**#N/A**	OK ☐ NG ☐	OK ☐ NG ☐
	Inspection3 visuelle FSM & Con			OK ☐ NG ☐	OK ☐ NG ☐
	**Test d'étancheité**			OK ☐ NG ☐	OK ☐ NG ☐
	Pression 24kPa immergé			**40** kPa	kPa
	**Inspection Image**			OK ☐ NG ☐	OK ☐ NG ☐
	Uses :	**entrée**	**sortie**		
		679			
	Test d'etancheité con.			OK ☐ NG ☐	OK ☐ NG ☐
	Test de condensation			OK ☐ NG ☐	OK ☐ NG ☐
	SW1			OK ☐ NG ☐	OK ☐ NG ☐
	SW2			OK ☐ NG ☐	OK ☐ NG ☐
	SW3			OK ☐ NG ☐	OK ☐ NG ☐
	SW4			OK ☐ NG ☐	OK ☐ NG ☐
	SW5			OK ☐ NG ☐	OK ☐ NG ☐
	EAU cc/min		**#N/A**	OK ☐ NG ☐	OK ☐ NG ☐
	Air cc/5sec		**#N/A**	OK ☐ NG ☐	OK ☐ NG ☐
	ASPI cc/min		**#N/A**	OK ☐ NG ☐	OK ☐ NG ☐
	Jet chanel			OK ☐ NG ☐	OK ☐ NG ☐
	**Canal Ballonnet 580UT/UR***			OK ☐ NG ☐	OK ☐ NG ☐
	FCT gauge		**#N/A**	OK ☐ NG ☐	OK ☐ NG ☐
	Rigidité variable			OK ☐ NG ☐	OK ☐ NG ☐
	Inspection	**haut**	**#N/A**	OK ☐ NG ☐	OK ☐ NG ☐
	Gauche / droite	**bas**	**#N/A**	OK ☐ NG ☐	OK ☐ NG ☐
	Haut/bas	**gauche**	**#N/A**	OK ☐ NG ☐	OK ☐ NG ☐
	Angulation	**droite**	**#N/A**	OK ☐ NG ☐	OK ☐ NG ☐
	Freins			OK ☐ NG ☐	OK ☐ NG ☐
Type de réparation	☑ **Mineur** ☐ **Majeur** ☐ **Inspection simple**				

Comment: * Le canal du ballon doit être vérifié avec l'ecouvillon