



Rui Filipe Cantarino Valente de Almeida

Master of Science

Development of a Tomographic Atmospheric Monitoring System based on Differential Optical Absorption Spectroscopy

Thesis plan submitted in partial fulfillment
of the requirements for the degree of

Doctor of Philosophy in
Biomedical Engineering

Adviser: Pedro Vieira, Auxiliar Professor,
NOVA University of Lisbon



FACULDADE DE
CIÊNCIAS E TECNOLOGIA
UNIVERSIDADE NOVA DE LISBOA

August, 2019

Contents

1	Introduction	1
1.1	Background and Motivation	1
1.1.1	Background	1
1.1.2	The Problem	1
1.1.3	Objectives	2
1.1.4	Methods	2
2	Research Question	5
2.1	Problem Introduction	5
2.2	Research Question	6
2.3	Hypothesis and Approach	7
3	Literature Review	9
3.1	Air pollution and pollutants	9
3.2	Tomographic algorithms and reconstruction techniques	10
3.2.1	Introduction	10
3.2.2	Initial Considerations	11
3.2.3	The Fourier Slice Theorem	12
3.2.4	The Filtered BackProjection Algorithm	13
3.3	DOAS tomography instrumentation	13
4	Research Methodology	15
4.1	Aimed contribution	15
4.2	Detailed work plan and scheduling	15
4.3	Validation methodology	15
4.4	Integration with other research activities	15
	Bibliography	17

List of Figures

3.1	A schematic representation of a projection.	10
3.2	Schematic representation for coordinate setting.	11
3.3	The FST, a schematic representation.	13

List of Tables

2.1	Main research question.	6
2.2	Secondary research questions.	7

Glossary

aliquam	tincidunt urna. Nulla ullamcorper vestibulum turpis. Pellentesque cursus luctus mauris.
computer	An electronic device which is capable of receiving information (data) in a particular form and of performing a sequence of operations in accordance with a predetermined but variable set of procedural instructions (program) to produce a result in the form of information or signals.
cras viverra	metus rhoncus sem. Nulla et lectus vestibulum urna fringilla ultrices. Phasellus eu tellus sit amet tortor gravida placerat.
donec nonummy	pellentesque ante. Phasellus adipiscing semper elit. Proin fermentum massa ac quam. Sed diam turpis, molestie vitae, placerat a, molestie nec, leo.
integer sapien	est, iaculis in, pretium quis, viverra ac, nunc. Praesent eget sem vel leo ultrices bibendum. Aenean faucibus.
lorem ipsum	dolor sit amet, consectetur adipiscing elit. Ut purus elit, vestibulum ut, placerat ac, adipiscing vitae, felis. Curabitur dictum gravida mauris.
maecenas lacinia	nam ipsum ligula, eleifend at, accumsan nec, suscipit a, ipsum. Morbi blandit ligula feugiat magna. Nunc eleifend consequat lorem.
morbi ac	orci et nisl hendrerit mollis. Suspendisse ut massa. Cras nec ante. Pellentesque a nulla. Cum sociis natoque penatibus et magnis dis parturient montes, nascetur ridiculus mus.

GLOSSARY

morbi dolor	nulla, malesuada eu, pulvinar at, mollis ac, nulla. Curabitur auctor semper nulla. Donec varius orci eget risus. Duis nibh mi, congue eu, accumsan eleifend, sagittis quis, diam. Duis eget orci sit amet orci dignissim rutrum.
nam lacus	libero, pretium at, lobortis vitae, ultricies et, tellus. Donec aliquet, tortor sed accumsan bibendum, erat ligula aliquet magna, vitae ornare odio metus a mi.
nam dui	ligula, fringilla a, euismod sodales, sollicitudin vel, wisi. Morbi auctor lorem non justo.
name arcu	libero, nonummy eget, consectetur id, vulputate a, magna. Donec vehicula augue eu neque. Pellentesque habitant morbi tristique senectus et netus et malesuada fames ac turpis egestas. Mauris ut leo.
nulla malesuada	porttitor diam. Donec felis erat, congue non, volutpat at, tincidunt tristique, libero. Vivamus viverra fermentum felis.
sed lacinia	nulla vitae enim. Pellentesque tincidunt purus vel magna. Integer non enim. Praesent euismod nunc eu purus. Donec bibendum quam in tellus.

Acronyms

AP	Air Pollution
CT	Computed Tomography
DOAS	Differential Optical Absorption Spectroscopy
EPA	Environmental Protection Agency (United States)
FBP	Filtered BackProjection
FFF	Forest Fire Finder
FST	Fourier Slice Theorem
FT	Fourier Transform
IFT	Inverse Fourier Transform
ML	Machine Learning
PM	Particulate Matter
ROI	Region Of Interest
RQ	Research Question
SLR	Systematic Literature Review
WHO	World Health Organization

Symbols

*

Introduction

1.1 Background and Motivation

1.1.1 Background

The idea behind this thesis was born in 2015, at NGNS-IS (a Portuguese tech startup). At the time, the company's flagship product was the Forest Fire Finder (**FFF**), which was a forest fire detection system that performed a spectroscopical analysis of the atmosphere and then, through some Machine Learning (**ML**) techniques, could detect the presence of a smoke column above the horizon and alert the operators.

The growing importance of Air Pollution (**AP**) in today's society, and the fact that the system was already scanning the atmosphere for some chemical components originated and motivated the idea behind this thesis. Although **FFF** was already a spectroscopic system, it was constructed to operate in remote and inhospitable locations, and its design had had no spatial constraints into account. In addition, the system scanning method meant that it was not appropriate for pollution measurement, as it could only detect a mean pollutant column density for each spectrum it took. A truly useful monitoring tool would be able to map these pollutants concentration, thus retrieving the same kind of information as a network of in-situ electro-chemical sensors.

There were, at the time, a few research projects that aimed to bring this kind of capacity to Differential Optical Absorption Spectroscopy (**DOAS**) based systems. However, there was ample room for improvement, as it was made clear by the lack of commercial systems and the scarcity of the literature on the subject. Realizing the type of research that accomplishing this project would entail, the company decided to publish this PhD Project, in a tripartite consortium with FCT-NOVA and the Portuguese Foundation for Science and Technology.

1.1.2 The Problem

Air Pollution (**AP**) is one of the grave concerns of modern day western society, with many decades worth of research proving that it can have a pronounced negative effect

on human, animal and plant life, as shown in Section ???. On humans, it has been shown to significantly increase risk of cardiovascular, pulmonary and even neuropsychiatric diseases [2, 4, 11]. Its implications on ecosystems are remarkably complex and difficult to quantify, but nonetheless extremely important, and have a huge impact on biodiversity [12].

Knowing all this brings us the responsibility of at least trying to mitigate some of these adverse consequences of the spectacular progress that we have achieved in the few last centuries. But we cannot act unless we also know what we must do; and to know this, we must have measurements.

1.1.3 Objectives

The overarching goal of this thesis was to theorize and design a bidimensional mapping tool for trace atmospheric pollutants such as NO_x and SO_x , using DOAS as the measurement technique. In order to maximize commercial value (and viability), the system had to be small and mobile. During the research, several "micro-objectives" appeared regularly. Some were kept and incorporated in the workplan, others discarded after initial exploration. The main secondary objectives were:

- To use a tomographic approach for the mapping procedure;
- To ensure the designed system would be small and highly mobile;
- To use a single collection point, minimizing material costs for the technology.

These objectives allowed setting several research questions, which are introduced in Section 2.2.

1.1.4 Methods

To address these goals, I assumed the development of this thesis to be essentially split in two parts, which are to be explored simultaneously. They can (coarsely) be addressed as *tomography* and *instrumentation*. On the tomography side, it will be necessary to study what are the more appropriate algorithms (and what type of tomography), how they can be physically deployed (i.e., the problem's geometry) and what type of reconstruction method is the more favorable. Regarding the instrumentation, there are also several points that need considering: decisions are required with regard to the mechanics, the controls and the optical components of the final system. A more detailed discussion of these topics can be found in Section 2.3 and Chapter ??.

One of the most important steps in the development of this work is designing and implementing a simulation software platform, that allows the validation of the acquisition strategy, the geometry selection and the reconstruction approach. This endeavor will also be of critical importance for component selection, since it will define

the component requirements for the whole system. On the optical instrumentation side, this project will require optimizing FFF's optical assembly. Although similar in purpose and types of components, this assembly is significantly larger than what is acceptable for this project and needs redesigning.

Research Question

2.1 Problem Introduction

AP is a very important topic of discussion in the current days, with scientists and researchers around the globe being very well aware of the potential effects it can have on the health of individuals and populations across all ecosystems. Not to mention its implications on climate change, which are generally regarded as one of the capital threats to life on Earth's survival (on par with a nuclear apocalypse). Defining AP can be a challenge. In fact, its effects and presence is so all-encompassing, that it would be fair to say that its definition changes with the angle with which one looks upon it. Nonetheless, it is important to at least try to define it, in order to approach it in some way [4, 12].

The United States Environmental Protection Agency (EPA) defined Air Pollution (AP) as "*the presence of contaminants or pollutant substances in the air that interfere with human health or welfare, or produce other harmful environmental effects*" [13]. This is (perhaps intentionally) a very broad definition, too broad to avoid vagueness. It does introduce a key concept: the term *pollutant*, which needs be discussed in order to complete the definition above.

It would be very hard to find someone who did not have an almost instinctive idea of what a pollutant is. We know something is amiss when we notice our air is full of smoke or smells strange, but our senses are not enough. There are many chemical components that are untraceable by unaided humans, and some that are only detected by our noses and eyes at concentration levels which are above the threshold where they can damage our health. This makes the task of separating pollutants from non-pollutants a non-trivial one. If we cannot rely solely on our senses to detect them, then it is up to the scientists and engineers to create methods that allow us to do so. Whats more, we must also rely on them to understand how can a normally harmless substance be a pollutant, depending on the circumstance. For instance, nitrous compounds are traditionally beneficial to the soils and cultures, but they can and do cause pulmonary and cardiovascular complications in humans [2, 4, 11].

Context matters to pollutants. The toxic nature of a certain chemical only is revealed when someone or something gets exposed to it. Even then, there are exposure levels which do not bear any effects, good or bad. At these levels, a pollutant is but an impurity. There too many potential pollutants in our modern day world to list here, but the World Health Organization (WHO) states that there are six major air pollutants:

- Particle Matter (PM);
- Ground level ozone (O₃);
- Carbon monoxide (CO);
- Sulfur Oxides (SO_x);
- Nitrous Oxides (NO_x);
- Lead (Pb).

Exposure to these pollutants have different effects on humans, ranging in seriousness from skin irritation to neuropsychiatric complications, depending on dose and on the time the exposure lasts.

2.2 Research Question

In Chapter 1, I have introduced the reasons which led NGNS-IS to pursue the development of an atmospheric monitoring system, and that what set it apart from other systems was the ability to spectroscopically map pollutants concentrations using tomographic methods, thus defining a primary objective for this thesis.

Two secondary objectives were born from the necessary initial research, which had a very heavy influence over the adopted methods:

- To use a tomographic approach for the mapping procedure;
- To ensure the designed system would be small and highly mobile;
- To use a single light collection point, minimizing material costs.

Taking all the above into account, we arrive at the main Research Question (RQ), presented in Table 2.1.

Table 2.1: Main research question.

RQ1	<i>How to design a miniaturized tomographic atmosphere monitoring system based on DOAS?</i>
------------	---

This is the main research question. It gave rise to four other more detailed research questions. These secondary questions allow a better delimitation of the work at hand and are important complements to RQ1. This questions are presented in Table 2.2

Table 2.2: Secondary research questions.

RQ1.1	<i>What would be the best strategy for the system to cover a small geographic region?</i>
RQ1.2	<i>What would be the necessary components for such a system?</i>
RQ1.3	<i>How will the system acquire the data?</i>
RQ1.4	<i>What should the tomographic reconstruction look like and how to perform it?</i>

2.3 Hypothesis and Approach

This work is based on the hypothesis that a system such as the one described in Chapter 1, which responds to the RQ in Table 2.1 and Table 2.2 can be achieved by careful selection of mathematical tomographic algorithms and instrumentation that is able to implement them correctly.

The first step in answering the entirety of the research questions should be to answer RQ1.1. In fact, it is not possible to make any other decision before this matter is settled. As with any technical problem, there are several ways to create a tomographic atmospheric monitoring tool. However, each and every one of them implies some kind of compromise, which determines the system's capabilities and requirements. Will the system use retro-reflection? Shall it move during the measurement? These are the kind of questions that determine the whole project.

When the measurement strategy is determined, one could start picking parts and components. However, a better first approach would be designing a software simulation. This simulator must include all major system features, so that it correctly mimics reality and is therefore able to mathematically validate the acquisition and reconstruction approach. The results obtained from the simulation will then dictate mechanical and control requirements.

One other aspect that needs addressing is the optical section. As mentioned before, the system will be inspired in FFF's basic optical capabilities. However, the smoke detector was not conceived with spatial restrictions in mind. This important set of components will thus need redesigning, so that it is in line with the size objectives of the new system.

Literature Review

In this chapter, I provide a literary review on the three most important subjects for the work of this thesis: [AP](#), tomographic algorithms and [DOAS](#) tomography instrumentation.

3.1 Air pollution and pollutants

As stated in Section 2.2, the definition of [AP](#) is dependent on the context. Here, I will focus especially on the effects of pollutants on human health. Whether these effects are the most significant problems stemming from [AP](#) is debatable (climate change is mostly caused by anthropogenic production of greenhouse gases, which are pollutants) but for this system and its intended uses, health effects are definitely more prominent. Human health implications of a polluted atmosphere are documented in very numerous studies throughout the literature. In this document, I will only present a small number of representative reviews and reports.

In 2004, [WHO](#) published a report summarizing what was then the most recent information on health effects of air pollution over Europe. This review concluded that, even with all the regulations on [AP](#) put in place by the European authorities, its levels were still posed a considerable burden on health throughout Europe [14].

Although there are several hundred potentially harmful components already that have already been found in the atmosphere, this report addresses only [PM](#), ground level Ozone and Nitrogen Dioxide. As many other studies had found, this Systematic Literature Review ([SLR](#)) identified several short-term and long-term exposure effects for the three pollutants. The study found that short-term exposure to all three substances were responsible for an increase in mortality and hospital admissions, and that both [PM](#) and O_3 increased the population's usage of medication. Long-term exposure to all three components have adverse pulmonary effects, but [PM](#) have many other negative effects. The most important of them a reduction in life expectancy, which the authors attribute to cardiopulmonary mortality and lung cancer.

Particulate Matter are described as airborne solid particles or droplets. These

particles vary in size, origin and composition, however, it is usual to classify them by size, since that is what governs particle deposition in the respiratory system. Urban **PM** are usually divided into three categories: coarse, fine and ultrafine. Convention dictates that coarse particles have an aerodynamic diameter of less than $10\mu\text{m}$ (PM_{10}), fine particles less than $2.5\mu\text{m}$ ($\text{PM}_{2.5}$). In this review, the authors stated that the role that ultrafine particles play in human health is still undetermined, but they noted that coarse and (especially) fine particles are highly correlated with an increase in mortality and the prevalence of respiratory syndromes.

3.2 Tomographic algorithms and reconstruction techniques

3.2.1 Introduction

Tomography is the cross-sectional imaging of an object through the use of transmitted or reflected waves, captured by the object exposure to the waves from a set of known angles. It has many different applications in science, industry, and most prominently, medicine. Since the invention of the Computed Tomography (**CT**) machine in 1972, by Hounsfield [5], tomographic imaging techniques have had a revolutionary impact, allowing doctors to see inside their patients, without having to subject them to more invasive procedures [10].

Mathematical basis for tomography were set by Johannes Radon in 1917. At the time, he postulated that it is possible to represent a function written in \mathbb{R} in the space of straight lines, \mathbb{L} through the function's line integrals. A line integral is an integral in which the function that is being integrated is evaluated along a curved path, a line. In the tomographic case, these line integrals represent a measurement on a ray that traverses the Region Of Interest (**ROI**). Each set of line integrals, characterized by an incidence angle, is called a projection (see Figure 3.1). To perform a tomographic reconstruction, the machine must take many projections around the object. To the set of projections arranged in matrix form by detector and projection angle, we call sinogram. All reconstruction methods, analytical and iterative, revolve around going from reality to sinogram to image [1, 3, 6–9].




Figure 3.1: A schematic representation of a projection.

There are two broad algorithm families when it comes to tomographic reconstruction, regarding the physics of the problem. The problem can involve either non-diffracting sources (light travels in straight lines), such as the X-Rays in a conventional CT exam; or diffracting sources, such as micro-waves or ultrasound in more research-oriented applications. In this document, I will not address the latter family, since I will not be applying them in my work. In the next few paragraphs, I will discuss the first family of algorithms, and describe how an image can be reconstructed from an object's projections when the radiation source is non-diffracting.

Let's consider the case in which we deal with a single ray of solar light entering the atmosphere at a given point. Since the atmosphere contains numerous absorbents and comparable atmospheric effects, the ray changes from the point where it enters the atmosphere to the point at which it is measured by a detector. Total absorption will depend on the pollutant species, their cross-section and their concentration, since it obeys Lambert-Beer's law. Looking from another angle, this absorption is also the line integral that we will use to reconstruct our image. With DOAS, it is possible to measure several pollutants at the same time, but for simplicity (and since it is one of the most studied compounds in the field), let's consider that the single pollutant in our atmospheric mixture is NO_2 .

3.2.2 Initial Considerations

The problem of tomographic reconstruction can be approached in a number of ways, depending mostly on the authors. In my literary search, I have found that Kak and Slaney [9] have certainly explained this problem in one of the clearer ways available. Therefore, I shall base the rest of my presentation in their writings, and complement with other authors' notes wherever necessary.

Considering the coordinate system displayed in Figure 3.2. In this schematic, the object is represented by the function $f(x, y)$. The (θ, t) parameters can be used to define any line in this schematic. Line AB in particular can be written:

$$x \cdot \cos(\theta) + y \cdot \sin(\theta) = t \quad (3.1)$$




Figure 3.2: Schematic representation for coordinate setting.

And if we were to write a line integral along this line, it would look like Equation 3.2, the Radon transform of function $f(x, y)$:

$$P_\theta(t) = \int_{-\infty}^{\infty} \int_{-\infty}^{\infty} f(x, y) \cdot \delta(x \cdot \cos(\theta) + y \cdot \sin(\theta) - t) dx dy \quad (3.2)$$

Where δ , the delta function, is defined in Equation 3.3.

$$\delta(\phi) = \begin{cases} 1, & \phi = 0 \\ 0, & \text{otherwise} \end{cases} \quad (3.3)$$

As I have mentioned previously, a projection is a set of line integrals such as $P_\theta(t)$. Geometry plays a very important role in how the integrals are written and solved for reconstruction. The simplest case is the one where the set is acquired in a row, describing what is called a parallel geometry. Another more complex case is when a single point source is used as origin for all rays, forming a fan. This is called a fanbeam array. There are other possible geometries, but they fall out of the scope of this work and will therefore not be addressed any further.

3.2.3 The Fourier Slice Theorem

The Fourier Slice Theorem (FST) is the most important component of the most important algorithm in tomographic inversion, the Filtered BackProjection algorithm (FBP). FST is based on the equality relation between the two-dimensional Fourier Transform (FT) of the object function and the one-dimensional FT of the object's projection at an angle θ . Let's start by writing the 2D FT for the object function, Equation 3.4, and the 1D FT of projection P_θ , in Equation 3.5.

$$F(u, v) = \int_{-\infty}^{\infty} \int_{-\infty}^{\infty} f(x, y) \cdot \exp[-j2\pi(ux + vy)] dx dy \quad (3.4)$$

$$S_\theta(\omega) = \int_{-\infty}^{\infty} P_\theta \cdot \exp[-j2\pi\omega t] \quad (3.5)$$

For simplicity, let's consider the 2D FT at the line defined by $v = 0$ in the frequency domain. We rewrite the 2D FT integral as:

$$F(u, 0) = \int_{-\infty}^{\infty} \int_{-\infty}^{\infty} f(x, y) \cdot \exp[-j2\pi\omega ux] dx dy \quad (3.6)$$

Notice that y is not present in the phase factor of the FT expression anymore, and this means we can rearrange the integral as:

$$F(u, 0) = \int_{-\infty}^{\infty} \left[\int_{-\infty}^{\infty} \mathbf{f}(\mathbf{x}, \mathbf{y}) d\mathbf{y} \right] \cdot \exp[-j2\pi\omega ux] dx \quad (3.7)$$

Now, the **bold** part of Equation 3.7 is similar to Equation 3.2. It is precisely that equation, considering $\theta = 0$ and a constant value of x , as in Equation 3.8.

$$P_{\theta=0}(x) = \int_{-\infty}^{\infty} f(x, y) dy \quad (3.8)$$

This in turn can be substituted in Equation 3.7, finally arriving at:

$$F(u, 0) = \int_{-\infty}^{\infty} P_{\theta=0}(x) \cdot \exp[-j2\pi ux] dx \quad (3.9)$$

And this is the one-dimensional **FT** for the projection at angle $\theta = 0$. Finally, the enunciation of the Fourier Slice Theorem:

The Fourier Transform of a parallel projection of an image $f(x, y)$ taken at angle θ gives a slice of the two-dimensional Fourier Transform, $F(u, v)$, subtending an angle θ with the u -axis (see Figure 3.3)




Figure 3.3: The **FST**, a schematic representation.

3.2.4 The Filtered BackProjection Algorithm

The idea behind the **FBP** algorithm is almost a conclusion for the **FST** theorem. If one assembles sufficient *parallel* projections in the Fourier space (and in their correct places), then the 2D **IFT** of the summation of all those projections will render an estimation for the object function, $f(x, y)$. In practice, things are not this smooth.

accumulation of low
frequency components

3.3 DOAS tomography instrumentation

Research Methodology

4.1 Aimed contribution

4.2 Detailed work plan and scheduling

4.3 Validation methodology

4.4 Integration with other research activities

Bibliography

- [1] P. P. Bruyant. “Analytic and iterative reconstruction algorithms in SPECT.” In: *Journal of nuclear medicine : official publication, Society of Nuclear Medicine* 43.10 (2002), pp. 1343–58. ISSN: 0161-5505. URL: <http://jnm.snmjournals.org/cgi/content/abstract/43/10/1343%7B%5C%%7D5Cnhttp://jnm.snmjournals.org/cgi/content/full/43/10/1343%7B%5C%%7D5Cnhttp://www.ncbi.nlm.nih.gov/pubmed/12368373>.
- [2] M. Carugno, D. Consonni, G. Randi, D. Catelan, L. Grisotto, P. A. Bertazzi, A. Biggeri, and M. Baccini. “Air pollution exposure, cause-specific deaths and hospitalizations in a highly polluted Italian region.” In: *Environmental Research* 147 (May 2016), pp. 415–424. ISSN: 00139351. DOI: 10.1016/j.envres.2016.03.003. URL: <https://linkinghub.elsevier.com/retrieve/pii/S0013935116300834>.
- [3] M. Defrise, P. E. Kinahan, and C. J. Michel. “4 Image Reconstruction Algorithms in PET * 2D Data Organization.” In: (2003), pp. 91–114.
- [4] A. Ghorani-Azam, B. Riahi-Zanjani, and M. Balali-Mood. “Effects of air pollution on human health and practical measures for prevention in Iran.” In: *Journal of Research in Medical Sciences* 21.1 (2016), p. 65. ISSN: 1735-1995. DOI: 10.4103/1735-1995.189646. URL: <http://www.jmsjournal.net/text.asp?2016/21/1/65/189646>.
- [5] R. Gunderman. *Essential Radiology: Clinical Presentation, Pathophysiology, Imaging, 2nd ed.* 2nd ed. Thieme, 2006. ISBN: 9781588900821.
- [6] G. T. Herman. “Image Reconstruction From Projections.” In: *Real-Time Imaging* 1.1 (1995), pp. 3–18. ISSN: 10772014. DOI: 10.1006/rtim.1995.1002.
- [7] G. T. Herman. *Fundamentals of Computerized Tomography*. Advances in Pattern Recognition. London: Springer London, 2009. ISBN: 978-1-85233-617-2. DOI: 10.1007/978-1-84628-723-7. URL: <http://link.springer.com/10.1007/978-1-84628-723-7>.

BIBLIOGRAPHY

- [8] G. T. Herman, A. Lent, and S. W. Rowland. “ART: Mathematics and applications. A report on the mathematical foundations and on the applicability to real data of the algebraic reconstruction techniques.” In: *Journal of Theoretical Biology* 42.1 (1973). ISSN: 10958541. DOI: [10.1016/0022-5193\(73\)90145-8](https://doi.org/10.1016/0022-5193(73)90145-8).
- [9] A. C. Kak. “Algebraic Reconstruction Algorithms.” In: *Computerized Tomographic Imaging*. 2001, pp. 275–296. ISBN: 978-0-89871-494-4. DOI: [10.1137/1.9780898719277.ch7](https://doi.org/10.1137/1.9780898719277.ch7).
- [10] A. C. Kak and M. Slaney. *Principles of Computerized Tomographic Imaging*. 1. Society for Industrial and Applied Mathematics, Jan. 2001, pp. 49–112. ISBN: 978-0-89871-494-4. DOI: [10.1137/1.9780898719277](https://doi.org/10.1137/1.9780898719277). URL: <http://epubs.siam.org/doi/book/10.1137/1.9780898719277>.
- [11] M. Kampa and E. Castanas. “Human health effects of air pollution.” In: *Environmental Pollution* 151.2 (Jan. 2008), pp. 362–367. ISSN: 02697491. DOI: [10.1016/j.envpol.2007.06.012](https://doi.org/10.1016/j.envpol.2007.06.012). URL: <https://linkinghub.elsevier.com/retrieve/pii/S0269749107002849>.
- [12] G. M. Lovett, T. H. Tear, D. C. Evers, S. E. Findlay, B. J. Cosby, J. K. Dunscomb, C. T. Driscoll, and K. C. Weathers. “Effects of Air Pollution on Ecosystems and Biological Diversity in the Eastern United States.” In: *Annals of the New York Academy of Sciences* 1162.1 (Apr. 2009), pp. 99–135. ISSN: 00778923. DOI: [10.1111/j.1749-6632.2009.04153.x](https://doi.org/10.1111/j.1749-6632.2009.04153.x). URL: <http://doi.wiley.com/10.1111/j.1749-6632.2009.04153.x>.
- [13] D. Vallero. *Fundamentals of air pollution*. 5th ed. Academic Press, 2014. ISBN: 9780124017337.
- [14] World Health Organisation Europe. “Health Aspects of Air Pollution Results from the WHO Project "Systematic Review of Health Aspects of Air Pollution in Europe"." In: *World Health* (2004), p. 30.