

EdLog Program: C:\Users\rcbusey\Dropbox\logger_programs\barrow\GRID_SITE_2011-10-02.CSI
CRBasic Program: C:\Users\rcbusey\Dropbox\logger_programs\barrow\GRID_SITE_2011-10-02.CR1

3	125	ProblemMessage, Step Loop Index must be manually converted. Probable solution is to modify loop count and use FOR loop STEP option [EDLogLine=20] [InstCode=90]	
5	166	ProblemMessage, Step Loop Index must be manually converted. Probable solution is to modify loop count and use FOR loop STEP option [EDLogLine=28] [InstCode=90]	
7	246	ProblemMessage,Output instruction found in IF block. CALLTABLE placed outside IF. Program must be reviewed.; [EDLogLine=42] [InstCode=70]	
8	250	ProblemMessage,Output instruction found in IF block. CALLTABLE placed outside IF. Program must be reviewed.; [EDLogLine=43] [InstCode=77]	
9	253	ProblemMessage,Output instruction found in IF block. CALLTABLE placed outside IF. Program must be reviewed.; [EDLogLine=44] [InstCode=71]	
10	273	ProblemMessage,Output instruction found in IF block. CALLTABLE placed outside IF. Program must be reviewed.; [EDLogLine=49] [InstCode=70]	
11	277	ProblemMessage,Output instruction found in IF block. CALLTABLE placed outside IF. Program must be reviewed.; [EDLogLine=50] [InstCode=77]	
12	280	ProblemMessage,Output instruction found in IF block. CALLTABLE placed outside IF. Program must be reviewed.; [EDLogLine=51] [InstCode=71]	