

**Sensors and actuators used:**

- Pressure: PT660, PT681
- Valve: FV581, FV582, FV583, FV584, FV585, FV587, FV640, FV641, FV642, FV643, Switch Valves
- Control valve: CV581, CV583.

**The user chooses:**

- Cryostat isolated or not isolated
- Mode : Vacuum/Liquid, Magnet
- Pressure setpoint: PT660Maxi

**Initial conditions:**

- Sequences 6 and 7 stopped
- Sequences from 10 to 20 stopped

