


Q: est & ess



Solution ① $e_{ss} = \lim_{k \rightarrow \infty} e(kT)$

$$= \lim_{z \rightarrow 1} (1 - z^{-1}) E(z)$$

Where $E(z) = R(z) - B(z)$

$$G_{ZH} G_p G_H(z) = \frac{B(z)}{E(z)} = G_H(z)$$

$$B(z) = G_H(z) E(z)$$

$$E(z) = R(z) - G_H(z) E(z)$$

$$(1 + G_H(z)) E(z) = R(z)$$

$$E(z) = \frac{R(z)}{1 + G_H(z)}$$

So $e_{ss} = \lim_{z \rightarrow 1} (1 - z^{-1}) E(z)$

$$= \lim_{z \rightarrow 1} (1 - z^{-1}) \frac{R(z)}{1 + G_H(z)}$$

$$\begin{aligned} \textcircled{2} e_{st} &= \lim_{k \rightarrow \infty} (r(k) - c(k)) \\ &= \lim_{z \rightarrow 1} (1 - z^{-1}) (R(z) - C(z)) \end{aligned}$$

$$\text{where } G_{cl}(z) = \frac{C(z)}{R(z)}$$

$$C(z) = R(z) G_{cl}(z)$$

$$\begin{aligned} e_{st} &= \lim_{z \rightarrow 1} (1 - z^{-1}) (R(z) - R(z) G_{cl}(z)) \\ &= \lim_{z \rightarrow 1} (1 - z^{-1}) R(z) (1 - G_{cl}(z)) \end{aligned}$$

稳态跟踪误差est: steady state tracking error

$$e_{st} = \lim_{k \rightarrow \infty} (r(k) - \textcircled{c}(k)) = \lim_{z \rightarrow 1} ((1 - z^{-1})R(z)[1 - G_d(z)])$$

稳态误差ess: steady-state error

$$e_{ss} = \lim_{z \rightarrow 1} \left[(1 - z^{-1}) \frac{1}{1 + GH(z)} R(z) \right]$$

直接R-C是跟踪 R-B是驱动
一个是减开环传函，一个是减闭环传函
反馈传函为1的话，两者一样，才能共用公式

$$GH(z) = (1 - z^{-1}) \mathcal{Z} \left[\frac{G_P(s)H(s)}{s} \right]$$




Figure 4-14: Discrete-time control system

$$E(z) = \frac{1}{1 + GH(z)} R(z)$$

The above analysis applies to the system in Figure 4-14.

上述分析适用于图4-14中的系统。

It is important to note that the above $E(z)$ is the actuating error $E(z) = R(z) - B(z)$. This is different from the tracking error $R(z) - C(z)$!

需要注意的是，上述 $E(z)$ 是驱动误差 $E(z) = R(z) - B(z)$ 。
这与跟踪误差 $R(z) - C(z)$ 不同！

For other system configurations where the sampler(s) are placed differently, the results have to be modified. A few examples are given in Table 4-5.

对于放置不同采样器的其他系统配置，必须修改结果。
如表4-5所示。