ITEM TO CHECK	PROVIDED Y/N	COMMENT
1. SPECIMEN		
Detailed description of specimen type and numbers	**Y**	Supplementary File 1
Sampling procedure (including time to storage)	**Y**	Supplementary File 1
Sample aliquotation, storage conditions and duration	**Y**	Immediately extracted
2. NUCLEIC ACID EXTRACTION		
Description of extraction method including amount of sample processed	**Y**	
Volume of solvent used to elute/resuspend extract	**Y**	
Number of extraction replicates	**Y**	No extraction replicates
Extraction blanks included?	**Y**	Growth medium
3. NUCLEIC ACID ASSESSMENT AND STORAGE		
Method to evaluate quality of nucleic acids		N/A
Method to evaluate quantity of nucleic acids (including molecular weight and calculations when using mass)		N/A (material too scarce)
Storage conditions: temperature, concentration, duration, buffer, aliquots	**Y**	20 µL of 1:30 dilution; up to 4 months at -20°C
Clear description of dilution steps used to prepare working DNA solution	**Y**	No further dilution
4. NUCLEIC ACID MODIFICATION		
Template modification (digestion, sonication, pre-amplification, bisulphite etc.)	**N**	N/A
Details of repurification following modification if performed	**N**	N/A
5. REVERSE TRANSCRIPTION	**NA**	
cDNA priming method and concentration		

One or two step protocol (include reaction details for two step)		
Amount of RNA added per reaction		
Detailed reaction components and conditions		
Estimated copies measured with and without addition of RT*		
Manufacturer of reagents used with catalogue and lot numbers		
Storage of cDNA: temperature, concentration, duration, buffer and aliquots		
6. dPCR OLIGONUCLEOTIDES DESIGN AND TARGET INFORMATION		
Sequence accession number or official gene symbol	**Y**	NC_012920
Method (software) used for design and *in silico* verification	**Y**	TaqMan custom tool, NCBI primer blast
Location of amplicon	**Y**	m.11725 - m.11817
Amplicon length	**Y**	93
Primer and probe sequences (or amplicon context sequence)**	**Y**	Forward: 5'TGCCTAGCAAACTCAAACTACGAA Reverse: 5' GGAGTAGAGTTTGAAGTCCTTGAGA
Location and identity of any modifications	**Y**	
Manufacturer of oligonucleotides	**Y**	Thermo Fisher
7. dPCR PROTOCOL		
Manufacturer of dPCR instrument and instrument model	**Y**	
Buffer/kit manufacturer with catalogue and lot number	**Y**	
Primer and probe concentration	**Y**	
Pre-reaction volume and composition (incl. amount of template and if restriction enzyme added)	**Y**	
Template treatment (initial heating or chemical denaturation)	**Y**	

Polymerase identity and concentration, Mg++ and dNTP concentrations***	**Y**	1X Perfecta Multiplex qScript ToughMix (VWR International, Cat. No. 733-2324)
Complete thermocycling parameters	**Y**	
8. ASSAY VALIDATION		
Details of optimisation performed	**Y**	Thermocylcing optimization using qPCR, input optimization using ddPCR
Analytical specificity (vs. related sequences) and limit of blank (LOB)	**N**	Data of all blanks provided
Analytical sensitivity/LoD and how this was evaluated	**y**	
Testing for inhibitors (from biological matrix/extraction)	**N**	Relative quantity of both mutations is assessed
9. DATA ANALYSIS		
Description of dPCR experimental design	**Y**	
Comprehensive details negative and positive of controls (whether applied for QC or for estimation of error)	**Y**	
Partition classification method (thresholding)	**Y**	ddpcRquant
Examples of positive and negative experimental results (including fluorescence plots in supplemental material)	**Y**	
Description of technical replication	**Y**	3 technical replicates of 6 samples
Repeatability (intra-experiment variation)	**N**	
Reproducibility (inter-experiment/user/lab etc. variation)	**Y**	technical replicates performed on separate days; experiments performed by two users
Number of partitions measured (average and standard deviation)	**Y**	Average: 21998; standard deviation: 3111
Partition volume	**Y**	0.43 ± 0.03 nL
Copies per partition (λ or equivalent) (average and standard deviation)	**Y**	average λ: 0.043173; standard deviation: 0.029124
dPCR analysis program (source, version)	**Y**	
Description of normalisation method	**Y**	mutational copy number vs total copy number
Statistical methods used for analysis	**Y**	Bland-Altman plot
Data transparency	able on request:	

Table S1. dMIQE2020 checklist for authors, reviewers and editors. Authors should fill detail whether information is provided. Where 'yes' is selected use comment box to detail location of information or to include the information. Where 'no' is selected use comment box to outline rationale for omission. Sections 4 and 5 may not apply depending on experiment.

* Assessing the absence of DNA using a no RT assay (or where RT has been inactivated) is essential when first extracting RNA. Once the sample has been validated as DNA-free, inclusion of a no-RT control is desirable, but no longer essential.

** Disclosure of the primer and probe sequence is highly desirable and strongly encouraged. However, since not all commercial pre-designed assay vendors provide this information when it is not available assay context sequences must be submitted (Bustin et al. Primer sequence disclosure: A clarification of the miqe guidelines. Clin Chem 2011;57:919-21.)

*** Details of reaction components is highly desirable, however not always possible for commercial disclosure reasons. Inclusion of catalogue number is essential where component reagent details are not available.