Supplemental Table 1: Obtained mutant loads at position m.8993 (T>G) in mixed (heteroplasmic) and pure (homoplasmic) gDNA samples after direct mutation site amplification by conventional PCR and after WGA followed by mutation site amplification

Expected mutant load (%)	Obtained mutant load (%)	
	Conventional PCR	WGA and conventional PCR
0.0	ND	ND
5.0	3.2	3.4
10.0	14.7	13.9
20.0	25.4	22.3
40.0	44.9	40.2
60.0	64.4	62.5
80.0	81.7	77.7
100	100	100

Mean coverage of the mutation site was 62,835 reads, ranging from 30,023-96,205. ND=not detected.