

Device Manager

Data Manager

Audit Trail



Device List

- ☒ ITAG-SingleUse
- ☐ ITAG-PDF
- ☐ ELog
- ☐ TempROD

 Connect

 Search

TempRod (TROD-A) Connected

Current Status: Logging
Memory: 63%
Battery: 89%



View Manager



Device Configuration



Summary

Data Graph

Data List

Report Editor

Sign

Save

Unsigned



Product Name: TempROD

Model: TROD-A

Description: Vaccine_Shanghai_to_Belmont

Serial Number: IMEC0997

Trip Number: IMEC0997

Logger Settings

Log Cycle: 7d15h35m

Start Mode: Manual Start

Log Interval: 30m15s

Start Delay: 30m

Logging Summary

Start Time: Jun 28, 2011 05:34:11 PM
Stop Time: Jul 11, 2011 10:09:03 PM

First Point: Jun 28, 2011 05:34:11 PM
Trip Length:
Data Points:

Highest Temperature: 33.5 @ Jun 29, 2011 08:29:22 PM
Lowest Temperature: 11.5 @ Jun 29, 2011 11:23:47 AM

Average Temperature: 27.2C
MKT:

Alarms

Alarm Settings

Alarm Type: Single Event

High Limit: 8C

Low Alarm: 2C

Alarm Type: Cumulative Alarm

Delay Time: 30m

Delay Time: 30m

Alarm Events

High Alarm (8C)

Low Alarm (2C)

Number of Events: 15

Number of Events: 27

Total Time Above HL: 5h49m

Total Time Below HL: 4h21m

Longest Time Above HL: 3h10m

Longest Time Below HL: 1h2m

First Triggered: Jun 29, 2011
11:17:57 AM

First Triggered: Jun 29, 2011
11:17:57 AM

Device Manager

Data Manager

Audit Trail

Device List

- ☒ ITAG-SingleUse
- ☐ ITAG-PDF
- ☐ ELog
- ☐ TempROD

 Connect

 Search

TempRod (TROD-A) Connected

Current Status: Logging
Memory: 63%
Battery: 89%



View Manager



Device Configuration



Summary

Data Graph

Data List

Report Editor

Sign

Save

Unsigned

Product Name: TempROD

Model: TROD-A

Description: Vaccine, Shanghai to Belmont

Serial Number: IMEC0997

Trip Number: IMEC0997

Logger Settings

Log Cycle: 7d15h35m

Start Mode: Manual Start

Log Interval: 30m15s

Start Delay: 30m

Logging Summary

Start Time: Jun 28, 2011 05:34:11 PM
Stop Time: Jul 11, 2011 10:09:03 PM

First Point: Jun 28, 2011 05:34:11 PM
Trip Length:
Data Points:

Highest Temperature: 33.5 @ Jun 29, 2011 08:29:22 PM
Lowest Temperature: 11.5 @ Jun 29, 2011 11:23:47 AM

Average Temperature: 27.2C
MKT:

Alarms

Alarm Settings

Alarm Type: Single Ev

Alarm Type: Cumulati

Alarm Events

High Alarm (8C)

Number of Events: 15

Total Time Above HL: 5h49m
Longest Time Above HL: 3h10m

First Triggered: Jun 29, 2011
11:17:57 AM

Low Alarm (2C)

Number of Events: 27

Total Time Below HL: 4h21m
Longest Time Below HL: 1h2m

First Triggered: Jun 29, 2011
11:17:57 AM

User Name

Password

Confirm

Available Meanings

- ☒ Approved
- ☐ Verified

Sign

Device Manager

Data Manager

Audit Trail



Device List

- ☒ ITAG-SingleUse
- ☐ ITAG-PDF
- ☐ ELog
- ☐ TempROD

 Connect

 Search

TempRod (TROD-A) Connected

Current Status: Logging
Memory: 63%
Battery: 89%



View Manager



Device Configuration



Summary

Data Graph

Data List

Report Editor

Sign

Save

▲ Signed

[1] Hugh J (Hugh Jackman)_Approved_Jun 21, 2011 01:22:36 AM



Product Name: TempROD

Model: TROD-A

Description: Vaccine_Shanghai_to_Belmont

Serial Number: IMEC0997

Trip Number: IMEC0997

Logger Settings

Log Cycle: 7d15h35m

Start Mode: Manual Start

Log Interval: 30m15s

Start Delay: 30m

Logging Summary

Start Time: Jun 28, 2011 05:34:11 PM
Stop Time: Jul 11, 2011 10:09:03 PM

First Point: Jun 28, 2011 05:34:11 PM
Trip Length:
Data Points:

Highest Temperature: 33.5 @ Jun 29, 2011 08:29:22 PM
Lowest Temperature: 11.5 @ Jun 29, 2011 11:23:47 AM

Average Temperature: 27.2C
MKT:

Alarms

Alarm Settings

Alarm Type: Single Event

High Limit: 8C

Low Alarm: 2C

Alarm Type: Cumulative Alarm

Delay Time: 30m

Delay Time: 30m

Alarm Events

High Alarm (8C)

Low Alarm (2C)

Number of Events: 15

Number of Events: 27

Total Time Above HL: 5h49m

Total Time Below HL: 4h21m

Longest Time Above HL: 3h10m

Longest Time Below HL: 1h2m

First Triggered: Jun 29, 2011
11:17:57 AM

First Triggered: Jun 29, 2011
11:17:57 AM

Device Manager

Data Manager

Audit Trail

Device List

- ☒ ITAG-SingleUse
- ☐ ITAG-PDF
- ☐ ELog
- ☐ TempROD

 Connect

 Search

TempRod (TROD-A) Connected

Current Status: Logging
Memory: 63%
Battery: 89%

 View Manager  Device Configuration



Summary

Data Graph

Data List


Report Editor

Sign

Save

▲ Signed

[1] Hugh J (Hugh Jackman)_Approved_Jun 21, 2011 01:22:36 AM



Device Manager

Data Manager

Audit Trial



Device List

- ☒ ITAG-SingleUse
- ☐ ITAG-PDF
- ☐ ELog
- ☐ TempROD

 Connect

 Search

TempRod (TROD-A) Connected

Current Status: Logging
Memory: 63%
Battery: 89%



View Manager



Device Configuration



Summary

Data Graph

Data List

Report Editor

Sign

Save



This window allows you to edit the appearance of the data report you want to present, any edit actions will not effect and alter the integrity of raw data
Please go to "Options" in the upper left corner to change Logo and Contact Information Settings

☐ Logo

☐ Contact Information

☐ Report Title

☐ Device Information

☐ Summary

Comments

Device Manager

Data Manager

Audit Trial

History Data

Filter

From
To

Record

Date

<input checked="" type="checkbox"/>	IMEC0997_IMEC0997	Jun 11, 2011
<input checked="" type="checkbox"/>	IMEC0997_IMEC0996	Jun 11, 2011
<input type="checkbox"/>	IMEC0997_IMEC0995	Jun 11, 2011
<input type="checkbox"/>	ANDK1723_SVGR9934	Jun 11, 2011
<input type="checkbox"/>	ANDK1723_SVGR9934	Jun 11, 2011
<input type="checkbox"/>	ANDK1723_SVGR9934	Jun 11, 2011
<input type="checkbox"/>	ANDK1723_SVGR9934	Jun 11, 2011
<input type="checkbox"/>	ANDK1723_SVGR9934	Jun 11, 2011
<input type="checkbox"/>	ANDK1723_SVGR9934	Jun 11, 2011
<input type="checkbox"/>	ANDK1723_SVGR9934	Jun 11, 2011
<input type="checkbox"/>	ANDK1723_SVGR9934	Jun 11, 2011
<input type="checkbox"/>	ANDK1723_SVGR9934	Jun 11, 2011
<input type="checkbox"/>	ANDK1724_SVGS56231	Jun 11, 2011
<input type="checkbox"/>	ANDK1724_SVGS56231	Jun 11, 2011
<input type="checkbox"/>	ANDK1724_SVGS56231	Jun 11, 2011
<input type="checkbox"/>	ANDK1724_SVGS56231	Jun 11, 2011
<input type="checkbox"/>	ANDK1724_SVGS56231	Jun 11, 2011
<input type="checkbox"/>	ANDK1724_SVGS56231	Jun 11, 2011
<input type="checkbox"/>	ANDK1724_SVGS56231	Jun 11, 2011
<input type="checkbox"/>	ANDK1724_SVGS56231	Jun 11, 2011

Compare

View Manager

☒ Show discription

Delete seleted records
Delete all records


Data Graph

Data List

Report Editor

Save

Temperature (C)



A IMEC0997_IMEC0997

B IMEC0997_IMEC0996

Date/Time

Filter

From
To

Record

Date ▼

[illegible]

Compare

Summary

Data Graph

Data List

Report Editor

Save


[illegible]

Device Manager

Data Manager

Audit Trail



 Audit Trail



- ☒ System Audit Trail ●
- ☒ Analysis Audit Trail ●

Filter Criteria

Action

User Name

Date From

Date To

	Action	User Name	Full Name	Date	Detail
●	Log on	Hugh J	Hugh Jackman	Jun 23, 2011 11:22:33	Success
●	Creat user account	Hugh J	Hugh Jackman	Jun 23, 2011 11:45:12	Maria D
●	Change password	Maira D	Maira Daisy	Jun 24, 2011 07:57:01	Maria D
●	Creat user account	Maira D	Maira Daisy	Jun 24, 2011 07:57:01	Nelson
●	Sign record	Nelson	Nelson Bourn	Jun 25, 2011 14:20:32	IMEC0997_IMEC0997
●	Add comment	Hugh J	Hugh Jackman	Jun 25, 2011 16:05:49	IMEC0997_IMEC0997
●	Creat record	Hugh J	Hugh Jackman	Jun 25, 2011 16:05:49	IMEC0998_IMEC0998