

Plasma Levels of NfL were quantified using Simoa assays. All assays were performed by the same operator. Simoa™ NF-light® Advantage (Quanterix, Billerica, MA, USA) kits were used according to the manufacturer's instructions. The highest dilution factor that allowed reliable quantification of samples was determined as 1:4 before [1]. Plasma was centrifuged at 14,000 x g for 4 minutes, and the upper 90 % transferred to the assay plate. Samples, controls, and blanks were analyzed in duplicates. The lower limit of quantitation (LLOQ) is defined as the lowest standard: (i) with a signal higher than the average signal for the blank plus 9 SDs, and (ii) allows a percent recovery $\geq 100 \pm 20$ %. Two internal control samples were assessed both at the start and end of an assay plate. In x runs over x days the LLOQ was xxx pg/mL. The average % coefficient of variation (%CV) of all samples measured in the study was xxx. The % repeatability was xxx% (sample 1) and xxx% (sample 2). The interassay variance between the runs across x days was xxx% (sample 1) and xxx% (sample 2).

References

1. Mengel D, Liu W, Glynn RJ, Selkoe DJ, Strydom A, Lai F, Rosas HD, Torres A, Patsiogiannis V, Skotko B, Walsh DM: **Dynamics of plasma biomarkers in Down syndrome: the relative levels of Abeta42 decrease with age, whereas NT1 tau and NfL increase.** *Alzheimers Res Ther* 2020, **12**:27.