

Application of Magnetic Resonance Fingerprinting for the Assessment of Stroke damage

Jack Allen, Supervisors: Peter Jeppard & James Kennedy

University of Oxford, Oxford, UK.

Abstract

Keywords: Magnetic Resonance Imaging, Fingerprinting, Stroke

1. Introduction

Magnetic resonance imaging (MRI) has become an established tool for diagnosis and disease monitoring in clinical environments. As with the phenomenon of nuclear magnetic resonance (NMR), MRI makes use of the behaviour of nuclear spins in a magnetic field. However MRI has crucial difference that it allows the acquisition of images from the behaviour of nucleic spins. When a strong, static, magnetic field B_0 is applied to a sample, there is a net alignment of the spins to B_0 , producing net magnetisation along the longitudinal axis of B_0 . Radio frequency (RF) excitation pulses can be applied to tip the net magnetisation of a particular species of nuclei so that it is no longer aligned with B_0 . This produces components of magnetisation that are orthogonal to the static field (transverse magnetisation) and results in reduced longitudinal magnetisation. The tipped spins then precess around the axis of the static field. Over time and during the precession, the transverse and longitudinal components decay and recover, respectively. This evolution occurs according to exponential terms.

The transverse components decay towards zero, according to Eq. (1), where t is the time elapsed. The constant T_2^* , described in Eq. (2), determines the rate of decay and can be reduced by spatial inhomogeneities in the static field. These

¹Since 1880.

inhomogeneities cause variations in the strength of the static field experienced by different spins. The frequency of precession is proportional to the static field strength, therefore static field inhomogeneities will cause spins in different locations will precess at slightly different rates. The effect of the inhomogeneities is represented by T_2' and so in an ideal situation, the magnetisation in Eq. (1) would decay according to solely the spin dephasing and could be described by T_2 .

$$M_{xy}(t) = M_{xy}(0) \exp^{-\frac{t}{T_2^*}} \quad (1)$$

$$\frac{1}{T_2^*} = \frac{1}{T_2} + \frac{1}{T_2'} \quad (2)$$

Equation (3) describes the behaviour of the longitudinal magnetisation, where t is the elapsed time, $M_z(0)$ is the initial longitudinal magnetisation and $M_{z,eq}$ is the equilibrium magnetisation that would occur, given B_0 , but in the absence of excitation pulses. The constant T_1

$$M_z(t) = M_{z,eq}(1 - e^{-\frac{t}{T_1}}) + M_z(0)e^{-\frac{t}{T_1}} \quad (3)$$

The precessing magnetisation induces current in receiver radio frequency coils, which decays as the magnetisation realigns with the static field. By observing the induced current, the relaxation parameters T_1 and T_2 can be measured on a pixel-wise basis. Field gradients are applied to produce controlled differences in spin frequency and phase, which can be used to enable the spatial origin of a signal to be determined [1].

A large portion of MRI research has been directed towards achieving accurate, efficient and reliable images of the human brain. The magnetisation properties of different tissue types can make it possible to achieve contrast between areas of different tissue types.

Commonly used approaches for measuring T_1 and T_2 , such as inversion recovery (IR) and spin echo (SE) require a significant period of time. * mention the fastest times?* *explain why T_1 and T_2 measurements take a long time*

Within the setting of emergency medicine, fast response and diagnosis is crucial for giving the patient the best chance of a positive outcome. Specifically, acute stroke patients benefit from fast treatment. Over factors, such as pH change, are observed, as they can be good indicators of the extent of damage.

Recently, a new approach known as magnetic resonance fingerprinting (MRF) [ma2013] has been developed and applied, which has been used to acquire parameters, such as T1 and T2, in a fraction of the time normally needed [some numbers]. During a fingerprinting experiment, a pseudo-random sequence is played out. This can be achieved by varying factors such as the time between subsequent RF pulses and any flip angles used.

Once a series of images have been acquired via the fingerprinting sequence, the pixel-wise signal variation over the experiment is compared to a previously built dictionary of time courses, in order to find the closest match. The dictionary is built by simulating the signal expected from the experiment, for a variety of difficult combinations of parameters (such as T1 and T2).

2. Materials and Method

2.1. Image acquisition

All images for this report were acquired at the acute vascular imaging centre (AVIC) at the John Radcliffe Hospital in Oxford. [mention the scanner and coil hardware used (3T, model, brand, etc)] [mention the type of sequence]

Using MATLAB [cite MATLAB as Ma 2013 did], a dictionary was composed for [mention T1 and T2 values used]

In order to test the method for a range of T1 and T2 values [if I manage to make another phantom, mention here that the values are similar to those expected in the brain and cite it]

3. Results and Discussion

4. Conclusions

References

- [1] E. M. Haacke, R. Brown, M. Thompson, R. Venkatesan, Magnetic resonance imaging: physical principles and sequence design. 1999, New York: A John Wiley and Sons.