Northwind Metrology - Calibration Certificate

Certificate No: NW-2026-009911

Technician: Ingrid Solheim

Customer contact: Tobias Wren

Phone: +44 20 7946 0555

Email: t.wren@example.org

Site address: 7 Lantern Row
Port Verity, PV9 8ZZ

Calibration date: 2026-05-02

Instrument model: PG-310

Pressure 50 % point reading	5.01	bar	4.98	5.02
Hysteresis reading	0.04	%	0.00	0.10
Supply voltage test point	24.02	V	23.50	24.50