Night Signal: ethics submission package (revised scope)

Gabriel Mamane · Science Fair 2027 · Human participants · Requested by Ms. Ireland on Oct 9 (original due date: October 13)

Status of this draft

On October 9, Ms. Ireland agreed that the project should focus on wearable measurement accuracy and oxygen desaturation detection, and asked for a revised proposal. **She accepts the project only after reviewing that revised proposal.** No ethics approval, certificate, consent or signature has been obtained yet, and no data has been collected from anyone. The student checks and submits this package; nothing here has been submitted.

How to use this package

1. Send the revised proposal to Ms. Ireland (she asked for it on October 9).
2. Fill in every **yellow** box: your grade and Ms. Ireland's email address.
3. Create an account and answer the pre-questionnaire at **approbation.technoscience.ca/pre-questionnaire**. Ms. Ireland asked for this now because of time constraints, even though it is formally required only for projects accepted to the regional fair. Section 1 lists the questions it will most likely ask; check them against the real platform.
4. The ethics form comes after the pre-questionnaire. Use sections 2 to 4 (summary, protocol, risk assessment), and attach the blank data sheets in section 6.
5. **Wait for the Certificate of Approval and signed consent** before any recorded test, even on yourself. Building the device and checking the electronics without wearing it can happen before then, if Ms. Ireland agrees.

Note: the approval website could not be opened while preparing this draft, so section 1 is a best guess at its questions. Per Ms. Ireland, the form appears after the pre-questionnaire and requires an account. The rest follows what Technoscience asks for: protocol, risk assessment, data collection tools, supervisor details and consent.

0. Fill in before you submit

Item	Your answer
Exhibitor	Gabriel Mamane
Grade / school	**Grade __** · Les écoles Azrieli Herzliah, Campus Sylvan Adams
Supervising teacher (adult)	Ms. Jessica Ireland, Science Department; Head teacher for Secondary 5
Supervisor's institution and contact	Les écoles Azrieli Herzliah, Campus Sylvan Adams, 5475 avenue Mountain Sights, Montréal, QC H3W 2Y8 · 514-739-2291, ext. 255 · **email**
Project title	Night Signal: SpO2 Accuracy (revised; original title was “Night Signal: Apnea Detector”)
Category	Health and Medical Sciences · Experimental project
Planned test dates	Only after Ms. Ireland accepts the revised proposal, ethics approval is issued and consent is signed; target October to December (before the January 11 final slides)

1. Pre-questionnaire: likely questions and your answers

Question	Answer
Does the project involve human participants, including yourself?	Yes. Me (the exhibitor). Optionally up to 4 adult volunteers for one 30-minute seated accuracy session, only if Ms. Ireland and the ethics committee agree.
Is there a questionnaire or survey?	No survey. Only a short yes/no comfort check after each session (sheet D).
Are biological samples taken (blood, saliva, tissue)?	No. The sensor shines light on the fingertip, like a store-bought pulse oximeter. Nothing enters the body.
Animals?	No.
Hazardous chemicals, microorganisms or radiation?	No.
Electricity or hazardous equipment?	Low voltage only: 5 V from a standard USB phone charger. There is no battery in the prototype and no mains wiring.
Physical activity or discomfort?	Optional, only if approved: short voluntary breath-holds (20 seconds at most) by the exhibitor only, seated, awake and supervised by an adult.
Is it a medical device or a diagnosis?	No. It is an educational prototype. It cannot diagnose or confirm sleep apnea or any other condition, and is never used to treat anyone.
Risk level	Low: no greater than everyday life (see section 4).

2. Project summary

Pulse oximeters estimate blood oxygen saturation (SpO_2) by shining red and infrared light through the fingertip. A short drop in SpO_2 is called an oxygen desaturation. Desaturations can happen during sleep apnea, but they also have other causes (movement, poor sensor contact, other conditions), and not every apnea causes a large drop. Diagnosing sleep apnea requires a sleep study that also measures airflow, breathing effort and sleep.

This project builds a low-cost wearable from an ESP32 microcontroller and a MAX30102 fingertip light sensor (with an MPU-6050 motion sensor used to exclude readings taken during movement). It records SpO_2 and heart rate once per second. The project measures how closely these readings agree with a separate Health Canada-licensed fingertip pulse oximeter, and whether software can reliably flag oxygen desaturation events in the wearable's own recordings.

> **What this project cannot do.** It measures only oxygen and heart rate. It cannot detect, diagnose or confirm sleep apnea, and it does not calculate an apnea–hypopnea index (AHI). Any event rate is reported as flagged desaturation events per hour of valid recording.

Question: How closely do SpO_2 and heart-rate readings from a low-cost wearable (MAX30102 + ESP32) agree with a Health Canada-licensed fingertip pulse oximeter, and can software reliably flag oxygen desaturation events (a drop of 3 percentage points or more lasting at least 10 seconds) in the wearable's recordings?

Hypothesis: If the wearable is worn correctly and only good-quality readings are used, then its SpO_2 will agree with the reference oximeter within about ±3 percentage points (95% limits of agreement), because both measure red/infrared light absorption, and the software will flag the dips that the reference also shows.

3. Research protocol (planned; nothing has been carried out yet)

3.1 Participants

Code	Who	Takes part in
P1	Gabriel Mamane (exhibitor, minor)	Accuracy sessions; if approved, the optional breath-hold test and overnight recordings. Parent/guardian consent required.
P2 to P5	Optional: up to 4 adult volunteers (18+), e.g. family members, only if approved.	Accuracy session only (30 min, seated, awake). No breath-holds, no nights.

Who can take part: healthy and with no known heart or lung disease. People with a history of fainting are not included.

Who cannot take part: anyone who does not want to, or who feels unwell on the day. Taking part is voluntary, and anyone can stop at any time without giving a reason.

3.2 Equipment

- Night Signal wearable: ESP32 board, MAX30102 fingertip sensor in a Velcro cuff, MPU-6050 motion sensor, powered by a 5 V USB phone charger.
- Reference: a fingertip pulse oximeter with a Health Canada medical device licence (**model and licence number to confirm**). Fingertip oximeters are typically specified to about ±2 percentage points; the exact figure will be taken from this model's manual. The comparison therefore measures agreement between two devices, not absolute truth.
- Laptop with the recording and analysis software; a timer; the blank data sheets (section 6); alcohol wipes.

3.3 Procedures

A. Device check (no participant). Before every session I run check_device.py, which confirms that data arrives every second, both sensors pass their self-test and readings are in a normal range. The session continues only if the device reports READY.

B. Accuracy sessions: the main test (P1, and volunteers if approved; about 35 min each). The participant sits and rests for 5 minutes with warm hands. They wear the wearable on one index finger and the reference oximeter on the other hand. Every 2 minutes for 30 minutes, I write down the time and both devices' SpO_2 and pulse on sheet A (15 pairs per session). They breathe normally throughout. P1 repeats this on several different days.

C. Signal-quality checks (data processing, no participant). A one-second reading is used only if a finger is detected, the sensor marks the reading valid, SpO_2 is between 70 and 100%, and there was no sudden movement within 5 seconds. Each report states the percentage of valid signal. Agreement is calculated as bias, 95% limits of agreement and mean absolute difference (Bland–Altman method).

D. Breath-hold test (optional; P1 only; only if approved; about 45 min; adult supervisor present). Seated and awake, after 2 minutes of normal breathing, I do short voluntary breath-holds: up to 5 each of 10, 15 and 20 seconds, with at least 2 minutes of normal breathing between holds. The supervisor times them and logs the start and end of each hold on sheet B. **Purpose and limit:** this only checks whether the software flags short, deliberately induced dips at the logged times. An awake breath-hold is not a sleep apnea, so this test does not validate detecting sleep apnea during sleep. Short holds may not lower SpO_2 by 3 points; if so, that is reported, and holds are never made longer. Rules:

- Never longer than 20 seconds, always seated, never standing or in water, and never while asleep.
- Stop immediately if there is any dizziness, discomfort or tingling. Stop the whole session if the reference oximeter reads below 90%.

E. Overnight recordings (optional; P1 only; only if approved; up to 5 nights at home). I wear the cuff on my index finger and the armband pouch on my forearm during my normal sleep. Nothing about my sleep is changed. The USB cable runs along my arm to the side of the bed, never near the neck, and is clipped to the mattress. I can take it off at any time. Each morning I fill in sheet C and the comfort check (sheet D). These recordings measure only the percentage of valid signal and the number of flagged desaturation events per hour of valid recording; because the reference oximeter does not record overnight, these events cannot be verified, and they are not interpreted as apneas.

3.4 Data and privacy

- Recorded: blood oxygen (%), heart rate, motion (used to exclude readings during movement), finger-contact flag and the time. No names, photos of faces or other personal information.
- Each participant gets a code (P1 to P5). The list linking codes to names is kept on paper by the supervisor, not with the data.
- Data stays on my own laptop and is not uploaded or shared. Only I and my supervisor see the raw data.
- Results are shown as anonymous graphs and averages. Volunteers' data is deleted after the science fair, or sooner on request.

3.5 If a reading looks unusual

The wearable is an educational prototype and I do not interpret anyone's health. If the reference oximeter shows a resting value below 92%, the session stops and the participant (and parent/guardian for a minor) is advised to speak to a doctor. If my own night recordings show many oxygen dips, I will tell my parents. The device cannot say whether anyone has sleep apnea; at most, such a result is a reason to ask a doctor.

4. Risk assessment

> **Overall: low risk.** The main test is sitting still with two fingertip sensors, which is what a pharmacy pulse oximeter does. The optional breath-hold test is limited to 20 seconds, seated and supervised, and happens only if approved.

Possible risk	Who	Likelihood	Severity	How it is prevented
Light-headedness during breath-holds (optional test)	P1	Low	Minor	Only if approved. Holds of 20 s or less, seated, 2 min rest between holds, adult present, stop at any discomfort, stop the session below 90% on the reference oximeter. Never while asleep.
Skin irritation or pressure from the finger cuff	All	Low	Minor	Cuff snug, not tight. Remove it at any numbness or tingling, and check the skin after each session.
Electrical	All	Very low	Minor	5 V USB from a certified phone charger. No battery, no mains wiring, insulated and taped connections.
Cable at night (tangling or tripping; optional nights)	P1	Low	Moderate	The cable runs along the arm to the side of the bed, never near the neck, and is clipped to the mattress. It can be removed at any time.
Poor sleep or discomfort (optional nights)	P1	Medium	Minor	The device can be removed at any time. Nights are not scheduled before tests or exams.
Hygiene (shared sensor)	Volunteers	Low	Minor	Sensor and cuff wiped with alcohol between people.
Privacy of health data	All	Low	Minor	Coded IDs, data kept only on my laptop, anonymous results, deletion after the fair.
Worry or false reassurance about a result	All	Low	Minor	It is explained beforehand that this is not a medical test and cannot detect or rule out sleep apnea. Unusual readings mean seeing a doctor, not a diagnosis.

5. Participant information and consent

> Draft for review. If Technoscience's Certificate of Approval is used as the consent form, give this page to each participant to read before they sign it. Keep this version as well for the parent/guardian signature for a minor. No one should sign until the project has been approved.

Project: Night Signal, a science fair project by Gabriel Mamane that tests how closely a low-cost fingertip sensor measures blood oxygen and pulse compared with a store-bought pulse oximeter.

What you would do: sit and rest for about 35 minutes while wearing a soft fingertip sensor on one hand and a store-bought pulse oximeter on the other. Every 2 minutes, both readings are written down. (The optional breath-hold test and night recordings are done only by the student, and only if approved.)

Risks: very low. The sensor shines a little red and infrared light on the fingertip, like a pharmacy pulse oximeter. The cuff may feel tight: tell me and it comes off right away.

Benefits: none directly for you. You help measure how accurate a low-cost oxygen sensor is.

Your rights: taking part is voluntary. You can stop at any time, without giving a reason. You can ask for your data to be deleted.

Privacy: your readings are stored under a code (for example P3), never with your name. Results are shown only as anonymous averages and graphs, and your data is deleted after the science fair.

Not a medical test: this is a student prototype. It cannot tell you anything about your health and cannot detect or rule out sleep apnea. If the store-bought oximeter shows an unusual value, you will be advised to talk to a doctor.

Questions: Gabriel Mamane, or the supervising teacher, Ms. Jessica Ireland (Les écoles Azrieli Herzliah, 514-739-2291, ext. 255)

I have read this information, my questions have been answered, and I agree to take part.

Participant name and code

Participant signature **Date**

Parent/guardian name (if the participant is under 18)

Parent/guardian signature **Date**

Student (Gabriel Mamane) signature **Date**

6. Data collection tools (blank)

Sheet A: Accuracy session

Participant code ____ Session # ____ Date _____ Start time _____ Room temperature ____ °C Wearable on: L / R index finger Device check READY? Y / N

#	Minute	Reference SpO_2 %	Reference pulse	Device SpO_2 %	Device pulse	Notes (movement, cold hands…)
1	0					
2	2					
3	4					
4	6					
5	8					
6	10					
7	12					
8	14					
9	16					
10	18					
11	20					
12	22					
13	24					
14	26					
15	28					
16	30					

Sheet B: Breath-hold test (optional; P1 only; only if approved; supervised)

Date _____ Supervisor _____ Resting reference SpO_2 before starting ____ %

#	Target (s)	Start time	End time	Felt OK? (Y/N)	Lowest ref. SpO_2 %	Notes
1	10					
2	10					
3	10					
4	10					
5	10					
6	15					
7	15					
8	15					
9	15					
10	15					
11	20					
12	20					
13	20					
14	20					
15	20					

Stop rules: stop immediately at any dizziness, discomfort or tingling. Stop the session if the reference oximeter reads below 90%. Hold for 20 seconds at most. Always seated. This test checks dip flagging only; it is not a test of sleep apnea.

Sheet C: Night log (optional; P1 only; only if approved)

Night	Date	Bedtime	Wake time	Device READY? (Y/N)	Removed during night?	Comfort 1–5	Notes
1							
2							
3							
4							
5							

Sheet D: Comfort and safety check (after every session)

Date	Code	Dizziness? (Y/N)	Skin redness? (Y/N)	Numbness / tingling? (Y/N)	Action taken / notes