Thermometer Calibration Log

Business: _____ **Month / year:** _____ **Elevation (ft):** _____

Food probes (°F scale): accurate to ±2°F — §4-203.11(B) · **Air / water thermometers (°F): ±3°F** — §4-203.12(B) · Calibrate per the manufacturer's instructions — §4-502.11(B) · **Ice-point check: 32°F (0°C)** in a full ice-water slurry · Boiling point is 212°F only at sea level; it drops about 1°F per ~500 ft of elevation · Check weekly, after a drop, and before each shift on high-risk days

Date	Time	Thermometer ID / type	Method (ice / boil)	Reading °F	Should read °F	Off by °F	Within limit? (Y/N)	Adjusted to °F	Action (recalibrated / replaced / removed)	Initials

Rules: FDA Food Code 2022 (same limits in the 2026 edition) §4-203.11, §4-203.12, §4-502.11. Not legal advice; your local health department's rules apply and may be stricter.
Keep completed logs for as long as your health department requires - ask your inspector how long that is.

Free printable from ThermoTally · gallerycleanai.github.io/thermotally