


Sample Name: SAMPLE_20251001-1100PM-ELECDON-ORGSA3

=====

Acq. Operator : SYSTEM
Sample Operator : SYSTEM
Acq. Instrument : Green House Gas GC Location : 1 (F)
Injection Date : 10/5/2025 5:15:18 PM
Inj Volume : Manually

Method : C:\Chem32\2\Methods\Gamma-N2O-2.M
Last changed : 10/5/2025 11:52:09 AM by SYSTEM
(modified after loading)



=====

External Standard Report

=====

Sorted By : Signal
Calib. Data Modified : 10/5/2025 11:52:09 AM
Multiplier : 1.0000
Dilution : 1.0000
Do not use Multiplier & Dilution Factor with ISTDs

Signal 1: ECD2 B, Back Signal

RetTime [min]	Type	Area [Hz*s]	Amt/Area	Amount [ng/ul]	Grp	Name
1.846	BB	591.35773	3.23624e-3	1.91377	N20	
2.847		-	-	-	N02	
Totals :				1.91377		

Signal 2: FID1 A, Front Signal

RetTime [min]	Type	Area [pA*s]	Amt/Area	Amount [ng/ul]	Grp	Name
1.141		-	-	-		Methane
1.945		-	-	-		C02
Totals :				0.00000		

Signal 3: TCD3 C, Aux Signal

RetTime [min]	Type	Area [25 µV*s]	Amt/Area	Amount [ng/ul]	Grp	Name
3.272		-	-	-		C02
4.177		-	-	-		N02
5.338		-	-	-		Oxygen
6.396		-	-	-		Nitrogen
8.853		-	-	-		Methane
Totals :				0.00000		

1 Warnings or Errors :

Warning : Calibrated compound(s) not found

*** End of Report ***