


Sample Name: SAMPLE_20250921-1245PM-D87LIQ-C01

=====

Acq. Operator : SYSTEM
Sample Operator : SYSTEM
Acq. Instrument : Green House Gas GC Location : 1 (F)
Injection Date : 9/25/2025 3:20:39 PM
Inj Volume : Manually

Method : C:\Chem32\2\Methods\Gamma-N2O-2.M
Last changed : 9/25/2025 2:53:46 PM by SYSTEM
(modified after loading)



=====

External Standard Report

=====

Sorted By : Signal
Calib. Data Modified : 9/25/2025 2:17:23 PM
Multiplier : 1.0000
Dilution : 1.0000

Do not use Multiplier & Dilution Factor with ISTDs

Signal 1: ECD2 B, Back Signal

RetTime [min]	Type	Area [Hz*s]	Amt/Area	Amount [ng/ul]	Grp	Name
1.845	BB	206.19214	3.75039e-3	7.73302e-1	N20	
2.847		-	-	-	N02	

Totals : 7.73302e-1

Signal 2: FID1 A, Front Signal

RetTime [min]	Type	Area [pA*s]	Amt/Area	Amount [ng/ul]	Grp	Name
1.141		-	-	-		Methane
1.945		-	-	-		C02

Totals : 0.00000

Signal 3: TCD3 C, Aux Signal

RetTime [min]	Type	Area [25 µV*s]	Amt/Area	Amount [ng/ul]	Grp	Name
3.272		-	-	-		C02
4.177		-	-	-		N02
5.338		-	-	-		Oxygen
6.396		-	-	-		Nitrogen
8.853		-	-	-		Methane

Totals : 0.00000

1 Warnings or Errors :

Warning : Calibrated compound(s) not found

*** End of Report ***