


Sample Name: Calibration_N2O-Standards-1010ppm-1

```

=====
Acq. Operator   : SYSTEM
Sample Operator : SYSTEM
Acq. Instrument : Green House Gas GC           Location :    1  (F)
Injection Date  : 9/26/2025 3:35:58 PM
                                           Inj Volume : Manually

Acq. Method     : C:\Chem32\2\Methods\Gamma-N2O-2.M
Last changed    : 9/26/2025 2:00:32 PM by SYSTEM
                  (modified after loading)
Analysis Method : C:\Chem32\2\Methods\Gamma-N2O-2.M
Last changed    : 9/26/2025 3:45:29 PM by SYSTEM
                  (modified after loading)
    
```



External Standard Report

```

=====
Sorted By       :      Signal
Calib. Data Modified :    9/26/2025 3:45:29 PM
Multiplier      :      1.0000
Dilution        :      1.0000
Do not use Multiplier & Dilution Factor with ISTDs
    
```

Signal 1: ECD2 B, Back Signal

RetTime [min]	Type	Area [Hz*s]	Amt/Area	Amount [ng/ul]	Grp	Name
1.857	BBAS	2.21318e5	4.63407e-3	1025.60277		N20
2.847		-	-	-		N02

Totals : 1025.60277

Signal 2: FID1 A, Front Signal

RetTime [min]	Type	Area [pA*s]	Amt/Area	Amount [ng/ul]	Grp	Name
1.141		-	-	-		Methane
1.945		-	-	-		C02

Totals : 0.00000

Signal 3: TCD3 C, Aux Signal

RetTime [min]	Type	Area [25 µV*s]	Amt/Area	Amount [ng/ul]	Grp	Name
3.272		-	-	-		C02
4.177		-	-	-		N02
5.338		-	-	-		Oxygen
6.396		-	-	-		Nitrogen
8.853		-	-	-		Methane

Totals : 0.00000

1 Warnings or Errors :

Warning : Calibrated compound(s) not found

=====
*** End of Report ***