

HINDUSTAN PETROLEUM CORPORATION LIMITED

(A Government of India Enterprise)

QUALITY CONTROL LABORATORY

MAZGAON TERMINAL

LUBE PLANT, HAY BUNDER ROAD, MAZGAON

MUMBAI,400033

LABORATORY TEST REPORT**ENKLO HLP 46 (210L)(0568241)**

Laboratory Serial No.	FB - 1626	Sample Drawn By	OPERATION OFFICER
JDE Serial No	1213570	Date Sample Drawn	13/01/23
Batch No / Lot No	M2-23-A-092	Date Received In Lab	13/01/23
Work Order No	1012583	Date Tested	14/01/23
Tank /Kettle/Location	WHS2	Particular of Sample	DRUM
Plant Code/ Name	14546	Mazgaon TML Lubes-(STP 0101)	Sampling Plan
			IS 1447

Test Line Number	Test Description/ Characteristics	Units	Test Method	Specifications		Result Value
				Min	Max	
1	APPEARANCE	.	VISUAL	B&C	B&C	B&C
3	COPPER STRIP CORROSION@100°C	.	P15 D130	-	1	1
4	DENSITY@29.5	g/ml	P16 D1298 D4052	-	-	.8785
5	EMULSION CHAR @54°C	Milliliter	P91 D1401	40-40-0	40-37-3	40-38-2
6	FLASH POINT	°C	P69 D92	185	-	226
10	FOAMSTABSEQ1	Milliliter	P67 D892	-	0	0
11	FOAMSTABSEQ2	Milliliter	P67 D892	-	0	0
12	FOAMSTABSEQ3	Milliliter	P67 D892	-	0	0
7	FOAMTENDSEQ1	Milliliter	P67 D892	-	150	10
8	FOAMTENDSEQ2	Milliliter	P67 D892	-	75	40
9	FOAMTENDSEQ3	Milliliter	P67 D892	-	150	30
17	KV @ 100	cSt	P25 D7042 D445	-	-	6.820
16	KV @ 40°C	cSt	P 25 D7042 D 445	42.00	50.00	47.11
14	PHOSPHORUS	%Wt	P54 D6443 D1091	.0270	.0600	.0332
13	POUR POINT	°C	P10 D97	-	-15	-15
15	RUST TEST 24HRS SEA	.	P96 D665			P
18	VISCOSITY INDEX (VI)	.	P56 D2270	90	-	98
19	ZINC	%Wt	D6443 EDTA	.026	.065	.037
2	ACIDITY TOTAL	mg KOH/g	P1 D974	-	-	.60

Remarks : Sample meets the specification in above tests.

Note:

- 1) The report refers only to the sample tested.
- 2) The test report shall not be reproduced except in full without prior written approval of the Lab Incharge.
- 3) This test report shall not be used in any advertising media or as evidence in the Court of Law without prior written consent of Laboratory.
- 4) Test results reported are valid at the time of testing
- 5) This is a system generated test report and hence does not require signature.

Reviewed and Authorized by:
DALVI P Y(30044160)